

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
#910, 410-22nd Street East
Saskatoon
Sask.
S7K 5T6
Bid Fax: (306) 975-5397

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
**Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution
Public Works and Government Services
Canada/Réception des soumissions Travaux publics et
Services gouvernementaux Canada
#910, 410 - 22nd Street East
Saskatoon
Sask.
S7K 5T6

Title - Sujet Dataloggers - Fredericton, NB	
Solicitation No. - N° de l'invitation KW405-120085/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client KW405-120085	Date 2012-07-03
GETS Reference No. - N° de référence de SEAG PW-\$STN-197-4491	
File No. - N° de dossier STN-2-35004 (197)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2012-07-30	Time Zone Fuseau horaire Central Standard Time CST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Holt, Judy	Buyer Id - Id de l'acheteur stn197
Telephone No. - N° de téléphone (306) 975-4051 ()	FAX No. - N° de FAX (306) 975-5397
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment to solicitation KW405-120085/A is issued to answer the following question:

Question:

After reviewing the details of the requirement, we would like clarification of some point. Requirement 12 indicates the DO sensor must use optical detection methods, while Requirement 10 requires the unit should have owner replaceable membrane or sensor. Based on our experience of optical DO sensors (optodes), they do not typically have foils or membranes that are easily replaced by the customer, and this type of requirement is usually applicable to electrochemical type DO sensors, such as the Oxyguard. Alternatively, is the end user requesting spare sensors be included in the proposal, and if yes, is the requirement for one spare for each unit?

Answer:

The short answer is that we are looking for sensors that have the ability to be serviced to extend useful life of whole unit. Optical sensors such as YSI ROX do have user replaceable membranes. The entire unit does not have to be replaced for service. If supplier optode has that ability, we would prefer that unit over others, provided all other criteria were met. No, we are not looking to include spare sensors as part of proposal. Expected life of unit as supplied should be longer than a year under use.