

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving - PWGSC / Réception des soumissions -
TPSGC
11 Laurier St./ 11 rue, Laurier
Place du Portage, Phase III
Core 0A1 / Noyau 0A1
Gatineau, Québec K1A 0S5
Bid Fax: (819) 997-9776

SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Scientific, Medical and Photographic Division /
Division de l'équipement scientifique, des produits
photographiques et pharmaceutiques
11 Laurier St./ 11 rue, Laurier
6B1, Place du Portage
Gatineau, Québec K1A 0S5

Title - Sujet PATIENT SIMULATORS		
Solicitation No. - N° de l'invitation W3931-130094/A		Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client W3931-130094		Date 2013-01-25
GETS Reference No. - N° de référence de SEAG PW-\$\$PV-940-61913		
File No. - N° de dossier pv940.W3931-130094	CCC No./N° CCC - FMS No./N° VME	
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2013-02-20		Time Zone Fuseau horaire Eastern Standard Time EST
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>		
Address Enquiries to: - Adresser toutes questions à: Hooper, Marlyn		Buyer Id - Id de l'acheteur pv940
Telephone No. - N° de téléphone (819) 956-2702 ()		FAX No. - N° de FAX (819) 956-3814
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:		

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment is raised to publish all answers to the questions received as of January 23, 2013, and to modify mandatory specification 9.2

Q1.

May you please tell us which Blood Pressure monitor it will be used with? We would need to know the make and model.

A1.

The patient simulators are designed to be connected to a wide range of medical devices that measure temperature, pressure, respiration, ECG, arritmias etc.

These parameters can be measured together in one medical device or one parameter per device.

Some of the medical devices that requires a patient simulator for calibration or test are the following:

- Propaq Welch Allyn mod:206
- Welch Allyn mod: CP200
- Philips mod: HeartStart MRx

Q2

For temperature, article 9.2, stated from 0 to 40 C, is this a requirement or a typing mistake, since there is no simulator that goes to 0 degrees.

A2

The requirement need to be change:

Temperature: Must be able to provide a simulation from the minimum range of 35 oC to 42 oC

Q3

Do you require all necessary cables or are those to be put as options?

A3

CF requires a fabricant to provide a list of cables or adapters the its product are able to connect. CF will decide which connector or cable to buy.

All the cables or connectors need to be put as optional for further purchase

Q4

Warranty: there is no mention of warranty; what is the warranty requirement?

A4

Warranty: The standard of 1 year

Q5

Is this a new procurement or is it replacing existing simulator(s)? If so, please specify model(s) and manufacturer(s)?

A5

This procurement aim to replace and add more service maintenance tools for the Biomedical Technologist. We are using LionHeart MP51, Fluke MPS450 and others.

Q6

Can you specify how, where and who will be using the simulators?

A6

Patient simulators will be used by the Biomedical Technologist as service maintenance tool for different medical devices as defibrillators, vital monitors, ecg recorders and many others.

Q7

Are the requirements set for a specific part task trainer to perform unique training tasks or is it for a high fidelity human patient simulator? Can you provide additional information including training needs?

A7

This simulator are high fidelity human patient simulator, but the provider need to offer multiple optional connectors to connect to the simulator. It not required training for this device

Q8

Could you provide specifications for training, installation and/or support & warranty for the simulators?

A8

Is not require a training, neither an installation although is mandatory a demo-presentation showing the benefits of such device. The warranty will be one year.

The fabricant must be able to provide calibration service or eventual repairs of such devices.

Q9

What is the anticipated level of expenditure (budget) for this project?

A9

The request is for one to six devices during 2013, with option to buy 6 more in the next three years.

Delete

9.2 Temperature: from 0 °C to 40 °C

Insert

9.2 Temperature: 35 °C to 42 °C

All other terms and conditions remain unchanged