

**RETURN BIDS TO:**  
**RETOURNER LES SOUMISSIONS À:**  
**Bid Receiving - PWGSC / Réception des soumissions -**  
**TPSGC**  
**11 Laurier St. / 11, rue Laurier**  
**Place du Portage, Phase III**  
**Core 0A1 / Noyau 0A1**  
**Gatineau, Québec K1A 0S5**  
**Bid Fax: (819) 997-9776**

**SOLICITATION AMENDMENT**  
**MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**  
**Raison sociale et adresse du**  
**fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Electrical & Electronics Products Division  
11 Laurier St./11, rue Laurier  
6B1, Place du Portage, Phase III  
Gatineau, Québec K1A 0S5

|   |   |
|---|---|
| <b>Title - Sujet</b><br>NETWORK ANALYZER  |   |
| <b>Solicitation No. - N° de l'invitation</b><br>W6369-130066/A  | <b>Amendment No. - N° modif.</b><br>001                                   |
| <b>Client Reference No. - N° de référence du client</b><br>W6369-130066   | <b>Date</b><br>2013-01-09   |
| <b>GETS Reference No. - N° de référence de SEAG</b><br>PW-\$\$HN-313-61817  |   |
| <b>File No. - N° de dossier</b><br>hn313.W6369-130066   | <b>CCC No./N° CCC - FMS No./N° VME</b>                                    |
| <b>Solicitation Closes - L'invitation prend fin</b><br><b>at - à 02:00 PM</b><br><b>on - le 2013-01-29</b>  | <b>Time Zone</b><br><b>Fuseau horaire</b><br>Eastern Standard Time<br>EST |
| <b>F.O.B. - F.A.B.</b><br><b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/> |   |
| <b>Address Enquiries to: - Adresser toutes questions à:</b><br>Kelly, Ruth  | <b>Buyer Id - Id de l'acheteur</b><br>hn313                               |
| <b>Telephone No. - N° de téléphone</b><br>(819) 956-3588 ( )  | <b>FAX No. - N° de FAX</b><br>( ) -                                       |
| <b>Destination - of Goods, Services, and Construction:</b><br><b>Destination - des biens, services et construction:</b>   |   |

**Instructions: See Herein**

**Instructions: Voir aux présentes**

|  |  |
|--|--|
| <b>Delivery Required - Livraison exigée</b>  | <b>Delivery Offered - Livraison proposée</b> |
| <b>Vendor/Firm Name and Address</b><br><b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>   |  |
| <b>Telephone No. - N° de téléphone</b><br><b>Facsimile No. - N° de télécopieur</b>   |  |
| <b>Name and title of person authorized to sign on behalf of Vendor/Firm</b><br><b>(type or print)</b><br><b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b><br><b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b> |  |
| <b>Signature</b>   | <b>Date</b>                                  |

---

**Amendment No. 001 is raised to provide answers to the following questions:**

**Question 1:**

The RFP indicates that a High Stability Timebase is required. Would a solution that offers a typical total stability of +/- 3 ppm be acceptable?

**Answer 1:**

A total stability of +/-3 ppm would be acceptable.

**Question 2:**

The RFP indicates that a 50 ohm Electronic Calibration Kit (ECal module) covering the frequency range of 300 kHz to 9 GHz is required. However, the network analyzer requirement for frequency range is limited to 3 GHz. Would a 50 ohm Electronic Calibration Kit (ECal module) covering the frequency range of 300 kHz to 3 GHz be acceptable?

**Answer 2:**

A 50 ohm Electronic Calibration Kit (ECal module) covering the frequency range of 300 KHz to 3 GHz would be acceptable.

**Question 3:**

The RFP indicates that a 75 ohm Electronic Calibration Kit (ECal module) is required. Such 75 ohm modules are going or have gone obsolete from most vendors. Would the 75 ohm ECal module simply be a desirable and not mandatory?

**Answer 3:**

A 75 ohm Electronic Calibration Kit (ECal module) is required (mandatory) since the the item will be used most of the time with 75 ohm systems.

**All other terms and conditions of the RFP remain unchanged.**