

HEALTH CANADA
1001 SAINT-LAURENT, LONGUEUIL, QUEBEC
ULTRAPURE WATER PRODUCTION SYSTEM MAINTENANCE SCHEDULE

1001 SAINT-LAURENT, LONGUEUIL, QC		First Base Year First Option Year												Second Base Year Second Option Year												Third Base Year Third Option Year											
		July	August	September	October	November	December	January	February	March	April	May	June	July	August	September	October	November	December	January	February	March	April	May	June	July	August	September	October	November	December	January	February	March	April	May	June
	MAINTENANCE WORK (Part A)																																				
Aa	REPLACEMENT OF REVERSE OSMOSIS AND ULTRAFILTRATION MEMBRANE FILTERS (at the outset of the contract and as required thereafter)	X																																			
Ab	CLEANING OF THE REVERSE OSMOSIS MEMBRANE FILTER (every month)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Ac	REPLACEMENT OF THE NUCLEAR GRADE RESIN FILTER MEDIUM (every 3 months)	X			X			X			X			X			X			X			X			X			X			X					
Ad	REPLACEMENT OF THE UV LAMPS (every 9 months)	X									X						X											X									
Ae	DISINFECTION OF THE TREATED WATER STORAGE TANKS AND THE PURITY METER AND CONDUCTIVITY METER PROBES (every 12 months)	X												X												X											
Af	REPLACEMENT OF THE ACTIVATED CARBON FILTER MEDIUM (every 24 months)	X																								X											
	REPAIR AND REPLACEMENT WORK (Part B)																																				
Ba	REPAIR AND REPLACEMENT WORK: MATERIALS (as required for 36 months)																																				
Bb	REPAIR AND REPLACEMENT WORK: LABOUR (as required for 36 months)																																				
	SPECIAL SERVICE, As Required (Part B)																																				
Bc	LABOUR FOR THE DESINFECTION OF 60 TAPS FOR PURE WATER (as required every 12 months); related work and co-ordination at point Ae	X												X												X											