

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving - PWGSC / Réception des soumissions
- TPSGC
11 Laurier St./ 11 rue, Laurier
Place du Portage, Phase III
Core 0A1 / Noyau 0A1
Gatineau, Québec K1A 0S5
Bid Fax: (819) 997-9776

SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Scientific, Medical and Photographic Division /
Division de l'équipement scientifique, des produits
photographiques et pharmaceutiques
11 Laurier St./ 11 rue, Laurier
6B1, Place du Portage
Gatineau, Québec K1A 0S5

Title - Sujet Reach-In Growth Chamber	
Solicitation No. - N° de l'invitation K8A70-139001/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client K8A70-139001	Date 2013-06-25
GETS Reference No. - N° de référence de SEAG PW-\$\$PV-938-62789	
File No. - N° de dossier pv938.K8A70-139001	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2013-07-02	Time Zone Fuseau horaire Eastern Daylight Saving Time EDT
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Roy, Paul	Buyer Id - Id de l'acheteur pv938
Telephone No. - N° de téléphone (819) 956-6919 ()	FAX No. - N° de FAX (819) 956-3814
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment is raised in order to post the questions and answers received during the site visit and following the site visit.

Q.1. Please confirm that we are to provide an enclosure from the chamber to the ceiling to isolate the sound from the compressor?

A.1. Yes

Q.2. Please confirm that the customer will provide an Ethernet connection?

A.2. Yes

Q.3. Please confirm that the customer will disconnect the existing electrical and chilled water from the current chamber?

A.3. Yes

Q.4. Please confirm we are responsible for connecting power to the new chamber?

A.4. Yes

Q.5. Please confirm we are responsible for the refrigeration piping, electrical, and mounting the new condenser?

A.5. Yes

Specification questions

Q.6. Regarding interior dimension specifications: Is a 95½" wide by 29¾" deep interior acceptable?

A.6 Yes

Q.7. Regarding exterior height: Is a chamber that is higher than 96" but less than 102" acceptable?

A.7 Yes

Q.8. Regarding Door: With reduced height growth chamber, door height has to be reduced so Is a 52" high door opening acceptable?

A.8 Yes

Q.9. Regarding growth height per tier: Is 18" of growth height on the bottom two tiers and 15" of growth height on the top tier acceptable in three tier mode?

A.9 Yes

Q.10. Regarding humidity control system: Is 90% RH lights off and 75% RH lights on acceptable?

A.10 Yes

Q.11. Regarding humidity control system: Is a +/-3% RH accuracy acceptable?

A.11 Yes

Q.12. Regarding growth area: Is 18ft² of growth area per tier acceptable?

A.12 Yes

Q.13. Regarding light intensity, Are higher light intensities required in two tier mode (ex. 750µmoles/m²/s per tier) and single tier mode (ex. 1,100µmoles/m²/s)? If yes please indicate the required light intensities?

A.13 Our requirements are always 300 ± per tier at soil level (approximately 6 inches from bottom of tier)

Solicitation No. - N° de l'invitation

K8A70-139001/A

Amd. No. - N° de la modif.

001

Buyer ID - Id de l'acheteur

pv938

Client Ref. No. - N° de réf. du client

K8A70-139001

File No. - N° du dossier

pv938K8A70-139001

CCC No./N° CCC - FMS No/ N° VME

Q.14. Regarding humidity control system: Is a spray nozzle system an acceptable method for the humidification system?

A.14 Yes

Q.15. Is the the forklift available for and unloading?

A.15. yes Property Management will provide and operate forklift but would prefer a representative their while equipment is being received.

All other terms and conditions remain the same