

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
PO Box 1408, Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3C 2Z1
Bid Fax: (204) 983-0338

SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
**Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution
Public Works and Government Services Canada -
Western Region
PO Box 1408, Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3C 2Z1

Title - Sujet Cage and Rack Washer	
Solicitation No. - N° de l'invitation 6D063-132061/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client 6D063-132061	Date 2013-08-20
GETS Reference No. - N° de référence de SEAG PW-\$WPG-202-8592	
File No. - N° de dossier WPG-3-36039 (202)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2013-09-09	Time Zone Fuseau horaire Central Daylight Saving Time CDT
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Kozak, Tammy	Buyer Id - Id de l'acheteur wpg202
Telephone No. - N° de téléphone (204) 984-8825 ()	FAX No. - N° de FAX (204) 983-7796
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This Amendment #002 is raised to address bidder enquiries as follows:

Q1. We require additional details for #1.25 of the glassware washer such as details about kind, dimensions and shapes of the glassware to be treated as well as the expected throughput in order to quote racks that will be useful and efficient.

A1. The site visit will provide a visual understanding of PHAC needs. Bidders will have the opportunity to take their own measurements of the items at that time.

Q2. We need more details about the dimensions of the cages, and bottles and the expected throughput for 2.4.6 for the Cage and Rack Washer.

Q2. The site visit will provide a visual understanding of PHAC needs. Bidders will have the opportunity to take their own measurements of the items at that time.

Q3. If the machine is pass-through we can't understand what are the loading and unloading sides, as there 2 walls in the corner where the machine should be installed.

Q3. Clarification will be provided at site visit.

Q4. Do you need to only treat the drain temperature or the water PH as well?

A4. Temperature of drain water cannot exceed 75 degrees Centigrade. Water PH must be between 6 and 8.

Q5. Is it possible to use cold water to cool down the drain temperature or is a chilled water circuit preferred?

A5. Cold water cooling is acceptable

Q6. What is the security level required for the bid?

A6. Refer to Part 6, article 1.1 Security Requirement, page 9 of 15 of the RFP.

REMINDER TO BIDDERS: THIS REQUIREMENT INCLUDES A MANDATORY SITE VISIT.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.