

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
PO Box 1408, Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3C 2Z1
Bid Fax: (204) 983-0338

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
**Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution
Public Works and Government Services Canada -
Western Region
PO Box 1408, Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3C 2Z1

Title - Sujet Flow Particle Image Analyzer	
Solicitation No. - N° de l'invitation 23295-140080/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client NRCAN	Date 2013-11-22
GETS Reference No. - N° de référence de SEAG PW-\$WPG-211-8725	
File No. - N° de dossier WPG-3-36089 (211)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2013-12-16	Time Zone Fuseau horaire Central Standard Time CST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Wiebe, Amanda	Buyer Id - Id de l'acheteur wpg211
Telephone No. - N° de téléphone (204) 983-7032 ()	FAX No. - N° de FAX (204) 983-7796
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment No. 001 is raised to modify solicitation 23295-140080/A to address questions from industry as follows:

Question:

Item 2 Particle Orientation: particles must orient in a way that places the largest area of the particle towards the Charge-Couple Device (CCD) camera.

The Product that I am looking at does not orient the particles toward the camera but is able to measure the area of particles as per the specifications

Answer from Technical Authority:

We are looking for an equipment to show us the maximum surface area of a particle not a random one. Thus, we do not accept this offer.

Question:

Item 3 – Particle Identification: Must be able to identify the particles in the range from 0.8 µm to 300 µm

The product I am looking at has a size range is 1µm to 600µm, would the customer be willing to consider this?

Answer from Technical Authority:

Yes, this is acceptable, however a size range commencing at any value above 1 µm will not be acceptable.

Refer to Annex A1 - Compliance Matrix at Item 3. Particle Identification:

DELETE:

...0.8 µm to 300 µm

INSERT:

...1 µm to 300 µm

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.