

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
B3J 1T3
Bid Fax: (902) 496-5016

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Acquisitions
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
B3J 3C9

Title - Sujet SUPPLY AND INSTALL AUTOCLAVE	
Solicitation No. - N° de l'invitation 01804-140373/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client 01804-14-0373	Date 2014-01-13
GETS Reference No. - N° de référence de SEAG PW-\$HAL-123-5045	
File No. - N° de dossier HAL-3-70094 (123)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2014-01-21	Time Zone Fuseau horaire Atlantic Standard Time AST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Lockyer (HAL), Jeff	Buyer Id - Id de l'acheteur hal123
Telephone No. - N° de téléphone (902) 496-5636 ()	FAX No. - N° de FAX (902) 496-5016
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment 001 is being created for the following:

In order to help us to provide a proposal, could you please clarify:

Existing installation

Q. Could you provide a few pictures of the current installation?

A. See attached images

Q. Please confirm make & model of the autoclave to be replaced.

A. The following autoclave will replace a Steris Amsco Century SV-120 Scientific Prevacuum Sterlizer

Please confirm availability of cold water supply:

- 10-21 degC **yes**
- 1" NPT line **1/2" inch**
- 30-50 psig dynamic pressure **yes**
- 6 GPM peak flow **yes**

Please confirm details concerning the RO water supply line from the building:

- 1/2" NPT line **yes**
- 20-50 psig dynamic pressure **yes**
- 1 GPM peak flow. 4 gallon per cycle **yes**
- Material of the supply line: plastic or stainless steel? **Plastic**

Electric services: Please confirm availability within 3 meters of the unit of:

- 120 V/1 ph/9.5 amps **yes**
- Disconnect switch for 120V **No**

-
- 208 V/3 phase, 83 amps **yes**
 - Disconnect switch for 208V, 3ph **yes**

Floor drain

- Presence of a floor drain able to accept peak 6 GPM **yes**

Location of services (utilities)

- Please confirm the services (cold water, RO water, electrical, drain) are available in the same room and will not require modifications to the building or passage of conduits within the ceiling or walls **Affirmative**

Please confirm that existing utilities (cold water, RO water) are equipped with:

- pressure gauges **No**
- isolation valves **Yes, not within the existing room, but located in the pipe chase.**

Delivery & installation by March 31st

Installation includes set-in-place, leveling, and assembly of cabinet panels and loading shelves.

- Please confirm the connections & startup are not required to be completed by March 31st 2014.

Yes the Autoclave must be working condition this would include connection & Startup.

***All other terms and conditions remain the same.**









