

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Public Works and Government Services Canada
Telus Plaza North/Plaza Telus Nord
10025 Jasper Ave./10025 ave. Jaspe
5th floor/5e étage
Edmonton
Alberta
T5J 1S6
Bid Fax: (780) 497-3510

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada
Telus Plaza North/Plaza Telus Nord
10025 Jasper Ave./10025 ave. Jasper
5th floor/5e étage
Edmonton
Alberta
T5J 1S6

Title - Sujet Lab Tables	
Solicitation No. - N° de l'invitation EW479-141857/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client PWGSC	Date 2014-01-28
GETS Reference No. - N° de référence de SEAG PW-\$EDM-023-10086	
File No. - N° de dossier EDM-3-36290 (023)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2014-02-19	Time Zone Fuseau horaire Mountain Standard Time MST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Taylor, Ian	Buyer Id - Id de l'acheteur edm023
Telephone No. - N° de téléphone (780) 497-3621 ()	FAX No. - N° de FAX (780) 497-3510
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This solicitation amendment has been raised to answer questions asked by potential bidders during the tendering period.

Questions:

Answers:

1. Would an acceptable method be to have a free standing booth style ventilation system with the table placed inside?

Answer: YES. Free standing is acceptable.

2. Or does the customer require a ceiling hung ventilation hood style unit so they can have access to load the table from all sides?.

Answer: NO. We do not need access from all sides.

3. Please specify the CFM (cubic feet per minute) required for the ventilation system on the Clean Room Bench

Answer: CFM of 740 will meet our needs.

If the answers affect your bid, you must forward an updated copy of your offer to the Bid Receiving Unit by closing date and time indicated on the front of the solicitation document.

If you have any further questions please forward them to the Contracting Authority in accordance with Part 2 - Bidder Instructions, Section 3 - Enquiries - Bid Solicitation contained within the solicitation document.

All other terms and conditions of the Request For Proposal remain unchanged.