



**RETURN BIDS TO:
RETOURNER LES SOUMISSIONS A :**

Procurement & Contracting Services
Bid Receiving Unit
VISITOR'S CENTRE
73 Leikin Drive
Ottawa, Ontario K1A 0R2

Services d'acquisitions et des marchés
Module de réception des soumissions
CENTRE DES VISITEURS
73 promenade Leikin
Ottawa (Ontario) K1A 0R2

QUESTIONS & ANSWERS

QUESTIONS & REPONSES

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments: - Commentaries :

Title – Sujet Forensic Comparison Microscope		Date June 3 rd , 2014
Solicitation No. – N° de l'invitation 201500259		Amendment No. – N° de la modification 002
Solicitation Closes – L'invitation prend fin		
At / à : 2:00 PM	EDT(Eastern Daylight Saving Time) HAE (heure avancée de l'Est)	
On / le : June 24th, 2014		
F.O.B. – F.A.B See herein — Voir aux présentes	GST – TPS See herein — Voir aux présentes	Duty – Droits See herein — Voir aux présentes
Destination of Goods and Services – Destinations des biens et services See herein — Voir aux présentes		
Instructions See herein — Voir aux présentes		
Address Inquiries to – Adresser toute demande de renseignements à Candice Therien		
Telephone No. – No. de téléphone 613-843-3826	Facsimile No. – No. de télécopieur 613-825-0082	
Vendor/Firm Name, Address and Representative – Raison sociale, adresse et représentant du fournisseur/de l'entrepreneur:		
Telephone No. – No. de téléphone	Facsimile No. – No. de télécopieur	
E-mail – Courriel		
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) – Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)		
Signature	Date	



This amendment is raised to address the following:

- To respond to questions received during the solicitation period

QUESTIONS AND ANSWERS

Question 1:

At Annex A section 3.7 - Interchangeable eyepieces with different magnifications. Are eyepieces with different magnifications to be included in the quote or is this just a specification that is required of the scope in the future? If you want different eyepieces can you please indicate which magnifications you would like? We would like to mention that the optimal way to increase magnification is to change the objective or to use the magnification changer. If you change the eyepiece magnification, this increase will not be seen on the camera.

Answer 1:

One (1) set of eyepieces per macroscope (10X magnification). Macroscope should have the capability of changing eyepieces if required in the future.

Question 2:

At Annex A, section 3.8 - Image capture system. Can you please provide more details on the camera and software?

Do you want a CCD or CMOS camera?

Which resolution of camera would you like, i.e. 2.5 MPixel, 5 MPixel or 8 MPixel?

Do you want measurement software included?

Answer 2:

CCD, 8 mega-pixel, **no** measurement software.

Question 3:

At Annex A, section 3.9 - Lighting systems - Can you please provide more details on the types of lights that you would like such as; fluorescent tubes, ring lights, line lights, spot lights with or without diffusers.

Answer 3:

LED spotlights, flexible fibre optic illuminator, fluorescent tube illuminators (any associated diffusers are required).

ALL TERMS AND CONDITIONS REMAIN THE SAME