

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
800 Burrard Street, Room 219
800, rue Burrard, pièce 219
Vancouver, BC V6Z 0B9
Bid Fax: (604) 775-7526

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada -
Pacific Region
219 - 800 Burrard Street
800, rue Burrard, pièce 219
Vancouver, BC V6Z 0B9

Title - Sujet Gas Chromatograph	
Solicitation No. - N° de l'invitation F1688-140012/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client F1688-140012	Date 2014-10-15
GETS Reference No. - N° de référence de SEAG PW-\$VAN-584-7335	
File No. - N° de dossier VAN-4-37121 (584)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2014-11-17	Time Zone Fuseau horaire Pacific Standard Time PST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Perez, Elizabeth	Buyer Id - Id de l'acheteur van584
Telephone No. - N° de téléphone (604) 775-7690 ()	FAX No. - N° de FAX (604) 775-7526
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation

F1688-140012/A

Client Ref. No. - N° de réf. du client

F1688-140012

Amd. No. - N° de la modif.

001

File No. - N° du dossier

VAN-4-37121

Buyer ID - Id de l'acheteur

van584

CCC No./N° CCC - FMS No/ N° VME

AMENDMENT 001

TITLE: Gas Chromatograph with Flame Ionization Detector

SEE ATTACHED.

AMENDMENT 001

This amendment is raised to present the following questions from bidders and the responses from the Project Authority, amend the Standard Instructions, Clauses and Conditions and the General Conditions of Resulting Contract Clauses.

- **Questions and Answers:**

Q1. 20mL vials are a bit odd for use with liquid injections, 2mL vials are the most common. I was wondering if this is typo or if it is an actual requirement. If it is an actual requirement please specify injection volume.

A1. The volume for the autosampler vials should be at least 2mL, not 20mL.

Q2. This solicitation calls for Fatty Acid analysis with both capillary and packed columns. Could you please clarify if two separate Flame Ionization Detectors are needed or just one?

A2. Only one FID is required.

Q3. Also, Fatty Acid analysis is typically done by capillary column so please clarify the need for packed column. Is there a specific analysis that requires packed columns?

A3. The requirement is for a gas chromatograph with the capability to use capillary columns for lipid and fatty acid analyses and with the capability to switch to packed columns for preparative and other types of analyses.

- **In PART 2 – BIDDERS INSTRUCTIONS**

Delete: Item 1. Standard Instructions, Clauses and Conditions

INSERT:

1. Standard Instructions, Clauses and Conditions

All instructions, clauses and conditions identified in the bid solicitation by number, date and title are set out in the Standard Acquisition Clauses and Conditions Manual (<https://buyandsell.gc.ca/policy-and-guidelines/standard-acquisition-clauses-and-conditions-manual>) issued by Public Works and Government Services Canada.

Bidders who submit a bid agree to be bound by the instructions, clauses and conditions of the bid solicitation and accept the clauses and conditions of the resulting contract.

The 2003 (**2014-09-25**) Standard Instructions - Goods or Services - Competitive Requirements, are incorporated by reference into and form part of the bid solicitation.

Subsection 5.4 of 2003, Standard Instructions - Goods or Services - Competitive Requirements, is amended as follows:

Delete: sixty (60) days

Insert: ninety (90) days

Solicitation No. - N° de l'invitation
F1688-140012/A
Client Ref. No. - N° de réf. du client
F1688-140012

Amd. No. - N° de la modif.
File No. - N° du dossier
VAN-4-37121

Buyer ID - Id de l'acheteur
van584
CCC No./N° CCC - FMS No/ N° VME

- **In PART 6 – RESULTING CONTRACT CLAUSES**

Delete: **Item 3.1 General Conditions**

INSERT:

3.1 General Conditions

2010A (2014-09-25), General Conditions - Goods (Medium Complexity), apply to and form part of the Contract.

- **ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED.**