

**RETURN BIDS TO:**  
**RETOURNER LES SOUMISSIONS À:**  
Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3  
Bid Fax: (613) 545-8067

## SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address  
Raison sociale et adresse du  
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution  
Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3

<b>Title - Sujet</b> Stereo PIV System	
<b>Solicitation No. - N° de l'invitation</b> W0114-155185/A	<b>Amendment No. - N° modif.</b> 002
<b>Client Reference No. - N° de référence du client</b> W0114-15-5185	<b>Date</b> 2014-11-26
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$KIN-610-6482	
<b>File No. - N° de dossier</b> KIN-4-42156 (610)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2015-01-05</b>	<b>Time Zone</b> Fuseau horaire Eastern Standard Time EST
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Correia-Reid, Vincent	<b>Buyer Id - Id de l'acheteur</b> kin610
<b>Telephone No. - N° de téléphone</b> (613) 545-8738 ( )	<b>FAX No. - N° de FAX</b> (613) 545-8067
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

Instructions: See Herein

Instructions: Voir aux présentes

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm (type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

**AMENDMENT 002**

This Amendment 002 is raised to respond to questions and modify Annex "A" - Requirement in the RFP.

**Questions and Answers**

Q1: Do you truly need more than 8 output channels and 2 input channels for the high speed controller.

A1: Yes; in anticipation of upgrading at later time to tomographic PIV, we require more than 8 channels.

Q2: Please confirm that for the high speed cameras, the specification should read " 1 M pixel at 1600 frames/sec at full resolution".

A2: Please see the change below.

**Changes**

Under Annex "A" - Requirement

DELETE:

Two High Speed Cameras at least 1G pixel resolution sensor, high frame rate (1600 frames/sec or higher at full resolution). At least 6 GB RAM (preferably upgradable).

And INSERT the following:

Two High Speed Cameras at least 1M pixel resolution sensor, high frame rate (1630 frames/sec or higher at full resolution). At least 6 GB RAM (preferably upgradable).

**ALL OTHER TERMS AND CONDITIONS OF THE RFP REMAIN UNCHANGED**