

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving - PWGSC / Réception des soumissions
- TPSGC
11 Laurier St. / 11, rue Laurier
Place du Portage, Phase III
Core 0A1 / Noyau 0A1
Gatineau, Québec K1A 0S5
Bid Fax: (819) 997-9776

SOLICITATION AMENDMENT

MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Vehicles & Industrial Products Division
11 Laurier St./11, rue Laurier
7A2, Place du Portage, Phase III
Gatineau, Québec K1A 0S5

Title - Sujet Metering Pump	
Solicitation No. - N° de l'invitation 23240-150237/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client 23240-150237	Date 2014-12-02
GETS Reference No. - N° de référence de SEAG PW-\$\$HP-404-66107	
File No. - N° de dossier hp404.23240-150237	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2014-12-22	Time Zone Fuseau horaire Eastern Standard Time EST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Michele Mak	Buyer Id - Id de l'acheteur hp404
Telephone No. - N° de téléphone (819) 956-3947 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation

23240-150237/A

Amd. No. - N° de la modif.

001

Buyer ID - Id de l'acheteur

hp404

Client Ref. No. - N° de réf. du client

23240-150237

File No. - N° du dossier

hp40423240-150237

CCC No./N° CCC - FMS No/ N° VME

This page has been intentionally left blank

Amendment #001 to Solicitation number 23240-150237A

Question(s) & Answers:

Q1. What fluid are the items pumping?

A1. Mildly corrosive and non corrosive reagents

Q2. Can we include FKM for a material in contact with the fluid?

A2. Yes

Q3. Can we provide an English manual with only part in French?

A3. English with Partial French YES

Q4. An analog signal control cable/connector would be required for operation of the pump via 4-20 mA, and will need to be quoted as a second line item.

A4. Please provide bids for pump only. However pump MUST have the capability of being controlled both manually and through 4- 20mA for future applications.

Q5. The requested item (GALA1008PVT126OUD112100) is now obsolete

A5. Product number provided is for reference only - equivalent model accepted.

Q6. Fluid: Ensure fluid used is compatible with wetted materials quoted.

A6. The fluids being pumped are : mildly corrosive and non corrosive reagents. The parts of the pump in contact with these fluids must be compatible (i.e. the wetted parts are usually PTFE,FKM)

Add Clause: A1009C Instructions to Bidders / Contractors (2008-05-12)

Authorized representatives of Canada must have access to any site where any part of the Work is being carried out at any time during working hours to make examinations and such tests of the Work as they may think fit.

All other terms and conditions remain unchanged.