

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3
Bid Fax: (613) 545-8067

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Title - Sujet Mobile X-Ray Unit	
Solicitation No. - N° de l'invitation 21401-146159/A	Amendment No. - N° modif. 003
Client Reference No. - N° de référence du client 21401-14-6159	Date 2014-12-23
GETS Reference No. - N° de référence de SEAG PW-\$KIN-620-6489	
File No. - N° de dossier KIN-4-42146 (620)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2015-01-12	
Time Zone Fuseau horaire Eastern Standard Time EST	
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Porter, Marta M.	Buyer Id - Id de l'acheteur kin620
Telephone No. - N° de téléphone (613) 547-7587 ()	FAX No. - N° de FAX (613) 545-8067
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation

21401-146159/A

Amd. No. - N° de la modif.

003

Buyer ID - Id de l'acheteur

kin620

Client Ref. No. - N° de réf. du client

21401-14-6159

File No. - N° du dossier

KIN-4-42146

CCC No./N° CCC - FMS No/ N° VME

See Attached

Amendment 003 – Mobile X-Ray Unit, is being raised for the following reason:

- 1. Provide a response back from Public Works and Government Services Canada to questions received from bidders.**
-

Q1. Please find attached an Addendum that was issued for Solicitation No. 21501-145469/A. This is in regards to the technical specifications.

Would you be able to apply this to this current RFP?

A1. Please see amendment to Annex "A" – Requirement below.

Refer: Annex "A" – Requirement, Mobile X-Ray Unit, 1 c, d, e, f, k

Delete: In its Entirety

Insert: Annex "A" – Requirement, Mobile X-Ray Unit, 1 c, d, e, f, k

1c) Kilovolts peak (kVp) range in 1 kVp steps minimum range of **50** to 125 kVp

1d) Milliamp (mA) Stations minimum range of 10 to **300** mA

1e) Exposure Time Range **0.004** - 8 seconds. Cannot be longer than 8 seconds exposure time range –

1f) Milliampere second (mAs) Range minimum range of **0.2** to 500 mAs

1k) XraY tubing (rotating anode) Focal spots **.6/1.0**kVp 140/150 Target angle 12 degrees

ALL OTHER TERMS AND CONDITIONS OF THIS SOLICITATION REMAIN UNCHANGED.

If your tender has already been forwarded and you wish to revise same, this revision should be faxed and reach the bid receiving area before the closing date. The tender number and the closing date are to be shown on the first page of your fax.