

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
PWGSC/TPSGC Acquisitions
1045 Main Street
1st Floor, Lobby C
Unit 108
Moncton, NB E1C 1H1
Bid Fax: (506) 851-6759

SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
NB / PEI Division - Moncton Acquisitions Office
1045 Main Street
1st Floor, Lobby C
Unit 108
Moncton, NB E1C 1H1

Title - Sujet Autoclave	
Solicitation No. - N° de l'invitation 01913-150378/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client 01913-150378	Date 2015-01-05
GETS Reference No. - N° de référence de SEAG PW-\$MCT-014-4927	
File No. - N° de dossier MCT-4-37096 (014)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2015-01-20	
Time Zone Fuseau horaire Atlantic Standard Time AST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: MacDonald, Charline	Buyer Id - Id de l'acheteur mct014
Telephone No. - N° de téléphone (506) 851-6067 ()	FAX No. - N° de FAX (506) 851-6759
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Title: Autoclave**Solicitation Amendment No. 001**

This solicitation is hereby amended to provide the following questions and answers:

Q1. Reference: Technical Specification 5

Please clarify the requirement for “Two full racks”; should we understand two full shelves?

A1. Yes, it should be two full shelves.

Q2. Reference: Technical Specification 6

“Bottom of chamber” is lower than the loading height. Should we understand “Bottom of chamber” or “Loading Height”?

A2. Both. As in: “Bottom of chamber is lower than the loading height”.

Q3. Reference: Technical Specification 8

Does “Extendable” mean shelves that can slide out? Please clarify.

A3. Yes, the shelves can slide out.

Q4. Reference: Technical Specifications 10 and 14

10 ask for Cabinet panels, but 14 ask for a recessed mounting. When “recessed”, the unit is not provided with panels. Please clarify if panels are needed or not.

A4. Panels are not essential.

Q5. Reference: Technical Specification 23

Micro Processors are considered as proprietary controls, and usually customers will prefer a non-proprietary control, such as an industrial type PLC (Allen Bradley, per example). The PLC allows 20 cycles, and competes programming flexibility. It also provides a superior control technology, and is not proprietary to the manufacturer of the sterilizer – allowing flexibility for supply of control components in the future. Please confirm that non-proprietary PLC control is accepted, as a value added control technology.

A5. Non-proprietary PLC control will be accepted.

Q6. Reference: Technical Specification 30

Our Gravity cycle (for solids) has a factory setup dry time of 5 minutes; however, this is fully adjustable to the requested 30 minutes, and will be programmed as per the owner request at start-up.

Please confirm if this is acceptable.

A6. This is acceptable.

Q7. Reference: Technical Specification 35

Please clarify this requirement: If door fails, requirement should be to automatically stop steam to the jacket & chamber. A baffle also helps to redirect the steam away from where the user may be working, and evenly distributes the steam into chamber and avoid overheat spots and excessive condensation. Should we understand that the requirement is to provide an interior steam baffle?

A7. We are looking for a safety feature that would ensure a manual override if required. If an interior steam baffle is a mechanism that is part of the system, then that is fine.

Q8. Reference: Technical Specification 42

Do we have to include an RO generating equipment in our proposal?

A8. Yes.

Q9. Reference: Technical Specification 42

Assuming RO is provided by the building to feed the steam generator, please confirm what conductivity (or resistivity) is this RO water.

A9. The RO water for the steam generator will be supplied by the new RO system that is specified in the tender. No existing building RO water will supply the steam generator

Q10. Reference: Technical Specification 43

We understand that the supply must coordinate the installation with the owner; please confirm the connections to utilities (water, drain, and electricity) will be done by the owner.

A10. All connections will be completed by the contractor. There will be no work completed by the owner.

If your bid has already been forwarded and you wish to revise same, this revision should be sent either in a sealed envelope and mailed to the above address or by facsimile (506) 851-6759 and reach the undersigned before the appropriate closing date. The solicitation number and the closing date are to be shown on the outside of the sealed envelope or on the facsimile transmission.

All other terms and conditions of the solicitation document remain unchanged.

All enquiries concerning this amendment are to be forwarded to:

Name: Charline MacDonald
Telephone No.: (506) 851-6067
Facsimile No: (506) 851-6759