

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
**Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada**
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
B3J 1T3
Bid Fax: (902) 496-5016

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
**Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution
Atlantic Region Acquisitions/Région de l'Atlantique
Acquisitions
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
B3J 3C9
Nova Scot

Title - Sujet Steam Sterilizer	
Solicitation No. - N° de l'invitation F5244-140185/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client F5244-14-0185	Date 2015-02-12
GETS Reference No. - N° de référence de SEAG PW-\$HAL-409-9425	
File No. - N° de dossier HAL-4-73162 (409)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2015-02-17	Time Zone Fuseau horaire Atlantic Standard Time AST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Taylor, Kathie	Buyer Id - Id de l'acheteur hal409
Telephone No. - N° de téléphone (902) 496-5510 ()	FAX No. - N° de FAX (902) 496-5016
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation

F5244-140185/A

Amd. No. - N° de la modif.

002

Buyer ID - Id de l'acheteur

ha1409

Client Ref. No. - N° de réf. du client

F5244-14-0185

File No. - N° du dossier

HAL-4-73162

CCC No./N° CCC - FMS No/ N° VME

Solicitation Amendment #02
See Attached

Solicitation No. - N° de l'invitation
F5244-140185/A
Client Ref. No. - N° de réf. du client
F5244-14-0185

Amd. No. - N° de la modif.
002
File No. - N° du dossier
HAL-4-73162

Buyer ID - Id de l'acheteur
hal409
CCC No./N° CCC - FMS No./N° VME

Solicitation Amendment #02

I. The following questions were posed by industry:

Q1: Could you please confirm what the height these carboys will be – picture or catalog drawings would be also helpful?

R1: One example of carboys used in the autoclave is in the photo. The carboy height just over 20" and less than 21" without the cap.



Solicitation No. - N° de l'invitation
F5244-140185/A
Client Ref. No. - N° de réf. du client
F5244-14-0185

Amd. No. - N° de la modif.
002
File No. - N° du dossier
HAL-4-73162

Buyer ID - Id de l'acheteur
hal409
CCC No./N° CCC - FMS No./N° VME

Q2. As far as we are informed, Automatic Blowdown does not meet the plumbing code (refer to recent RFQ 01913-150378/A for PEI similar autoclave).

R2: No, we do not have the authority to release the supplier from the responsibility of deviation to the plumbing code. The requirement for the building where the autoclave is to be installed is that boiling water cannot enter the building drainage system. Only warm water is allowed to enter the drains. The boiling water exiting the steam sterilizer must be modified so that it is warm water being released into the drains.

In answering the previous question posed on January 22 (Question 2), the response indicated that the drain can be modified to not send boiling water down into the plumbing as is the case for other units in the building and this response upon further clarification and information was incorrect. **The drain is not and cannot be modified.**

Q3: If mandatory Auto Blowdown request is maintained, please confirm if we have to add the CRN stamped condenser receiving blowdown tank, to release pressure and cooldown condensate?

R3: If a condenser receiving blowdown tank is required for the autoblowdown to release warm water into the plumbing system, then yes.

All other terms and conditions remain unchanged.