

**RETURN BIDS TO:**  
**RETOURNER LES SOUMISSIONS À:**  
Bid Receiving Public Works and Government  
Services Canada/Réception des soumissions Travaux  
publics et Services gouvernementaux Canada  
Room 1650, 635 8th Ave. S.W.  
Calgary  
Alberta  
T2P 3M3

**SOLICITATION AMENDMENT**  
**MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**  
**Raison sociale et adresse du**  
**fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Public Works and Government Services  
Canada/Travaux publics et Services gouvernementaux  
Canada  
Room 1650, 635 8th Ave. S.W.  
Calgary  
Alberta  
T2P 3M3

<b>Title - Sujet</b> Portable Network Vector Analyzer		
<b>Solicitation No. - N° de l'invitation</b> W7702-155721/A		<b>Amendment No. - N° modif.</b> 003
<b>Client Reference No. - N° de référence du client</b> W7702-155721		<b>Date</b> 2015-02-19
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$CAL-135-6313		
<b>File No. - N° de dossier</b> CAL-4-37070 (135)		<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2015-03-09</b>		<b>Time Zone</b> Fuseau horaire Mountain Daylight Saving Time MDT
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>		
<b>Address Enquiries to: - Adresser toutes questions à:</b> Denadi, Gilies		<b>Buyer Id - Id de l'acheteur</b> cal135
<b>Telephone No. - N° de téléphone</b> (403) 612-9648 ( )		<b>FAX No. - N° de FAX</b> ( ) -
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b> DEPT OF NATIONAL DEFENCE DEFENCE R&D CANADA - SUFFIELD Building 560 Receiving Ralston, AB, T0J 2N0 CANADA		

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

Solicitation No. - N° de l'invitation

W7702-155721/A

Amd. No. - N° de la modif.

003

Buyer ID - Id de l'acheteur

cal135

Client Ref. No. - N° de réf. du client

W7702-155721

File No. - N° du dossier

CAL-4-37070

CCC No./N° CCC - FMS No/ N° VME

---

This page has been left blank intentionally.

Solicitation No. - N° de l'invitation  
W7702-155721/A  
Client Ref. No. - N° de réf. du client  
W7702-155721

Amd. No. - N° de la modif.  
003  
File No. - N° du dossier  
CAL-4-37070

Buyer ID - Id de l'acheteur  
cal135  
CCC No./N° CCC - FMS No./N° VME

---

This amendment 003 is being raised to correct an administrative error in amendment 002 and to incorporate the following information into Solicitation W7702-155721/A:

**Questions and Answers:**

Q1: What applications require measurements to 14GHz?

A1: Experimental test requirements have moved up to and including 14 GHz.

Q2: Why is pulse and power required for cable & antenna measurements? Is there radar pulses measurements?

A2: Yes, measurements required will include but is not limited to radar applications.

Q3: The pulse power measurements call for an external peak power sensor, do you then need the external sensor with that? What frequency and power range is required and what pulse width?

A3: The option for peak power measurements is in the requirement. i.e. the Portable Vector Network Analyzer must have the option available/installed for pulse measurements. The external pulse sensor is application specific and is not required in this acquisition.

Q4: What is the rise time, and pulse repetition rate needed?

A4: N/A

Q5: Would you require calibration kit, and cables (what length)?

A5: Calibration kit not required for this acquisition.

**All other terms and conditions remain the same.**