

**RETURN BIDS TO:**  
**RETOURNER LES SOUMISSIONS À:**  
Bid Receiving Public Works and Government  
Services Canada/Réception des soumissions Travaux  
publics et Services gouvernementaux Canada  
800 Burrard Street, Room 219  
800, rue Burrard, pièce 219  
Vancouver, BC V6Z 0B9  
Bid Fax: (604) 775-7526

## SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

### Comments - Commentaires

Vendor/Firm Name and Address  
Raison sociale et adresse du  
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution  
Public Works and Government Services Canada -  
Pacific Region  
219 - 800 Burrard Street  
800, rue Burrard, pièce 219  
Vancouver, BC V6Z 0B9

<b>Title - Sujet</b> DATA ACQUISITION SYSTEM	
<b>Solicitation No. - N° de l'invitation</b> 31019-143885/A	<b>Amendment No. - N° modif.</b> 002
<b>Client Reference No. - N° de référence du client</b> 31019-143885	<b>Date</b> 2015-02-27
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$VAN-524-7439	
<b>File No. - N° de dossier</b> VAN-4-37287 (524)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2015-03-09</b>	<b>Time Zone</b> Fuseau horaire Pacific Daylight Saving Time PDT
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Yamamoto, Albert	<b>Buyer Id - Id de l'acheteur</b> van524
<b>Telephone No. - N° de téléphone</b> (604) 775-7549 ( )	<b>FAX No. - N° de FAX</b> (604) 775-7526
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

Instructions: See Herein

Instructions: Voir aux présentes

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> Raison sociale et adresse du fournisseur/de l'entrepreneur	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm (type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

## Questions and Answers:

Q1. What is the level of vibration that is to be measured?

A1 : Peak  $\pm$  100g.

Q2. Of the 6 channels how many will monitor vibration and how many will monitor sound?

A2 : Typically 3 channels for vibration and 3 channels for acoustics.

Q3. Is the system to be AC Mains powered with battery backup or ONLY DC powered?

A3 : System must be able to work in both situations: with AC power in lab, and in remote area where AC power is not available.

Q4. Is the stored data to be only time-domain?

A4 : Yes.

Q5. What does 'compatible with existing PULSE Type 3160-A-042' mean ?

A5 : Compatible means both in hardware and software. The system must, as an extendable hardware, be able to work with the existing 4-channel hardware together to build a 10-channel test capability. The software need to be shared. The data must be consistent and comparable in processing.

Q6. Is there an IP rating specification for the enclosure?

A6 : No.