



Annex A

Statement of Work (SOW)

Service Agreement for Medical Device Inspection Program

1. OBJECTIVE

Correctional Service of Canada (CSC) intends to contract a qualified firm to provide expert inspection services for the medical equipment located at Correctional Service of Canada (CSC) federal correctional facilities in the Atlantic Region (Nova Scotia and New Brunswick).

This requirement involves the provision of technical support to Correctional Service of Canada (CSC) at the following sites:

- Dorchester Complex Minimum Sector, Dorchester (NB)
- Dorchester Complex Medium Sector, Dorchester (NB)
- Shepody Healing Center, Dorchester (NB)
- Atlantic Institution, Renous (NB)
- Springhill Institution, Springhill (NS)
- Nova Institution, Truro (NS)
- Regional Pharmacy, Moncton (NB)
- Regional Headquarters, Moncton (NB)

2. BACKGROUND

2.1 Project Background

The Health Services Sector of CSC has developed a Medical Device Inspection Program (MDIP) which outlines the processes and responsibilities for the effective inspection, upgrading, and replacement of medical devices and equipment used in CSC Health Services clinics, Hospitals, Pharmacies and Regional Head Quarters (RHQ).

The MDIP ensures that medical devices and equipment are safe and functional. It also provides guidance to identify and address potential problems with medical devices and equipment that may result in injury to staff or patients and requires a follow up on medical equipment or devices involved in an adverse event.



To support the MDIP, CSC requires the services of expert resources to conduct an annual inspection of all relevant medical equipment that falls within the inspection profile of the MDIP (hereinafter referred to as “Equipment”).

2.2 Objectives

As part of its mandate, Correctional Service Canada (CSC) has built and maintains a large number of secure sites with a diverse security profile. In order to provide the requisite health services to the inmate population within these sites, each Institution has a wide variety of medical systems and components that, in all aspects, constitutes a mission critical infrastructure that must be maintained.

To that end, CSC has a requirement to contract technical support for the inspection of the medical systems and associated equipment described in Annexes C, D, E, F, G, H, I, J.

3. SCOPE

3.1 Nature of Work

The work involves the provision of qualified resources to conduct the inspection of the Equipment identified by CSC in its federal correctional facilities (hereinafter referred to as “Institutions”). The Contractor must perform the Work, as described in this Statement of Work (SOW).

3.2 Nature of Services

The Contractor will provide technical support services designed and delivered to ensure the quality and effectiveness of CSC Equipment and associated components.

3.3 Institutional Medical Equipment

The medical devices (Equipment) within an Institution’s Health Services clinic can be broken down into two distinct categories:

1. Equipment that should be repaired when it ceases to function; and
2. Equipment that should be replaced when it ceases to function.



3.4 Equipment at End of Life

In all of the above cases, the useful service life of Equipment is established by the recommendations of the manufacturer and captured by Health Services in its Lifecycle Management Program. The Technician must assess the lifecycle status of each piece of Equipment based on the manufacturer's recommended service life as well as through the performance of a physical examination. If the Equipment is at or near the end of its useful service life, the Technician must note that in his or her report to the chief of health or project authority at RHQ.

4. INSPECTION PERSONNEL

4.1 General

The Contractor must perform this work using specialized resources with expertise in the inspection of medical equipment (hereinafter referred to as "Technicians").

The Contractor's Technicians will be responsible for:

1. Travelling to the required Institutions
2. Providing an annual report to the chief of health care, Regional Pharmacist or Manager of Clinical Services and the project authority at RHQ.

4.2 Technician Qualifications

Each technician performing the work must be an employee of the contract or an employee of a subcontractor.

Each Technician must have graduated from one of the following recognized programs:

- Biomedical Electronic Technologist or Technician program
- Biomedical Engineering Technologist or Technician program.
- Electronic Engineering Technologist or Technician program

Each technician must have at least three (3) years experience performing inspections on medical equipment.

4.3 Technician Certifications

Many manufacturers of medical equipment provide certification programs for the inspection and service of its product lines (including the inspection thereof). The Contractor must provide Technician(s) to complete the Work that have the



necessary certifications and training to inspect all of the Equipment appearing in Annexes C, D, E, F, G, H, I, J of this document.

It is reasonable that the Contractor may provide more than one Technician to an Institution at disparate times to ensure that all of the Equipment is maintained by someone with the proper certifications and training to perform that part of the Work.

5. TASKS AND ACTIVITIES

5.1 Test Equipment and Tools

The provision, inspection, management and control of all required test equipment will be the responsibility of the Contractor. A complete list of required testing equipment shall be presented to the chief of health prior to each site visit.

It should be noted that CSC will have final approval of all proposed test equipment before it is deployed on-site. This is due to a possible situation where some forms of equipment qualify as contraband within the precincts of an Institution and are therefore not permitted on-site.

5.2 Travel to CSC Institutions

The Contractor's Technician(s) must travel to Institutions identified by CSC as a facility that has Equipment that requires Inspection. The Contractor must arrange for the travel and accommodations for the Technician(s) and will remain entirely responsible for the travel requirements of its Technicians throughout the project.

The Contractor must ensure that the target Institution receives prior notification of the planned Inspection activities at least ten (10) working days prior to the site visit and the Technician has undergone a CPIC check in advance of that date.

The Technician will be accompanied by an individual assigned by the Institution to escort the Technician throughout the Institution. The Contractor must ensure that the Technician will abide by the instructions regarding the escort given by the Institutional Head or its delegate to the Technician.

6. CONSTRAINTS

The following constraints must be taken into account in the performance of the Work as described in this document:



6.1 Security Requirements

Any person entering a correctional Institution must undergo a clearance performed by the Canadian Police Information Centre (CPIC). This mandatory "CPIC Check" must be initiated at least two (2) weeks in advance of the scheduled date of entry into the Institution and completed without issue before the person is permitted to enter.

Delays caused by the CSC security clearance requirement will not incur any additional costs to CSC.

7. DELIVERABLES

The contractor shall provide the following services with regards to medical devices listed in Annexes C, D, E, F, G, H, I, J.

- a. Incoming on-site inspections to be conducted on an annual basis. Time of inspection will be determined after contract award between the Chief Health Services, Regional pharmacist or Manager of Clinical Services and the contractor. Each inspection shall include industry standard verifications for each piece of equipment listed in Annexes C, D, E, F, G, H, I, J.

Equipment that undergoes inspection must be tested to ensure it meets quality control standards as per the manufacturer's guidelines.

- b. Annual performance testing, inspection and calibration of electro medical devices is required in order to comply with standards set forth by the respective provinces.
- c. During inspection, assist in identifying requirements. This will be in regards to replacing existing equipment or purchasing new equipment.
- d. During inspection, repair all issues or problems with the equipment which do not require any parts.
- e. The report must be provided in writing **within 30 days of the site inspection**, and must detail all pieces of equipment inspected, any concerns identified and/or repairs recommended and or completed.

List of equipment

From time to time, equipment may be added or removed from the lists of equipment when required during the period of the contract. These additions or deletions may come from the respective site authority (Chief of Health Services) or from the Project Authority at the Regional office.



Any changes greater than 10% in the total number of items to be inspected as part of the contract or any changes greater than 30% in the number of items at anyone specific site will require a review of the inspection fees to take into consideration the additional or reduced workload. Should this situation occur, the Project Authority will request an updated quote from the contractor for work to be done.

NOTE: This clause excludes the addition or removal of diagnostic imaging systems and equipment. Any addition or removal of such equipment would require a new quote.

Response Time:

The Contractor shall provide these services Monday to Friday from 8:00 a.m. to 4:30 p.m.

Due to the nature of the environment in which the programs will be delivered, the scheduled dates of inspections/repairs will be subject to change due to urgent institutional circumstances such as institutional lockdowns, closures due to inclement weather, security requirements, or other emergency situations at no cost to the CSC. In these cases, the Project Authority or his/her representative will advise the Contractor and work will be scheduled.

Travel and Living Expenses

Travel and Living Expenses will be the responsibility of the contractor to and from CSC work sites. Vehicle transportation, driver's permits, insurances are also the responsibility of the contractor.

Project Authority

All work will be performed subject to the inspection and acceptance by the Project Authority or the on-site Chief of Health Services.

Project Authority name and phone number will be provided in the contract document.

The Project Authority or his designated representative will be responsible for monitoring the progress of the work and will be responsible for the technical requirements; the acceptance and approval of the deliverables. Any proposed changes to the scope of work may be discussed with the Project Authority, but any resulting changes can only be authorized by an amendment issued by the Contracting Authority.



APPENDIX 'A' – Glossary of Terms and Acronyms

| Term | Definition |
|-------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Contractor | A private business entity that is awarded the contract to perform the Work for Canada. |
| Equipment | Medical equipment at CSC Institutions that falls within the inspection profile of the Medical Device Maintenance Program. |
| Institution | Correctional facilities under the care of Correctional Service Canada where incarcerated offenders serve criminal sentences of two years or more as inmates. Institutions are secure facilities not broadly open to the public. |
| Technician | A specialized resource provided by the Contractor that is a trained Biomedical Technician with the requisite expertise and experience in the inspection of medical equipment. |



APPENDIX 'B' – List of CSC Institutions and Addresses

ATLANTIC REGION

Atlantic Institution: Attention – Chief of Health Care

13175 Route 8, P.O. Box 102
Renous, New Brunswick
E9E 2E1

Dorchester Complex Minimum Sector: Attention – Chief of Health Care

4902 Main Street
Dorchester, New Brunswick
E4K 2Y9

Dorchester Complex Medium Sector: Attention – Chief of Health Care

4902 Main Street
Dorchester, New Brunswick
E4K 2Y9

Nova Institution: Attention – Chief of Health Care

180 James Street
Truro, Nova Scotia
B2N 6R8

Springhill Institution: Attention – Chief of Health Care

330 McGee Street, P.O. Box 2140
Springhill, Nova Scotia
B0M 1X0

Shepody Healing Centre: Attention – Nursing Supervisor

4902 Main Street
Dorchester, New Brunswick
E4K 2Y9

Correctional Service Canada: Attention – Regional Pharmacist or Manager

Clinical Services Health Services (RHQ – Atlantic)

1045 Main Street
Moncton, New Brunswick
E1C 1H1

Appendix C – Equipment List Westmorland Institution

| Item | Equipment Code Code d'équipement | Institution Établissement | Location Localisation | Manufacturer Nom du fabricant | Model Modèle | Serial Number Numéro de série |
|------------------------|-------------------------------------|------------------------------|--------------------------|----------------------------------|-----------------|----------------------------------|
| Aerosol Compressor | | Westmorland Institution | Storage rm | Airlife | RT500 | 169547 |
| Aerosol Compressor | | Westmorland Institution | Storage rm | Invacare | IRC1196 | 99H328019 |
| Aerosol Compressor | | Westmorland Institution | Storage rm | Airlife | RT500 | 169547 |
| Aspirator | 221-006303 | Westmorland Institution | Dr room | Gomco | 300 | L10395 |
| Blanket warmer | 221-002028 | Westmorland Institution | Medication Room | Pedigo | P-2010S | 545084-000 |
| Blood drawing chair | 221-002015 | Westmorland Institution | Nurses Office | Winco | 574 | 574100103 |
| Camscope angled wand | 221-002023 | Westmorland Institution | Dr room | Sometech | DCS-104T (NTSC) | DSCN-11B-061 |
| Camscope receiver | 221-002023 | Westmorland Institution | Dr room | Sometech | DCSM (XGA) | DCSM-11C-029 |
| Camscope straight wand | 221-002023 | Westmorland Institution | Dr room | Sometech | P101 | DCSM-11C-115 |
| Cast Saw | 221-002021 | Westmorland Institution | Dr room | Stryker Canada | 840 | 92111273 |
| Caulk Dentsply | 221-0005000 | Westmorland Institution | Dentists rm | Dentsply International | VM-D | 71970 |
| Centrifuge | 221-002014 | Westmorland Institution | Nurses Office | Thermo Scientific | CL-2 | 42619248 |
| Compressor | 221-009082 | Westmorland Institution | storage rm | Midmark | 277OCE50-660A | |
| Curing Light | | Westmorland Institution | Dentists rm | Coltene/Whaledent INC. | C7970100115 | 110507017 |
| Curing Light | | Westmorland Institution | Dentist rm | 3M ESPE | Elipar S10 | 939112017014 |
| Dental Assistant Chair | | Westmorland Institution | Dentist rm | Belmont | 093G | 36911 |
| Dental Assistant Chair | | Westmorland Institution | Dentist rm | Belmont | 092Q | 36910 |
| Dental cart equip | 221-002027 | Westmorland Institution | Dentists rm | Aidec | PACI | C912329 |
| Dental Chair | 221-009078 | Westmorland Institution | Dentists rm | Belmont | X-Caliber | AN09A0073 |
| Dental chair equip | | Westmorland Institution | Dentists rm | Belmont | BDS0050 | 13344 |

| | | | | | | |
|-------------------------------|------------|-------------------------|----------------|-------------------------------------------|--------------|-----------------|
| Dental Light | | Westmorland Institution | Dentist rm | Belmont | AL501T5CR3 | AV09A0040 |
| Dental X-ray Arm | 221-009079 | Westmorland Institution | Dentist rm | Belmont | 096-A | EA08K0779 |
| Dental X-Ray Controler | 221-009079 | Westmorland Institution | Dentists rm | Belmont | 96 | EX08K0779 |
| Dental X-Ray Head | 221-009079 | Westmorland Institution | Dentists rm | Belmont | 096-H | EH08K0779 |
| Electrocardiograph | 221-002006 | Westmorland Institution | Dr room | Welch Allyn Canada Ltd. | CP10 | 105068614163 |
| Examining Light | 221-002009 | Westmorland Institution | Dr room | Welch Allyn Canada Ltd. | LS200 | ref # 74227 |
| Examining Light | 221-002008 | Westmorland Institution | Dr room | Welch Allyn Canada Ltd. | 48100 | |
| Geriatric chair | | Westmorland Institution | Storage Room | AMG Medical Inc. | 770-050 | 10860-0609-0003 |
| Grinder / polisher | 221-000535 | Westmorland Institution | Dentists rm | Red-wing | 26A | 816-227 84 |
| Interferential Ultrasound | 221-002007 | Westmorland Institution | Dr room | Dynatronics Research Corp. | D850PLUS | 860726 |
| IV Pump | 221-009028 | Westmorland Institution | Nurses office | Plum A+ | 84095270002 | 13898345 |
| Mister | 221-006303 | Westmorland Institution | Dr. room | Schuco | 5711102 | 188532 |
| Monitor – NIBP | 221-002013 | Westmorland Institution | Dr room | Critikon / Johnson & Johnson | 100V2 | 043M0299009 |
| Monitor – NIBP | 221-002026 | Westmorland Institution | Nurses office | Critikon/Johnson & Johnson | DCP400N-CE | 2019205-001 |
| O2 sat monitor | 221-002017 | Westmorland Institution | Nurses Office | Masimo Corporation | RAD5 | 500255 |
| O2 sat monitor | | Westmorland Institution | Dr room | Pulse Oximeer | M50E09CS0 | 9060429 |
| Ophthalmoscope | 221-002018 | Westmorland Institution | Nurses office | Welch Allyn Canada Ltd. | 74710 | 92192 |
| Ophthalmoscope | | Westmorland Institution | Dr room | Welch Allyn Canada Ltd. | 767 | 5290 |
| Physio Treatment Table | 221-002005 | Westmorland Institution | Dr room | Cardon Rehabilitation & medical equipment | R28535 | 039-08011 |
| Polycom Practitioner Cart | 221-001623 | Westmorland Institution | Dr room | Polycom | PPCE-120BP | L100614525 |
| Prophylaxis unit | 221-000499 | Westmorland Institution | Dentists rm | Dental Equipment | 2002 | 103-05362 |
| Sterilizer | 221-002024 | Westmorland Institution | Dentists rm | L&R | T-14B | Z85 83120 |
| Stretcher | 221-002020 | Westmorland Institution | Nurses Office | Midmark | K520 | TAV003996 |
| Temperature Monitor | 221-002018 | Westmorland Institution | Nurses office | Welch Allyn Canada Ltd. | 690 | 4106441 |
| Temperature Monitor | 221-002010 | Westmorland Institution | Dr room | Welch Allyn Canada Ltd. | 690 | 09361171 |
| Vaccine Refrigerator | | Westmorland Institution | Inf Dis Office | Panasonic | SR-L6111W-PA | 121101390 |
| Vibration machine | | Westmorland Institution | Dentists rm | | | G-10-1-4-V |
| Welch Allyn Temperature Probe | | Westmorland Institution | Nurses Office | Welch Allyn | 690 | 14310025 |

Appendix D - Equipment List Dorchester Medium

| Item | Equipment Code Code d'équipement | Institution Établissement | Location Localisation | Manufacturer Nom du fabricant | Model Modèle | Serial Number Numéro de série |
|-------------------------------------|-------------------------------------|------------------------------|--------------------------------|----------------------------------|----------------------------|----------------------------------|
| Amalgamator – Dental | 220-004041 | Dorchester Institution | Dentist's Office | The L.D Caulk Comp. | VM-C | 54154 |
| Aspirator / Suction Machine | | Dorchester Institution | Trauma Room | Gomco by Allied | 405 | 20080331005 |
| Aspirator 690 | 220-009018 | Dorchester Institution | Treatment room 1 | Contemporary Products | 6260 | AS21745 |
| Assembly cam-14 | | Dorchester Institution | Treatment room 1 | GE Medical Systems | CAM-14 | RNC11015386PA |
| Auto Refractometer Cradle | | Dorchester Institution | Optometry Room | NIDEK | AR20ST1 | 130493 |
| Auto refractometer Hand Held | | Dorchester Institution | Optometry Room | NIDEK | AR20 | 230454 |
| Battery pack for head lamp | 220-009033 | Dorchester Institution | X-ray room | Heine | M Pack | 2157354 |
| BIPAP Machine | | Dorchester Institution | Range | Philips | | C04772616 |
| Bladder scan | 220-009012 | Dorchester Institution | Treatment room 2 | Bladder Scan | 0570-0188 | PA011179 |
| Blanket Warmer | | Dorchester Institution | Medical Supply Room, 1st floor | Pedigo | P-20105 | 545088-000 |
| Blanket Warmer | | Dorchester Institution | Clean Utility Room | Pedigo | P-20105 | 545088-000 |
| Blood pressure/Pulse/O2 sat monitor | 220-009181 | Dorchester Institution | Storage room 3rd floor | Critikon | Dynamap Compact BP monitor | ref. 117207 |
| Blood Pressure/Temp/O2 Sat Monitor | 220-009004 | Dorchester Institution | Treatment room 1 | GE Medical Systems | Dynamap DPC400N-CE | 2019205-001 |
| Camscope box | 220-009022 | Dorchester Institution | Nursing supervisor's office | Sometec | DCSM(X6A) | DCSM-10G-167 |
| Camscope wand (angled) | 220-009022 | Dorchester Institution | Nursing supervisor's office | Sometec | DCS-104T (NTSC) | DCSN-10E-217 |
| Camscope wand (straight) | 220-009022 | Dorchester Institution | Nursing supervisor's office | Sometec | P101 | DCSN-11C-116 |
| Cassette autoclave | 220-006702 | Dorchester Institution | Dentist's lab | SciCan | 1102 | 228756A |
| Centrifuge | 220-007436 | Dorchester Institution | Treatment room 1 | Thermo IEC | CENTRA CL-2 | 42613405 |

| | | | | | | |
|--------------------------------|-----------------------------------------------------|------------------------|---------------------------------|--------------------------|----------------------|---------------|
| Chart Projector | 220-009030 + 220-006025 = 220-005098 Stand | Dorchester Institution | X-ray room | Marco | -- | 22143 |
| Compressed Air | | Dorchester Institution | Dentist's office | Belmont | 4000 | 20213 |
| Compressor | | Dorchester Institution | Dentist's office | Dominic Hunter | DHDRD10 | 3701300004 |
| Controls for X-ray machine | 220-009036 | Dorchester Institution | X-ray room | GE Medical System | 2320993 | 07115877 |
| Curing Light | | Dorchester Institution | Dentist's office | Coltène/Whaledent | Coltolux 75 | 060128666 |
| Dental Cart | 220-009024 | Dorchester Institution | Treatment room 1 | -- | 447192 | C912329 |
| Dental Chair | 220-009023 | Dorchester Institution | Dentist's office | Takara Belmont | 447193 | 012057301 |
| Dental Chair Instruments | | Dorchester Institution | Dentist's office | Belmont | | 12057301 |
| Dental Equipment | 220-09169 | Dorchester Institution | DP Minimum Sector - Dental Room | Dexis | DEXIS LLC - PLU5102C | 0134633003 |
| Dental light | 220-009025 | Dorchester Institution | Dentist's office | Belmont | AL501T | T040461 |
| Dental X-ray processor | 220-009026 | Dorchester Institution | Dentist's office | Dell | Studio Hybrid 140g | 18HVBM1 |
| Diagnostic X-ray wallstand | 220-009035 | Dorchester Institution | X-ray room | GE Hualun Medical System | 600-0301 | 50997HL6 |
| Digitizer | 220-009034 + 220-07878 | Dorchester Institution | X-ray room | AGFA | CR30-X | 2676 |
| Drill - Console - Surgical | 220-009024 | Dorchester Institution | Dentist's office | -- | -- | -- |
| Eclipse Proteus Collimator | 220-009038 | Dorchester Institution | X-ray room | GE Hualun Medical System | 2379827 | 53339HL8 |
| Electro therapy unit | 220-009015 | Dorchester Institution | Treatment room 2 | Chattanooga Canada | Intellect Adv | 4978 |
| Electrocardiograph | 220-009002 | Dorchester Institution | Treatment room 1 | GE Medical Systems | MAC5500 | 11-41583 |
| Emergency Cart | 220-009042 | Dorchester Institution | Storage room 3rd floor | Harloff Co. | | 46269 |
| Frame Heater | | Dorchester Institution | Optometry Room | GFC | A600 | 682 |
| Grinder Polisher Lath - Dental | | Dorchester Institution | Dentist's Office | Red Wing | 16B | 243/10/11/044 |
| Head lamp | 220-009033 | Dorchester Institution | X-ray room | Heine | -- | 1130545 |
| Hyfrecator | 220-000209 | Dorchester Institution | Treatment Room 1 | Birtcher | | 2157354 |
| Ice maker | | Dorchester Institution | B-6 End of hallway | Whirlpool | GI15NDXTB1 | EW1501128 |
| Integrated Diagnostic System | | Dorchester Institution | Trauma Room | Welch Allyn | GS 777 WA 7791-MX | |
| IV Pump | | Dorchester Institution | Trauma Room | Hospira | 1239-104-07 | 17446358 |
| IV Pump – Infusion | 220-009001 | Dorchester Institution | Treatment room 1 | Hospira | 1239-104-03 | 13898441 |

| | | | | | | |
|------------------------|------------|------------------------|-------------------------------|------------------------|-----------------|-------------|
| IV Pump – Infusion | 220-009028 | Dorchester Institution | Medical supply room 1st floor | Hospira | 1239-104-03 | 13898345 |
| Keratometer | 220-006033 | Dorchester Institution | Optometry Room | Marco | | 21226 |
| Laser module | 220-009037 | Dorchester Institution | Treatment room 2 | Chattanooga Canada | 2762CC | 2033 |
| Lensmeter | 220-009037 | Dorchester Institution | X-ray room | Marco | 201 | M2474 |
| Lensometer | | Dorchester Institution | Optometry Room | | | |
| Lift | | Dorchester Institution | Range | Steady Aid | PWR TSA3510X | 7159 |
| Lift - Patient – Floor | | Dorchester Institution | Range | BHM Medical | Ergolift | ERLI2746 |
| Lift - Patient – Floor | | Dorchester Institution | Range | | 0102-8851 | 96511054100 |
| Lift - Tub lift | | Dorchester Institution | Range | TR EQUIPMENT | 9650-0542020 | 1210-20929 |
| LIFT-PATIENT-FLOOR | | Dorchester Institution | | TRE EQUIPMENT | 96511054010 | 80616861 |
| Light – Examining | 220-009014 | Dorchester Institution | Treatment room 2 | Welch Allyn Canada Ltd | 44200 | -- |
| Mister – Pump | 220-002007 | Dorchester Institution | Storage 1st Floor | Schuco | 5711102 | |
| Monitor - O2 sat | 220-009003 | Dorchester Institution | Treatment room 2 | Masimo Corporation | RAD5 | 532646 |
| Monitor - O2 sat | 220-009008 | Dorchester Institution | Treatment room 1 | Masimo Corporation | RAD5 | 530779 |
| Monitor - SPO2 | | Dorchester Institution | To be confirmed | | | |
| Monitor - SPO2 | | Dorchester Institution | To be confirmed | | | |
| Monitor - SPO2 | | Dorchester Institution | To be confirmed | | | |
| Monitor – Temp | | Dorchester Institution | IDN Office | Welch Allyn | 690 | 08272340 |
| Monitor – Temp | | Dorchester Institution | | Welch Allyn | 690 | 11110835 |
| Monitor – Temp | | Dorchester Institution | To be confirmed | Welch Allyn | | |
| Monitor – Temp | | Dorchester Institution | To be confirmed | Welch Allyn | | |
| Monitor – Temp | | Dorchester Institution | To be confirmed | Welch Allyn | | |
| Monitor – Temp | | Dorchester Institution | Trauma Room | Welch Allyn | 690 | 12512584 |
| Monitor Temp | 220-009010 | Dorchester Institution | | Welch Allyn | 690 Thermometer | 08047847 |
| Monitor Temp | | Dorchester Institution | To be confirmed | Welch Allyn | | |
| Ophthalmic Chair | 220-07672 | Dorchester Institution | | | FX920 | 93010707007 |
| Ophthalmoscope | 220-009021 | Dorchester Institution | Treatment Room 2 | Welch Allyn | | |
| Ophthalmoscope | 220-009006 | Dorchester Institution | Treatment Room 1 | Welch Allyn | | |
| Ophthaomalogy Chair | 220-009031 | Dorchester Institution | X-ray room | Power Tilt | FX920L | 93010707007 |
| Optometry equipment | 220-09170 | Dorchester Institution | Optometry Room | NIDEK | | 130493 |
| Optometry Equipment | 220-07671 | Dorchester Institution | Optometry Room | | 7900 | 79010702014 |
| Optometry Equipment | 220-002017 | Dorchester Institution | Optometry Room | Masimo | | 500255 |
| Patient Chair | | Dorchester Institution | Storage room | Invacare | IH6077A | 060981 |
| Portable Aspirator | | Dorchester Institution | Health Services | | L190-GR | 0813 |
| Practitioner cart | | Dorchester Institution | Nursing supervisor's office | Polycom | PPCE-120BP | N091215032 |

| | | | | | | |
|------------------------------------|---------------------------|------------------------|-------------------------------|---------------------------|--------------------------|-----------------|
| Projector - Ophthalmic | 220-009030 | Dorchester Institution | X-ray room | Marco | 50 | M22143 |
| Proteus elevating table | 220-009038 | Dorchester Institution | X-ray room | GE Hualun Medical System | 2259988 | 50508HL9 |
| Proteus System Console | 220-009038 | Dorchester Institution | X-ray room | GE Hualun Medical System | 2259976 | 50529HL7 |
| Proteus XR-A System | 220-009038 + 220-07879 | Dorchester Institution | X-ray room | GE Hualun Medical System | 2259973-2 | 50567HL7 + 5795 |
| Pump – Compressor | | Dorchester Institution | Trauma Room | AirLife | RT500 | 169542 |
| Pump – Infusion | 220-006718 | Dorchester Institution | Medical supply room 1st floor | Abbott Laboratories | PLUMXL 11555-04-01 | 11204399 |
| RAD-14X-ray-tube | 220-009038 | Dorchester Institution | X-ray room | GE Hualun Medical System | 2259981 | 95296-7U |
| Refractor | 220-0004032 | Dorchester Institution | X-ray room | Marco | RT-I | 10975 |
| Refrigerator - Pharmacy | 220-07993 | Dorchester Institution | IDN Office | | SAN-SRL6111W | 120200257 |
| Refrigerator - Pharmacy 17.2 cu ft | 220-07992 | Dorchester Institution | Secondary Pharmacy | | SAN-MPR514 | 12010029 |
| Ritter automatic sterilizer | | Dorchester Institution | Dentist's lab | Midmark | M9 UltraClave M9-022 | V546354 |
| Saw – Cast | 220-009016 | Dorchester Institution | Treatment room 2 | Stryker | 8127 | E34984 |
| Scale – Patient | 220-009017 | Dorchester Institution | Treatment Room 1 | Pelstar | 500KL | 5000005579 |
| Slit Lamp | 221-006503 | Dorchester Institution | Dr room | Top Con PS-20A | SL-2E | 6211049 |
| Slit Lamp | 220-009040 | Dorchester Institution | Optometry Room | HAAG Streit | BQ900 | 10654 |
| Solid state ultrasonic sterilizer | 220-000201 | Dorchester Institution | Dentist's lab | L & R Manufacturing | T-14B | 150 |
| Stand - optometrist equipment | 220-009032 | Dorchester Institution | X-ray room | Reliance Medical Products | 7900 | 79010702014 |
| Steam Sterilizer Table Top | | Dorchester Institution | Dentist's Office | SciScan | Statim 2000S - 01-121105 | 101212J00019 |
| Stretcher | 220-006439 | Dorchester Institution | Treatment room 1 | Midmark | 520 | |
| Stryker Stretcher | 220-009087 | Dorchester Institution | | Stryker | 0747-000 | 1105034705 |
| Stryker Stretcher | 220-009088 | Dorchester Institution | | Stryker | 0747-000 | 1105034706 |
| Table - Physio-treatment | 220-009013 | Dorchester Institution | Treatment room 2 | Chattanooga Canada | Triton DTS TRT-600 | 2214 |
| Tonometer | | Dorchester Institution | X-ray room | Clement Clarke | Perkins tonometer MK2 | 1999T2073 |
| Traction unit | | Dorchester Institution | Treatment room 2 | Chattanooga Canada | 2841/3336 | |
| Trolley | | Dorchester Institution | Treatment room 1 | GE Medical Systems | MAC 5000 | SCD11051932PA |
| Ultra sound applicator | | Dorchester Institution | Treatment room 2 | Chattanooga Canada | 27335 | 37018 |
| Ultrasonic Cleaner | 220-07243 | Dorchester Institution | Dentist's Office | Densply | cAVITRON SPS | 119-27469 |
| Ultraviolet Light | | Dorchester Institution | Dentist's Office | Elipar | S10 | 939112017246 |
| UV inspection lamp | | Dorchester Institution | Treatment room 1 | Luxo | 0316040 | 89705 |
| Vacume Pump - Dental | | Dorchester Institution | Hall outside Dental Office | VacStar | VS20 | 208990 |
| Vital signs monitor | 220-009011 | Dorchester Institution | Treatment room 2 | Welch Allyn | Spot Vital Signs Lxi | 20110201548 |
| Washer / Disinfectant | | Dorchester Institution | Hall | FRANKE | | |
| Water Bath | | Dorchester Institution | Dentist's Office | Whip Mix | | 07110195 |
| X-ray Controller | 220-009027 | Dorchester Institution | Dentist's office | Belmont | 097-CM | EX10J0141 |
| X-ray Head | 220-009027 | Dorchester Institution | Dentist's office | Belmont | 097-H | EH10J0510 |
| X-ray printer | 220-009039 | Dorchester Institution | X-ray room | Sharp AGFA | Drystar 5300 | |

| | | | | | | |
|--------------------|------------|------------------------|------------------|---------|-------|-----------|
| X-ray Regular Cone | 220-009027 | Dorchester Institution | Dentist's office | Belmont | 097-R | ER10J0510 |
|--------------------|------------|------------------------|------------------|---------|-------|-----------|

Appendix E – Equipment List Shepody Healing Center

| Item | ~~Equipment Code~~Code d'équipement~~ | ~~Institution~~Établissement~~ | ~~Location~~Localisation~~ | ~~Manufacturer~~Nom du fabricant~~ | ~~Model~~Modèle~~ | ~~Serial Number~~Numéro de série~~ |
|-------------------------|---------------------------------------|--------------------------------|----------------------------|------------------------------------|-------------------|------------------------------------|
| JAMAR DYNAMETER | | Shepody Healing Centre | PSR Unit | | | |
| MONITOR NIBP SPO2 TEMP | | Shepody Healing Centre | Acute Unit | Welch Allyn Canada Ltd | 42NTB | |
| MONITOR NIBP SPO2 TEMP | | Shepody Healing Centre | PSR Unit | Welch Allyn Canada Ltd | 42NTB | |
| MONITOR NIBP SPO2 TEMP | | Shepody Healing Centre | PSR Unit | Welch Allyn Canada Ltd | 42NMTB | 201419223 |
| MONITOR NIBP SPO2 TEMP | | Shepody Healing Centre | Acute Unit | Welch Allyn Canada Ltd | 42 MTB | 201419692 |
| MONITOR SP02 | | Shepody Healing Centre | Acute Unit | O-Two Systems International Inc | 01MN1000 | |
| MONITOR SP02 | | Shepody Healing Centre | Acute Unit | O-Two Systems International Inc | 01MN1000 | |
| MONITOR SP02 | | Shepody Healing Centre | PSR Unit | O-Two Systems International Inc | 01MN1000 | |
| MONITOR TEMP | | Shepody Healing Centre | Acute Unit | Welch Allyn Canada Ltd | 690 | |
| MONITOR TEMP | | Shepody Healing Centre | Acute Unit | Welch Allyn Canada Ltd | 690 | |
| MONITOR TEMP | | Shepody Healing Centre | PSR Unit | Welch Allyn Canada Ltd | 690 | |
| MONITOR TEMP TYMPANIC | | Shepody Healing Centre | Acute Unit | Welch Allyn Canada Ltd | PRO4000 | |
| SCALE PATIENT x2 | | Shepody Healing Centre | PSR Unit | Health-O-Meter | 402KL | 4020079413 |
| SPHYGMOMANOMETER ANALOG | | Shepody Healing Centre | Acute Unit | Heine | GAMMA G5 | |
| Stretcher | | Shepody Healing Centre | PSR Unit | Techlem | 500 | TMRR1303029 |
| Vaccine fridge Acute | | Shepody Healing Centre | Acute Unit | Haier (LG Electronics) | | NSA36LACG |
| Vaccine fridge PSR | | Shepody Healing Centre | PSR Unit | LG | GC151RCW | |

Appendix F – Equipment List Atlantic Institution

| Item | Equipment Code Code d'équipement | Institution Établissement | Location Localisation | Manufacturer Nom du fabricant | Model Modèle | Serial Number Numéro de série |
|------------------------------------|-------------------------------------|------------------------------|--------------------------|------------------------------------|-------------------|----------------------------------|
| FRIDGE | 200-004118 | Atlantic Institution | Pharmacy | | 428730 | EW3317486 |
| Fridge- Vaccine | 231-004141 | Atlantic Institution | Pharmacy | Panasonic | SR-L6111W-PA | 120700476 |
| Auto Reader | 231-3335 | Atlantic Institution | D306 | 3M Attest | 305807 | 208585 |
| Defibrillator, Automated, E | 231-004502 | Atlantic Institution | Hall | | 403018 | X08B147213 |
| Electrocardiograph | 231-003476 | Atlantic Institution | D303 | Welch Allyn Canada Ltd. | CD200 | 20004343 |
| TONOMETER | 231-000285 | Atlantic Institution | D104 | Topcon (Koenigkramer) | SL4E | 643280 |
| Light-examining | 231-000273 | Atlantic Institution | D303 | Heraeus | 55002286 | 100005 |
| Monitor - SPO2 | | Atlantic Institution | | Nonin | 8500 | 500257365 |
| Monitor - SPO2 | 231-004527 | Atlantic Institution | Unit 3 | Nonin | 8500 | 500714515 |
| Monitor NIBP-SPO2-Temp- Heart Rate | 231-003761 | Atlantic Institution | | General Electric Mesystems Limited | 2009284001 | AAW080502335 |
| Pump - Infusion | 231-001972 | Atlantic Institution | D303 | Abbot Laboratories | XL | 11203144 |
| Pump - Infusion | 231-004501 | Atlantic Institution | D303 | Hospira Worldwide Inc. | 123910403 | 13898447 |
| Strecher Ferno | 231-3270 | Atlantic Institution | Hall | FERNO | 287857 | 785076 |
| Fridge- Vaccine | 231-003427 | Atlantic Institution | Pharmacy | Follett | 128934 | B89235-35605 |
| Warmer, Blanket | 231-3582 | Atlantic Institution | D306 | Pedigo | 398442 | 545085-00 |
| MONITOR-BP-ECG-NIBP-REC-SPO2 | 231-004512 | Atlantic Institution | Unit 3 | Dinamap | V100-PBXA-AXAA-DX | SH6111400455A |
| OPHTHALMIC INSTRUMENT STAND | 231-004504 | Atlantic Institution | D104 | Topcon (Koenigkramer) | IS2000 | 351541 |
| CLEANER-ULTRASONIC | 231-004519 | Atlantic Institution | D308 | L & R | T14B | Q87122194T |
| STERILIZER-STEAM-TABLE-TOP | 231-001988 | Atlantic Institution | D308 | STATim | 1102 | 228426A |
| CHAIR-DENTAL | 231-003832 | Atlantic Institution | D308 | DENTECH | ULTIMA2 | CM2315-3-K1 |

| | | | | | | |
|----------------------------------|--------------|----------------------|----------------------------------|--------------------------|----------------|--------------|
| XRAY-UNIT | 231-003766 | Atlantic Institution | D308 | Sinclair Dental | BELRAY 096 | EX08K0769 |
| TABLE-PHYSIO-TREATMENT | 231-004508 | Atlantic Institution | D304 | Centry Medical Dist | TRT600 | 2245 |
| TRACTION-UNIT | 231-004509 | Atlantic Institution | D304 | Triton-Chattanooga | MP1 | 3358 |
| TABLE-EXAMINING | 231-002886 | Atlantic Institution | NOT IN USE - In Basement(Tunnel) | | 056-605 | 2702002001 |
| BED-ELECTRIC | 231-003076 | Atlantic Institution | D2017 - Dry Cell | Hill Room | P8400D3867 | E086AL6139 |
| LENSOMETER | 231-000281 | Atlantic Institution | D104 | | 322056 | 322056 |
| ASPIRATOR-PORTABLE | 231-004510 | Atlantic Institution | Unit 3 | | L190-GR | NONE |
| THERAPY SYSTEM - Physio | 231-004506 | Atlantic Institution | D304 | Intelect Advanced | 2762CC | 4879 |
| PUMP-COMPRESSOR | | Atlantic Institution | D303 | Nedquip | RT500 | 170682 |
| ASPIRATOR | 231-004505 | Atlantic Institution | D303 | Advantage | L190-GR | NONE |
| DEFIB-AED | 231-004503 | Atlantic Institution | Hall | AEDPLUS | X08D152305 | 7478539 |
| MONITOR-SPO2 | 231-004513 | Atlantic Institution | Unit 3 | | 8500 | 500327436 |
| Optometrist Chair | 231-004576 | Atlantic Institution | D104 | TOPCON | | |
| Autoscaler | | Atlantic Institution | Room D308 | South East- Henry Schein | | |
| Fridge | CSC-004229 | Atlantic Institution | | Danby | 2306 | |
| Dr. Camscope - Otoscope | | Atlantic Institution | D112 | Sometech Inc | DCS-104 (NTSC) | DCSM-118-059 |
| Dr. Camscope - Intra Oral Camera | | Atlantic Institution | D112 | Sometech Inc. | DCS-P101 | DCSN-11C-118 |
| Dr. Camscope - C.V.P. System | | Atlantic Institution | D112 | Sometech Inc. | DCSM (XGA) | DCSM-11C-030 |
| AMD-2500 | | Atlantic Institution | D112 | AMD Global Telemedicine | AMD-2500 | |
| AMD-3700 | | Atlantic Institution | D112 | AMD Global Medecine | | 21151 |
| AMD-400 | | Atlantic Institution | D112 | AMD Global Telemedicine | | |
| Hot Water Bath | 231-308 | Atlantic Institution | | Eagle | | |
| Lathe | 231-000312 | Atlantic Institution | D308 | BDMI | | 816-22-1287 |
| Bed, Examination | 23100-004036 | Atlantic Institution | D112 | Global Medco Ltd. | 9105-30 | |
| Bed, Examination | 23100-004035 | Atlantic Institution | Unit 3 | Global Medco Ltd. | 9105-30 | |
| Stretcher | 231- 003579 | Atlantic Institution | D303 | Stryker | 1007 | |
| Amalgamator | 231-24502 | Atlantic Institution | D308 | Caulk | | |

| | | | | | | |
|----------------------------------|--------------|----------------------|----------------------------|-------------------------|----------------|--------------|
| Light - Examining (Optical) | 231-0001697 | Atlantic Institution | D104 | Hein | | |
| Phoropter (Optical) | 231-001698 | Atlantic Institution | D104 | Reichert | 11625 AR | 10806-2 |
| KEROMETER (Optical) | 231-284 | Atlantic Institution | D104 | Bausch & Lomb | B & L 71-21-35 | 18256 |
| Microscope Double Zoom (Optical) | 231-000284 | Atlantic Institution | D104 | Amtek | CP 40 | 105072 |
| Dry Imager | 23100-004521 | Atlantic Institution | D304 | AGFA | Drystar 5300 | |
| Proteus XR/A | 23100-004515 | Atlantic Institution | D304 | Ge | AL01F | 02246 |
| X-Ray Table | 23100-04516 | Atlantic Institution | D304 | GE | | |
| Digitizer | 23100-004522 | Atlantic Institution | D304 | Agfa | CR 30-X | |
| Chest X-Ray Machine | 231-289 | Atlantic Institution | D304 | | 260-076-R | 1462 |
| Sterilizer | 23100-004140 | Atlantic Institution | D308 | Statim | Statim 2000 G4 | 101212j00026 |
| Refrigerator, Laboratory | 23100-004141 | Atlantic Institution | Pharmacy | Panasonic | SR-L6111W-PA | 120700476 |
| Console Xray Unit | | Atlantic Institution | D304 | GE | 2259976 | |
| Light Ultraviolet | | Atlantic Institution | | DSP | RADII PLUS | |
| Monitor Temp Tympanic | | Atlantic Institution | | Welch Allyn | Braun 6021 | |
| Tube - XRay Unit | | Atlantic Institution | D304 | GE | 2269646 | |
| Monitor SPO2 | 23100 004169 | Atlantic Institution | Unit 5 | NONIN | SPECTRO2 | 4047289 |
| MONITOR-BP-ECG-NIBP-REC-SPO2 | 23100-004170 | Atlantic Institution | Unit 5 | Welch Allyn Canada Ltd. | | 20140201184 |
| MONITOR-BP-ECG-NIBP-REC-SPO2 | 23100-004172 | Atlantic Institution | Unit 5 | Welch Allyn Canada Ltd. | | 20140201183 |
| Otto/Ophthalmoscope-Handheld | 23100-004171 | Atlantic Institution | Unit 5 | Welch Allyn Canada Ltd. | | 05259-M |
| STRETCHER-FERNO | 23100-004142 | Atlantic Institution | Hallway | Ferno | 35X proflex | 12N232511 |
| Monitor SPO2 | | Atlantic Institution | Unit 5 | Nonin | 8000AA | 500716022 |
| Stretcher | 23001-004142 | Atlantic Institution | Hallway | FERNO | | 238-1493-01B |
| Centrifuge | | Atlantic Institution | Pharmacy | Unico | Power Spin FX | 250-520-02 |
| Stretcher | | Atlantic Institution | Basement-For use in Unit 5 | FERNO | 287784 | BBB113616 |

Appendix G – Equipment List Springhill Institution

| Item | ~~Equipment Code~~Code d'équipement~~ | ~~Institution~~Établissement~~ | ~~Location~~Localisation~~ | ~~Manufacturer~~Nom du fabricant~~ | ~~Model~~Modèle~~ | ~~Serial Number~~Numéro de série~~ |
|------------------------------------------|---------------------------------------|--------------------------------|----------------------------|------------------------------------|---------------------|------------------------------------|
| AMTER KERATOMETER | 210-003190 | Springhill Institution | Optometrist | Topcon | 238714 | 01074G |
| Light Head - HEINE OMEGA 180 INDIRECT OP | 210-003429 | Springhill Institution | Optometrist | Heine | 268299 | 0/11848 |
| TOPCON COUNTERBALANCED MIDD | 210-003194 | Springhill Institution | Optometrist | Topcon | 238725 | 203100017B |
| TOPCON HALOGEN CHART PROJEC | 210-003193 | Springhill Institution | Optometrist | Topcon | 238723 | 262933G |
| TOPCON MOTORIZED OPLTHALMIC Chair | 210-003191 | Springhill Institution | Optometrist | Topcon | 238717 | 222466 |
| TOPCON POHTHALMIC STAND | 210-003192 | Springhill Institution | Optometrist | Topcon | 238721 | 350236 |
| TOPCON VISION TESTER | 210-003189 | Springhill Institution | Optometrist | Topcon | 238712 | 3617864 |
| TONOMETER | 210-001322 | Springhill Institution | Optometrist | TONO-PENXL | | T5932 |
| Dina care Procure (BP temp, O2 sats) | 210-005006 | Springhill Institution | Treatment | GE | 2009284001 | AAW080502158A |
| ECG Machine | 210-004364 | Springhill Institution | Treatment | GE | MAC5500 | SCD10087696PA |
| Gomco Suction Aspirator | 210-005009 | Springhill Institution | Treatment | Gomco | 4005 | 20071218026 |
| Portable Suction-Aspirator | 210-005004 | Springhill Institution | Treatment | Devil Bless | 7305 | PD235889 |
| HILL ROM STRETCHER | 210-003427 | Springhill Institution | Treatment | Hill Rom | 268261 | F093AW7612 |
| IV Pump | 210-005008 | Springhill Institution | Treatment | Hospira PlumA+ | 12910405 | 13898186 |
| Blanket Warmer | 210-005008 | Springhill Institution | Treatment | Pedigo | P-2010S | 545087 |
| Pulse Oximeter Nellcor NPB4 | 210-003983 | Springhill Institution | | | 365367 | G00842987 |
| Belmont Light | | Springhill Institution | dentist | Belmont 501T | 105000 Dental Light | H71903 |
| Curing Light-Densply | | Springhill Institution | dentist | Bobcat | 506 QHL75 | 506-30012 |
| Speed Clave Sterilizer | | Springhill Institution | dentist | Midmark | M7011 Speed Clave | MH003986 |
| Ultra Sonic Unit | | Springhill Institution | dentist | DENSPLY | Cavitron | 11546484 |
| Vacuum, water spurt | | Springhill Institution | dentist | Henry Schien | | 07344-563850 |
| Dental Chair | | Springhill Institution | dentist | Belmont | X-Calibur | 73203603 |

| | | | | | | |
|-------------------------------------------------------|------------|------------------------|-----------------|-------------------------|---------------------------|--------------|
| X-Ray Head | 210-004318 | Springhill Institution | dentist | Belmont | 428384 | EH07I0201 |
| Amalgamator | | Springhill Institution | dentist | Ultra Mat 2 | | 71903 |
| Ultrasound (physio) | 210-005015 | Springhill Institution | | Chattanooga Canada | 2762CC Intellect Advanced | 4295 |
| Hot Wax machine Inva care | | Springhill Institution | | Chattanooga Canada | Para Care | 24753 |
| Steam pack heating unit Hydro collator | 210-005014 | Springhill Institution | | Chattanooga Canada | Hydrocollator -2102C | 2328C |
| X-Ray System | 210-000434 | Springhill Institution | X-ray | GE | 5892919 | 4034 |
| X-Ray Unit | 210-000435 | Springhill Institution | X-ray | GE | 600-030 | 46335HL6 |
| X-Ray Table | 210-00436 | Springhill Institution | X-ray | GE | X171137 Picker | B25863 |
| X-Ray Digital Panoramic System | 210-004213 | Springhill Institution | X-ray | GENDEX | Orthoralix 8500 | 35-1858143DP |
| Medication Refridgerator | | Springhill Institution | Medication Room | Sanyo | MPR721 | 11110815 |
| Physio Treatment Table | | Springhill Institution | Doctor's Office | Chattanooga Canada | TRT400 | 11717 |
| Monitor SPO2 | | Springhill Institution | | Nellcor Puritan Bennett | NPB40 | G00842991 |
| Dr. Camscope Intra Oral Camera Telemedicine | | Springhill Institution | Back office | Sometech Inc | DCSN | DCSN-11C-117 |
| Dr. Camscope Otoscope - Telemedicine | | Springhill Institution | Back Office | SomeTech | DCS-104(NTSC) | DCSN-11B-058 |
| Dr. Camscope C.V.P. System - Telemedicine Control Box | | Springhill Institution | Back Office | | DCSM(XGA) | DCSN-11B-002 |
| Ophthalmoscope | | Springhill Institution | Dr Office | Welch Allyn Canada | 767 | |
| Monitor-Temp | | Springhill Institution | | Welch Allyn | 690 Suretemp | 743382 |
| Monitor-Temp | | Springhill Institution | | Welch Allyn | 679 Suretemp | 3137764 |
| Ophthalmoscope | | Springhill Institution | | Welch Allyn | 767 | |
| Dental Ultra Sonic | | Springhill Institution | Dentist | L&R Medical | Quantrex140 | 196231 |
| Dental Cart | | Springhill Institution | Dentist | ADEC | | |
| PUMP Compressor | | Springhill Institution | Treatment Room | Devilbiss | 5650C Pulmoaide | C170892 |
| Examining Light | | Springhill Institution | Treatment Room | Welch Allyn | 44300/LS135 | |

| | | | | | | |
|-------------------------------------------------------------|--------|------------------------|-------------------------|--------------------|--------------------|-------------|
| Pump Compressor | | Springhill Institution | Treatment | Invacare Corp | IRC1196 - Passport | 59400000890 |
| Pump Compressor | | Springhill Institution | Treatment | Invacare Corp | IRC1196 - PASSPORT | 59400000908 |
| Pump Compressor | | Springhill Institution | Treatment | Invacare Corp | IRC1196 - PASSPORT | 59400000897 |
| Monitor-NIBP-SP02-Temp | | Springhill Institution | | Welch Allyn | 42MTB Spot Vital | 201103450 |
| Ophthalmoscope | | Springhill Institution | | Welch Allyn | 767 | |
| Lensometer | | Springhill Institution | | Bausch & Lomb | 712655 | 512045 |
| Vaccine Refrigerator - RRC Panasonic Labratory Refrigerator | | Springhill Institution | Nursing Office - RRC | Panasonic | SR-L6111W | 120700510 |
| Monitor-NIBP-SP02-TEMP | | Springhill Institution | RRC - Nursing Office | Welch Allyn | 420 Series | 201216120 |
| Monitor-NIBP-SP02-TEMP - Blood Pressure Machine | | Springhill Institution | Dr. Office | Welch Allyn | 420 Series | 201216780 |
| Pulse Oximeter Nellcor N-65 | | Springhill Institution | Dr. Office | Nellcor | N-65P | P12827521 |
| Vaccine Fridge - Panasonic Vaccine Fridge - Treatment Room | | Springhill Institution | Treatment Room | Panasonic | | 121001043 |
| Cast Saw | | Springhill Institution | Treatment Room | BSN Medical Inc. | 0295-200 | 28330 |
| PortaCount | 004382 | Springhill Institution | Fit Testing Office | Portacount Pro+ | 8038 | 80389433 |
| Particle Generator (part of fit tester) | 004382 | Springhill Institution | Fit Testing Office | Particle Generator | 8026 | 80263697 |
| AGFA | | Springhill Institution | X-Ray room | AGFA | CF30X | 1380 |
| AGFA | | Springhill Institution | X-ray room | AGFA | Drystar 5300 | |
| Pedigo Stretcher | | Springhill Institution | Treatment Room | Pedigo | PP7500-W-Spec | 0199025 |
| Blood pressure/O2Stat and Temp | | Springhill Institution | Treatment Room/Portable | Welch Allyn | 42NTB | 201403441 |
| Monitor-NIBP-SP02-Temp | | Springhill Institution | | Welch Allan | 42NTB | |

| | | | | | | |
|-----------------------|--|------------------------|--|-------------|---------|--|
| MONITOR-TEMP-TYMPANIC | | Springhill Institution | | Welch Allan | 6021 | |
| MONITOR-TEMP-TYMPANIC | | Springhill Institution | | Welch Allan | 6022 | |
| Pump Compressor | | Springhill Institution | | Innovocorp | IRC1196 | |
| Scale - Patient | | Springhill Institution | | DET | 6437 | |
| Tonometer | | Springhill Institution | | CLC | MK2 | |
| Wheelchair | | Springhill Institution | | AMG | 700-631 | |
| Wheelchair | | Springhill Institution | | AMG | 700-625 | |
| Wheelchair | | Springhill Institution | | AMG | 700-621 | |
| Wheelchair | | Springhill Institution | | AMG | 700-627 | |
| Wheelchair | | Springhill Institution | | AMG | 700-632 | |

Appendix H – Equipment List Nova Institution for Women

| Item | ~~Equipment Code~~Code d'équipement~~ | ~~Institution~~Établissement~~ | ~~Location~~Localisation~~ | ~~Manufacturer~~Nom du fabricant~~ | ~~Model~~Modèle~~ | ~~Serial Number~~Numéro de série~~ |
|------------------------|---------------------------------------|--------------------------------|----------------------------|------------------------------------|---------------------------|------------------------------------|
| Amalgam separator | 250-001441 | Nova Institution | Dental suite (pump room) | Henry Schein | Solmetex/Vacstar/Internat | 25669/89824/5858009 |
| Amalgamator | | Nova Institution | Dental suite | Silamat | S5 | |
| Applanation Tonometer | 25000-003995 | Nova Institution | Dr office | Axis Medical | R-Type | 05117 |
| Aspirator | 250-001576 | Nova Institution | Tx room | Gomco | 4005 | 20071218030 |
| Bladder Scanner | | Nova Institution | Treatment Room | Verathon | Phanton BVI-6100 | ST300728 |
| Blanket warmer | | Nova Institution | Treatment Room | Olympic | Warmette | 13507 |
| Canal measurement tool | 250-001442 | Nova Institution | Dental suite | Sinclair Dental | Root2XII | |
| CPAP machine | | Nova Institution | Storeroom | ResMed | H5i/S9 | 22112011401/22111819 |
| Curing light | 250-001445 | Nova Institution | Dental suite | Demetron | A.2 | |
| Cylinders Phoroceptor | 25000-003994 | Nova Institution | Dr office | Reichert | Minus | 582150613 |
| Dental chair | 250-001449 | Nova Institution | Dental suite | Xcalibur | | 94492407 |

| | | | | | | |
|--------------------------------------|--------------|------------------|----------------------|-----------------|--------------------|----------------|
| Dental examination lamp | 250-001446 | Nova Institution | Dental suite | ALT | 508T12N | CN07L0139 |
| Desk viewer, X-ray backlight | | Nova Institution | Dental suite | Dux | | |
| Digital scale | 250-001572 | Nova Institution | Tx room | Health-o-Meter | 2842KL | 7E620450000069 |
| Dinamap V100 (Bloodpressure machine) | 25000 001652 | Nova Institution | Treatment Room | Dinamap | 2038172-001-219568 | SH 615045495A |
| Doppler | 250-001560 | Nova Institution | Ax room | Medisonics | | FB06949 |
| Drill station | 250-001452 | Nova Institution | Dental suite | Adec Pac | 3420 | C888511 |
| ECG | 250-001458 | Nova Institution | Tx room | Welch Allyn | CP2AS | 20004345 |
| ECG | | Nova Institution | Tx Room | Welch Allyn | CP 100 | 10021867 |
| Exam table | 250-001568 | Nova Institution | Secure Unit | Midmark | Ritter 104 | |
| Exam table | 25000-001381 | Nova Institution | Ax room | Ritter | Midmark 222 | V1168926 |
| Examination lamp | 250-001578 | Nova Institution | Tx room | Luxo | | |
| Examination lamp | 250-001566 | Nova Institution | Secure Unit | | LS-135 | |
| Examination lamp | 250-001551 | Nova Institution | Ax Room | Welch Allyn | GS300 | |
| Half Shadow Mirror | | Nova Institution | Dr office | Axis Medical | Plain white | none |
| Handheld BP Monitor | 250-001555 | Nova Institution | Emerg kit- stockroom | Welch Allyn | 3400 | 100000050811 |
| Heat source | 250-001443 | Nova Institution | Dental suite | System B | | |
| Hot pack warmer | 250-001118 | Nova Institution | Tx room | | E-1 | 137706 |
| Hydrocollator | | Nova Institution | Treatment Room | Chattanooga | 2402 | T7783C |
| ID Fridge | 250-001467 | Nova Institution | Med Room | Follett | Symphony series | |
| ID Fridge | 250-001557 | Nova Institution | Med Room | Follett | REF series | |
| Infusion pump | 250-001577 | Nova Institution | Tx room | Plum A+ | | 13897820 |
| Integrated Diagnostic System | 250-001558 | Nova Institution | Ax Room | Welch Allyn | | |
| Integrated Diagnostic System | 250-001580 | Nova Institution | Tx Room | Welch Allyn | | |
| Intraoral x-ray unit | 250-001450 | Nova Institution | Dental suite | Gendex/Visualix | | 1514B-3945ABG |
| LCM control module | 250-001440 | Nova Institution | Dental suite | Sinclair Dental | LUX115V | |
| Natus Neurology Elite - Doppler | | Nova Institution | | Natus | 200 | ECO1776 |

| | | | | | | |
|------------------------------|--------------|------------------|-------------------------|--------------|----------------|-------------------|
| Nebulizer | 250-001559 | Nova Institution | Ax room | Sunmark | 4075 | 045143 |
| Nebulizer | 250-001565 | Nova Institution | Secure Unit | Sunmark | 4075 | 040471 |
| Ophthalmic Stand | 25000-003997 | Nova Institution | Dr office | Marco | Deluxe II | 1310499 |
| Optometry Recline Chair | 25000-003999 | Nova Institution | Dr office | Axis Medical | CS-418 | CP-PZY-0418AJ |
| Oximeter | | Nova Institution | Ax room | black | | |
| Oximeter | 250-001117 | Nova Institution | Emerg kit | | NPB40 | G03860573 |
| Oximeter | 250-001225 | Nova Institution | Secure Unit | Rad-5V | | 538887 |
| Oxygen tank regulator | | Nova Institution | Emerg kit - stockroom | Medline | | |
| Oxygen tank regulator | | Nova Institution | H/C Unit emerg kit | Medline | | |
| Panoramic x-ray unti | 250-001438 | Nova Institution | Dental suite | Gendex | 8500DDE | 35-1858141DP |
| Physio bed | 250-001460 | Nova Institution | Tx room | Triton | TRT-600 | 2217 |
| Pneumatic Exam Stool | | Nova Institution | Dr office | Brewer | none | 61609 |
| Pulse rate monitor | 250-001462 | Nova Institution | Ax room | Dynamap | DPC400N | 2019205-001 |
| Pulse rate monitor | 250-001265 | Nova Institution | Secure Unit | Dynamap | DCP400N | SDT10110608SP |
| Scale – Adult | 250-001556 | Nova Institution | Inmate Bathroom | Detecto | | |
| Scale – infant | 250-001573 | Nova Institution | Tx Room | Detecto | | |
| Slit Lamp Microscope | 25000-003996 | Nova Institution | Dr office | Huvitz | HS-5000 | 5HS0001310002 |
| Speculum light with charger | | Nova Institution | Ax Room | Welch Allyn | 739 series | 20100720/20100714 |
| Sphygmomanometer | | Nova Institution | Tx Room | TYCOS | | 90121134041 |
| Sphygmomanometer | | Nova Institution | Ax Room | TYCOS | | 9508090077 |
| Sphygmomanometer | | Nova Institution | Emerg kit- stockroom | Welch Allyn | analog | |
| Sphygmomanometer | | Nova Institution | Emerg kit - secure unit | Welch Allyn | analog | |
| Stretcher | 250-001457 | Nova Institution | Tx room | Hill Rom | P8000 | K043AQ4641 |
| Telehealth control unit | | Nova Institution | Tx Room | Dr. Camscope | DCSM(XGA) | DCSM-10G0169 |
| Telehealth ophthal/autoscope | | Nova Institution | Tx Room | Dr. Camscope | DCS-104T(NTSC) | DCSN-10E-285 |
| Telehealth stethoscope | | Nova Institution | Tx Room | Dr. Camscope | AMD-3700 | 21150 |
| Telehealth ultrasound wand | 20100-004578 | Nova Institution | Tx Room | Dr. Camscope | DCS-P101 | DCSN-08B-341 |

| | | | | | | |
|----------------------------|--------------|------------------|-----------------------|------------|------------------|------------------|
| TENS machine | 250-001575 | Nova Institution | Tx room | Physiotens | | 5330588 |
| TENS machine | 250-001574 | Nova Institution | Tx room | Physiotens | | 447523 |
| TENS machine | | Nova Institution | Tx Room | PhysioSD | TS-1211 | 11-049683 |
| TENS machine | | Nova Institution | Tx Room | PhysioSD | TS-1211 | 11-049685 |
| TENS Machine | | Nova Institution | | PhysioTens | | S330585 |
| Therapeutic ultrasound | 250-001460 | Nova Institution | Tx room | Intelect | 2762CC | 4948 |
| Thermometer | 250-001553 | Nova Institution | Tx Room | Braun | Thermoscan Plus | 10105478202 |
| Thermometer | 250-001554 | Nova Institution | Ax Room | Braun | Thermoscan 6022 | 11110850664 |
| Thermometer | 250-001563 | Nova Institution | Emerg kit - stockroom | Braun | 6022 | 11110890735 |
| Thermometer | 250-001567 | Nova Institution | Secure Unit emerg kit | Braun | Thermoscan | 11110846300 |
| Thermometer | | Nova Institution | Secure Unit | Braun | 6013 | 10100054947 |
| Traction unit | 250-001460 | Nova Institution | Tx room | Chatanooga | 2841 | 3317 |
| Ultraclave | 250-001455 | Nova Institution | Dental suite | Ritter | Midmark M9-022 | V460343 |
| Ultrasonic cleaner (large) | 250-001454 | Nova Institution | Dental suite | Sulton | Prosonic 1000MT | QDG020710129E |
| Ultrasonic cleaner (small) | 250-001444 | Nova Institution | Dental suite | Densply | Cavitron GEN-131 | 9788 |
| Vibrator | | Nova Institution | Dental suite | Buffalo | No. 1A | |
| Visual Acuity Kit | 25000-003998 | Nova Institution | Dr office | Huvitz | HDC9000N | 9HC00N1310009 |
| Wheelchair (regular) | 250-001570 | Nova Institution | Tx Room | Medline | | |
| Wheelchair (regular) | | Nova Institution | Stockroom | Airgo | Procare IC | |
| Wheelchair (wide) | 250-001459 | Nova Institution | Healthcare Unit | AMG | 700-631 | 100050-0512-0026 |
| X-ray arm | 250-001453 | Nova Institution | Dental suite | Gendex | Expert DC | 12-1852698DP |

Appendix I – Equipment List Regional Pharmacy

| Item | ~~Equipment Code~~ Code d'équipement~~ | ~~Institution~~ Établissement~~ | ~~Location~~ Localisation~~ | ~~Manufacturer~~ Nom du fabricant~~ | ~~Model~~ Modèle~~ | ~~Serial Number~~ Numéro de série~~ |
|------|-------------------------------------------|------------------------------------|--------------------------------|----------------------------------------|-----------------------|----------------------------------------|
|------|-------------------------------------------|------------------------------------|--------------------------------|----------------------------------------|-----------------------|----------------------------------------|

| | | | | | | |
|----------------------------------------|-------------|-------------------|----------------------|--------------------|------------------|-----------------------|
| Analyzer - Gas | | Regional Pharmacy | | TSI | 8038 | |
| BIISTER PACKAGER - Automed Heat Sealer | 201-0045554 | Regional Pharmacy | Packaging Room | Healthmark | Autogen | AG-009C |
| BIISTER PACKAGER - Automed Heat Sealer | 201-004550 | Regional Pharmacy | Packaging Room | Healthmark | Autogen | AG-003C |
| BIISTER PACKAGER - Automed Heat Sealer | 201-004552 | Regional Pharmacy | Packaging Room | Healthmark | MTS -350 | 350-TC-004 |
| Blister Packager - Manual HEAT SEALER | 210-005023 | Regional Pharmacy | Shipping / Receiving | Hunt Corporation | Bienfang | MH-00804 |
| COUNTING MACHINE | 201-004553 | Regional Pharmacy | Dispensary | McKesson | Parata Accucount | NTEP CC No. 96-116A3P |
| REFRIGERATOR | 201-005024 | Regional Pharmacy | Methadone Room | True Food Services | TS-49 - True | 4978832 |
| REFRIGERATOR | 201-005025 | Regional Pharmacy | Dispensary | True Food Services | TS-49 - True | 4947312 |
| REFRIGERATOR | 201-004595 | Regional Pharmacy | Basement | ESBE Scientific | SAN-MPR1411 | 10020036 |
| REPEATER PUMP | 201-004551 | Regional Pharmacy | Methadone Room | Baxa | Baxa | 2952 |
| REPEATER PUMP | 201-004555 | Regional Pharmacy | Methadone Room | Baxa | Baxa | 2039 |
| Scale | | Regional Pharmacy | | PTA | Accu Count | |

Appendix J – Equipment List Regional Headquarters

| Item | ~Equipment Code~Code d'équipement~ | ~Institution~Établissement~ | ~Location~Localisation~ | ~Manufacturer~Nom du fabricant~ | ~Model~Modèle~ | ~Serial Number~Numéro de série~ |
|-------------------|------------------------------------|-----------------------------|-------------------------|---------------------------------|----------------|---------------------------------|
| Camera Controller | | Regional Headquarters | | | | |

| | | | | | | |
|----------------------------|--------------|-----------------------|---------------------------------------|-----------------|----------|--------------|
| General Video Exam Scope | 20100-004579 | Regional Headquarters | | Sometech | 456384 | DCSM-11C-015 |
| IntraOral Hand Piece DCS-P | 20100-004578 | Regional Headquarters | | Sometech | 456383 | DCSN-08B-341 |
| Regrigerator (SAN-MPR1411) | 20100-004597 | Regional Headquarters | Currently in storage (basement @ RHQ) | ESBE Scientific | 458353 | 40202019 |
| Stethoscope - Electronic | | Regional Headquarters | | | AMD-3700 | |
| Stethoscope - Electronic | | Regional Headquarters | | | AMD-3700 | |
| Telephonic Stethoscope AMD | 20100-004547 | Regional Headquarters | | AMD | 448659 | SN 21151 |
| Telephonic Stethoscope AMD | 20100-004548 | Regional Headquarters | | AMD | 448660 | SN 22150 |