

**RETURN BIDS TO:**  
**RETOURNER LES SOUMISSIONS À:**  
Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3  
Bid Fax: (613) 545-8067

## SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

### Comments - Commentaires

Vendor/Firm Name and Address  
Raison sociale et adresse du  
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution  
Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3

<b>Title - Sujet</b> Hospital Bed	
<b>Solicitation No. - N° de l'invitation</b> 21401-151857/A	<b>Amendment No. - N° modif.</b> 002
<b>Client Reference No. - N° de référence du client</b> 21401-15-1857	<b>Date</b> 2015-08-14
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$KIN-508-6632	
<b>File No. - N° de dossier</b> KIN-5-44043 (508)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2015-08-26</b>	<b>Time Zone</b> Fuseau horaire Eastern Daylight Saving Time EDT
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Weaver, Tammy	<b>Buyer Id - Id de l'acheteur</b> kin508
<b>Telephone No. - N° de téléphone</b> (613) 484-1809 ( )	<b>FAX No. - N° de FAX</b> (613) 545-8067
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

Instructions: See Herein

Instructions: Voir aux présentes

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

Solicitation No. - N° de l'invitation

21401-151857/A

Client Ref. No. - N° de réf. du client

21401-15-1857

Amd. No. - N° de la modif.

002

File No. - N° du dossier

KIN-5-44043

Buyer ID - Id de l'acheteur

kin508

CCC No./N° CCC - FMS No/ N° VME

---

See attached

Solicitation No. - N° de l'invitation  
21401-151857/A  
Client Ref. No. - N° de réf. du client  
21401-15-1857

Amd. No. - N° de la modif.  
002  
File No. - N° du dossier  
KIN-5-44043

Buyer ID - Id de l'acheteur  
KIN508  
CCC No./N° CCC - FMS No./N° VME

### **AMENDMENT #002**

For the supply of hospital beds for Correctional Service of Canada, Bath Institution, Millhaven, Ontario.

The following question has been submitted:

Q1. Can you please advise what the preferred mattress length is and what the preferred overall length of the bed is?

A1. As long as it follows the specifications listed and my project authority also has stated the following:

Mattresses must fit snugly within the bed to prevent patients from becoming entrapped between the mattress and bed frame. Therefore the mattress should be at most 1-1.5" less in length than the interior bed frame length. This will still allow the mattress to have sheets and blankets changed with little disruption to nurse's hands however reduce entrapment risk.

Q2. I am trying to determine if the bed width of 40" - 45" and bed weight capacity specifications are mandatory requirements.

A2. The minimum specification was posted with the solicitation min 40" and max 45" and Maximum patient weight of 450 lbs. As such they are what we are looking for and therefore mandatory requirements.

Q3. Will a bed that is 35" wide be considered for this RFP.

A3. Specification requirement is minimum 40" and maximum 45", therefore 35" does not fall within the requirement.