

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving - PWGSC / Réception des soumissions
- TPSGC
11 Laurier St. / 11, rue Laurier
Place du Portage, Phase III
Core 0B2 / Noyau 0B2
Gatineau, Québec K1A 0S5
Bid Fax: (819) 997-9776

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Electrical & Electronics Products Division
11 Laurier St./11, rue Laurier
7B3, Place du Portage, Phase III
Gatineau, Québec K1A 0S5

Title - Sujet SPECTRUM ANALYZER	
Solicitation No. - N° de l'invitation W8486-162773/A	Amendment No. - N° modif. 004
Client Reference No. - N° de référence du client W8486-162773	Date 2015-08-28
GETS Reference No. - N° de référence de SEAG PW-\$\$HN-323-67655	
File No. - N° de dossier hn323.W8486-162773	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2015-09-11	Time Zone Fuseau horaire Eastern Daylight Saving Time EDT
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Dubeau, Stéphane	Buyer Id - Id de l'acheteur hn323
Telephone No. - N° de téléphone (819) 956-1533 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment is raised 1) to answer questions from industry and 2) to correct NSN number in section Line Item detail section as follows:

1) to answer questions from industry

Question 1

For requirement h: Would an RF power measurement range of -67 dBm to +23 dBm with a frequency range of 10MHz to 10 GHz min be acceptable?

Answer 1

Yes. We have noticed that para h. in the Critical Requirement List has a typo error. It referenced +40dBm. It was supposed to read +20dBm so this is definitely acceptable.

Question 2

For requirement i: Would a Tracking Generator with a frequency range up to 8 GHz and an output level up to 0 dBm be acceptable?

Answer 2

Yes. Para i. the Critical Requirement is incorrect and is likely not achievable by anyone. The words "Independent signal source" should not have been there at all and the 20 dBm reference was a typo. It should have read "0 dBm".

This sentence should have been referring to a "Tracking Generator" for sure. This sentence needs to be deleted in its entirety and the following inserted:

Tracking Generator capability: Frequency range of 5 MHz to 8 Ghz minimum with output power level up to 0 dBm minimum.

Question 3

I think there is a problem with your invitation to tender. In your Annex B you request that the product meets the following specifications:

h. RF Power meter: Must include peak and average power meter (min -60 dBm to +40 dBm up to 10 Ghz).

However, the model down here that you ask for in your invitation to tender does not even meet this specification!!!!

4 NSN - NNO: 6625-12-360-4684
SENSOR, CURRENT

•

The bidder is requested to
indicate the manufacturer and the
part number offered below.

•

Proposed manufacturer: _____

•

Proposed Part: _____

•

NSCM/CAGE - COF/CAGE: D0894

Part No. - N° de la partie:

1165.1909.02

R&SFSH-Z18 Order No. 1165.1909.02

Power Sensor

10 MHz to 18 GHz

http://www.rohde-schwarz.us/en/product/fshz18-productstartpage_63493-8610.html

Response 3

We have noticed that para h. in the Critical Requirement List has a typo error. It referenced +40dBm as you indicated but it was supposed to read +20dBm. This is corrected to read “min –60 dBm to +20 dBm up to 10 Ghz”.

It should also be noted that we have noticed a problem with para i. Delete in its entirety. It is corrected to read:

Tracking Generator capability: Frequency range of 5 MHz to 8 Ghz minimum with output power level up to 0 dBm minimum.

2) to correct NSN number in section Line Item detail section as follows:

At page 7 of 8, line item detail section, item 007 on original solicitation dated 2015-07-14

Delete: NSN : 6625-12-339-5760

Insert: NSN : 6625-12-399-5760

All other terms and conditions remain unchanged.