

**RETURN BIDS TO:**  
**RETOURNER LES SOUMISSIONS À:**  
**Bid Receiving - PWGSC / Réception des soumissions**  
**- TPSGC**  
**11 Laurier St./ 11 rue, Laurier**  
**Place du Portage, Phase III**  
**Core 0B2 / Noyau 0B2**  
**Gatineau, Québec K1A 0S5**  
**Bid Fax: (819) 997-9776**

## **SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

### **Comments - Commentaires**

**Vendor/Firm Name and Address**  
**Raison sociale et adresse du**  
**fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Scientific, Medical and Photographic Division /  
Division de l'équipement scientifique, des produits  
photographiques et pharmaceutiques  
11 Laurier St./ 11 rue, Laurier  
6B1, Place du Portage  
Gatineau, Québec K1A 0S5

<b>Title - Sujet</b> AUTOCLAVE STERILIZER	
<b>Solicitation No. - N° de l'invitation</b> 39903-160214/A	<b>Amendment No. - N° modif.</b> 003
<b>Client Reference No. - N° de référence du client</b> 39903-160214	<b>Date</b> 2015-09-10
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$\$PV-916-67669	
<b>File No. - N° de dossier</b> pv916.39903-160214	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2015-09-15</b>	
<b>Time Zone</b> <b>Fuseau horaire</b> Eastern Daylight Saving Time EDT	
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Emond, Linda A.	<b>Buyer Id - Id de l'acheteur</b> pv916
<b>Telephone No. - N° de téléphone</b> (819) 956-4014 ( )	<b>FAX No. - N° de FAX</b> (819) 956-3814
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

Solicitation No. - N° de l'invitation

39903-160214/A

Amd. No. - N° de la modif.

003

Buyer ID - Id de l'acheteur

pv916

Client Ref. No. - N° de réf. du client

39903-160214

File No. - N° du dossier

pv91639903-160214

CCC No./N° CCC - FMS No/ N° VME

---

This page has been left blank intentionally.

This amendment is raised to publish all answers to the questions received

- 1) With the bid being extended into September, what is the new required delivery date? **March 1<sup>st</sup> 2016**
- 2) As per answers #4 and #5 below, it is now our understanding that CFIA requires a double door pass-through type sterilizer autoclave for use in a level 3 environment, however there was no indication of this being either a double door or at Level 3 application in the provided specification of the tender documents. Please confirm this, and if so, we would like to request an addendum to the tender specifications with this new information, detailing your requirement for:
  - Level 3 compliance **As mentioned, double door autoclave equipped with interlocking doors and lighting system indicating status, this is standard for any CI#3 as per CBSG**
  - Bioseal specification **Neoprene, there is no particular specification only that it provides an approved bioseal from containment to non containment**
  - Effluent sterilization requirements: **None, the effluent once leaving the autoclave drain is piped to an ETS**
- 3) Control Panel requirements: **on both, loading side and clean side**
- 4) As there is no 3 phase power source close to the autoclave, we can propose to extend an existing 3 phase source to sufficient proximity of the sterilizer, is this possible? **No**
- 5) A 3 phase power source is required in order to take advantage of our vacuum pump feature which can reduce water consumption per cycle by as much as 75% over standard sterilizers and water ejector systems. Please provide us with the distance from the existing 3 phase source to the sterilizer's planned location. **We are not entertaining any 3 phase additions to our Main labs electrical system.**

All other terms and conditions remain unchanged