

RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Public Works and Government Services Canada
ATB Place North Tower
10025 Jasper Ave./10025 ave. Jasper
5th floor/5e étage
Edmonton
Alberta
T5J 1S6
Bid Fax: (780) 497-3510

SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada
ATB Place North Tower
10025 Jasper Ave./10025 ave Jasper
5th floor/5e étage
Edmonton
Alberta
T5J 1S6

Title - Sujet Oscilloscope	
Solicitation No. - N° de l'invitation W7702-165738/A	Amendment No. - N° modif. 003
Client Reference No. - N° de référence du client W7702-165738	Date 2015-10-05
GETS Reference No. - N° de référence de SEAG PW-\$EDM-099-10535	
File No. - N° de dossier EDM-5-38105 (099)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2015-10-08	Time Zone Fuseau horaire Mountain Daylight Saving Time MDT
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Leslie, Sandra	Buyer Id - Id de l'acheteur edm099
Telephone No. - N° de téléphone (780) 616-2057 ()	FAX No. - N° de FAX (780) 497-3510
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: Defence R & D Canada Suffield Bldg 560 Receiving Ralston, AB T0J 2N0 Canada	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment No. 003 is being issued to answer potential bidder's questions:

QUESTIONS AND ANSWERS

Question No 1: Is there is any part number or NSM number leads to this Oscilloscope?

Answer No. 1: No, there is no specific model we are requesting.

Question No. 2: On Page 10 M7 regarding input sensitivity; Is 2mV/Div with high Zoom capability acceptable?

Answer No. 2: No, M7 input sensitivity range, specifies the range of input sensitivity or "zoom" from less than or equal to 1mV/div to greater than or equal to 1V/div.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME