



**RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:**

**Bid Receiving
PWGSC
33 City Centre Drive
Suite 480C
Mississauga
Ontario
L5B 2N5
Bid Fax: (905) 615-2095**

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

**Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution
Public Works and Government Services Canada
Ontario Region
33 City Centre Drive
Suite 480
Mississauga
Ontario
L5B 2N5

| | |
|---|--|
| Title - Sujet Material Mechanical Testing System | |
| Solicitation No. - N° de l'invitation 01E91-160370/A | Amendment No. - N° modif. 002 |
| Client Reference No. - N° de référence du client 01E91-160370 | Date 2016-01-08 |
| GETS Reference No. - N° de référence de SEAG PW-\$TOR-031-6976 | |
| File No. - N° de dossier TOR-5-38124 (031) | CCC No./N° CCC - FMS No./N° VME |
| Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2016-01-14 | |
| Time Zone Fuseau horaire Eastern Standard Time EST | |
| F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/> | |
| Address Enquiries to: - Adresser toutes questions à: Schmidt, Jeff | Buyer Id - Id de l'acheteur tor031 |
| Telephone No. - N° de téléphone (905) 615-2058 () | FAX No. - N° de FAX (905) 615-2060 |
| Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: | |

Instructions: See Herein

Instructions: Voir aux présentes

| | |
|--|--|
| Delivery Required - Livraison exigée | Delivery Offered - Livraison proposée |
| Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur | |
| Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur | |
| Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie) | |
| Signature | Date |

Solicitation No. - N° de l'invitation
01E91-160370/A
Client Ref. No. - N° de réf. du client
01E91-160370

Amd. No. - N° de la modif.
002
File No. - N° du dossier
TOR-5-38124

Buyer ID - Id de l'acheteur
tor031
CCC No./N° CCC - FMS No./N° VME

Amendment 002 has been issued to answer bidder questions and make changes to the solicitation.

At Page 1, Solicitation Closes

Delete: 2016-01-11
Insert: 2016-01-14

Bidder Questions:

Q1: In regards to Mandatory Technical Specification 2.4, The MMTS must have an extensometer. Can Agriculture and Agri-Food Canada (AAFC) provide the following information:

- i. Type of extensometer (contact like a strain gauge extensometer, or non-contact like video extensometer)
- ii. Dimension of samples to be tested.
- iii. Gauge length
- iv. Elongation

R1: The extensometer should have the following specs:

- i. Type of extensometer - either non-contact video extensometer or the clip-on extensometer will meet AAFC's requirement;
- ii. Dimension of samples (thickness) to be tested – This will be varied. For example: from 4 mm to 14 mm;
- iii. Gauge length – This will be varied. For example: from 7 mm to 50 mm;
- iv. Elongation – This will be varied depending on the formulation and result from AAFC's research.

Q2: In regards to Mandatory Technical Specification 2.10 Position accuracy (must be less or equal than): $\leq \pm 0.02$ millimeter (mm) or $\pm 0.06\%$ of displacement. Would position precision of $\leq \pm 0.01$ mm when indicated value is below 10mm and 0.1% when above?

R2: As stated in the Mandatory Technical Specification, the Position accuracy (must be less or equal than): $\leq \pm 0.02$ millimeter (mm) or $\pm 0.06\%$ of displacement.