



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

**Bid Receiving - PWGSC / Réception des soumissions
- TPSGC**
11 Laurier St. / 11, rue Laurier
Place du Portage, Phase III
Core 0B2 / Noyau 0B2
Gatineau, Québec K1A 0S5
Bid Fax: (819) 997-9776

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Vehicles & Industrial Products Division
11 Laurier St./11, rue Laurier
7A2, Place du Portage, Phase III
Gatineau, Québec K1A 0S5

Title - Sujet 15 HP VSD AIR COMPRESSOR	
Solicitation No. - N° de l'invitation HT234-151102/A	Amendment No. - N° modif. 007
Client Reference No. - N° de référence du client HT234-151102	Date 2016-01-27
GETS Reference No. - N° de référence de SEAG PW-\$\$HP-404-68320	
File No. - N° de dossier hp404.HT234-151102	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2016-02-11	
Time Zone Fuseau horaire Eastern Standard Time EST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Michele Mak	Buyer Id - Id de l'acheteur hp404
Telephone No. - N° de téléphone (873) 469-3338 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation
HT234-151102/A

Amd. No. - N° de la modif.
007

Buyer ID - Id de l'acheteur
hp404

Client Ref. No. - N° de réf. du client
HT234-151102

File No. - N° du dossier
hp404HT234-151102

CCC No./N° CCC - FMS No./N° VME

Amendment n° 007 is raised to amend Annex "A" :

Delete Item 15: Installation in its entirety. Annex "A", Revision 2 attached.

All other terms and conditions remain unchanged.

Scope of Work:

The Banting Retrofit is a multiyear project with the purpose of relocating the entire Environmental Health Center (EHC) laboratories into two new retrofitted lab floors within the Sir Frederick Banting building. The air compressors will be attached to nitrogen generator modules to provide the lab floors with a centralized nitrogen distribution system.

	Requirement / Description	Met	Not Met
1	Supply Two 15 HP Variable Speed Drive Air Compressors <ul style="list-style-type: none"> Required to meet the requirements of the Nitrogen Gas Generator 		
2	Complete with Integral air dryers and both filters		
3	Each Compressor should not exceed the following dimensions- Length: 38in /Width: 27in /Height:47in /Weight: 827 lb		
4	The Compressor must have a voltage of 575 VAC		
5	Adequate Air Dryers with a Dryer Refrigerant dew point of no more than 35°F. <ul style="list-style-type: none"> Adequate drying and filtration is required to ensure no condensable water and oil in the output stream 		
6	Capacity of CFMs: 18.4 - 50.4 CFM's at 145 psi		
7	The Compressor system must have suitable Filters; the Coalescing type with a final efficiency of 0.1u and auto drains <ul style="list-style-type: none"> These are required to complete the removal of oil and H2O prior to entering the N2 Generator 		
8	The Compressor system must include the following Receiver Tank Version- 80 Gallon Tank Mounted Receiver Tank with the following dimensions: Length-70in /Width-27in /Height-67in /Weight: 264.5 lb		
9	The Compressor must have a voltage of 575 VAC		
10	A X41 System Controller for the two compressors <ul style="list-style-type: none"> This is required to control and monitor the compressors function. 		
11	A 120 Vertical air receiver tank for nitrogen storage; to include all necessary components, back pressure regular and solenoid valves. <ul style="list-style-type: none"> Act as a buffer for unforeseen flow demands. 		

12	The system must possess a POLY-SEP PSG-15 oil/water condensate system <ul style="list-style-type: none"> This is required to control all the condensate from the systems components. To ensure no oil exits into municipal drains 		
13	Compressor needs to be 140 PSI output. (Max. pressure = 65-135 psi)		
14	English Manuals / 2 copies		
16	Labor and commissioning for start-up of a completed Compressed Air System		
18	Service Response Time <ul style="list-style-type: none"> 24 hours minimum to 48 hours maximum 		