



**RETURN BIDS TO:**

**RETOURNER LES SOUMISSIONS À:**

Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3  
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**  
Raison sociale et adresse du  
fournisseur/de l'entrepreneur

**Issuing Office - Bureau de distribution**  
Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3

<b>Title - Sujet</b> Spectrum Analyzer	
<b>Solicitation No. - N° de l'invitation</b> W3474-162811/A	<b>Amendment No. - N° modif.</b> 003
<b>Client Reference No. - N° de référence du client</b> W3474-16-2811	<b>Date</b> 2016-02-16
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$KIN-710-6807	
<b>File No. - N° de dossier</b> KIN-5-44142 (710)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2016-02-29</b>	
<b>Time Zone</b> Fuseau horaire Eastern Standard Time EST	
<b>F.O.B. - F.A.B.</b> Specified Herein - Précisé dans les présentes <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input type="checkbox"/> <b>Other-Autre:</b> <input checked="" type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Dunphy, Ken	<b>Buyer Id - Id de l'acheteur</b> kin710
<b>Telephone No. - N° de téléphone</b> (613) 449-5116 ( )	<b>FAX No. - N° de FAX</b> (613) 545-8067
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> Raison sociale et adresse du fournisseur/de l'entrepreneur	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> (type or print) <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

Solicitation No. - N° de l'invitation  
W3474-162811  
Client Ref. No. - N° de réf. du client  
W3474-16-2811

Amd. No. - N° de la modif.  
003  
File No. - N° du dossier  
KIN-5-44142

Buyer ID - Id de l'acheteur  
KIN710  
CCC No./N° CCC - FMS No./N° VME

---

**This amendment is being issued in response to Bidder's questions and Canada's responses.**

Q1. Does the Client already have an external noise source they plan to use when making the measurements? A noise source is needed for any Spectrum/Signal analyzer with a Firmware personality for making noise figure/gain measurements. If the Client does not have an external source, then we request further clarification to be sure we are in compliance with our bid. If a noise source is required, is it required as part of this bid, and if so, what specifications (frequency range, ENR figure, connector type) are required in order to be compliant? To just say that a signal analyzer can make noise figure measurements as an upgrade via SW key only is not true. An external noise source is always needed.

A1. We do not own an external noise source at the moment. It does not have to be included in the bid. It is an item we will buy later when required.