



**RETURN BIDS TO:**

**RETOURNER LES SOUMISSIONS À:**

Public Works and Government Services Canada  
ATB Place North Tower  
10025 Jasper Ave./10025 ave. Jasper  
5th floor/5e étage  
Edmonton  
Alberta  
T5J 1S6  
Bid Fax: (780) 497-3510

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**  
**Raison sociale et adresse du**  
**fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Public Works and Government Services Canada  
ATB Place North Tower  
10025 Jasper Ave./10025 ave Jasper  
5th floor/5e étage  
Edmonton  
Alberta  
T5J 1S6

<b>Title - Sujet</b> Disk Stack Separation System		
<b>Solicitation No. - N° de l'invitation</b> 31029-169942/A		<b>Amendment No. - N° modif.</b> 001
<b>Client Reference No. - N° de référence du client</b> 31029-169942		<b>Date</b> 2016-06-27
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$EDM-100-10776		
<b>File No. - N° de dossier</b> EDM-6-39014 (100)	<b>CCC No./N° CCC - FMS No./N° VME</b>	
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2016-07-13</b>		<b>Time Zone</b> Fuseau horaire Mountain Daylight Saving Time MDT
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>		
<b>Address Enquiries to: - Adresser toutes questions à:</b> Fisher, Andrew		<b>Buyer Id - Id de l'acheteur</b> edm100
<b>Telephone No. - N° de téléphone</b> (780) 901-4270 ( )		<b>FAX No. - N° de FAX</b> (780) 497-3510
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b> National Research Council 148 – 111 Research Drive Saskatoon, SK S7N 3R2		

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

Solicitation No. - N° de l'invitation  
31029-169942/A  
Client Ref. No. - N° de réf. du client  
31029-169942

Amd. No. - N° de la modif.  
File No. - N° du dossier  
EDM-6-39014

Buyer ID - Id de l'acheteur  
EDM100  
CCC No./N° CCC - FMS No./N° VME

**This amendment has been raised to respond to the following question and make the following revisions:**

**Question:** Under *Sanitation*, “Clarifier must allow maintenance of sterility”, does the unit need to have the ability to be steam sterilized or will chemical sanitization suffice?

**Answer:** Both steam and chemical sterilization will work.

**All other terms and conditions remain the same.**