



RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
Room 100,
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6
Bid Fax: (204) 983-0338

SOLICITATION AMENDMENT

MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada -
Western Region
Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6

Title - Sujet Microtome Cryostat	
Solicitation No. - N° de l'invitation K4E21-160008/A	Amendment No. - N° modif. 003
Client Reference No. - N° de référence du client K4E21-160008	Date 2016-08-24
GETS Reference No. - N° de référence de SEAG PW-\$WPG-119-9940	
File No. - N° de dossier WPG-6-39057 (119)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2016-09-07	Time Zone Fuseau horaire Central Daylight Saving Time CDT
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Zdan, Tyler	Buyer Id - Id de l'acheteur wpg119
Telephone No. - N° de téléphone (204) 509-5743 ()	FAX No. - N° de FAX (204) 983-7796
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: Environment and Climate Change Canada, CWS 11455 Saskatchewan Drive, University of Alberta CW405 BioSciences Building Edmonton, AB, T6G 2E9	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment, No. 003, is raised to address a question regarding Solicitation K4E21-160008/A as follows:

Question: Can you change the specification about chamber cooling to be cooling for each component?

Answer: A specific minimum temperature for each component is not required. However, specification 3.2 has been modified to include a minimum specimen temperature. Please amend specification 3.2 as follows:

Delete: Minimum cooling chamber temperature of -35°C.

Insert: Minimum cooling chamber/specimen temperature of -35°C.

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED.