



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

| | |
|--|--|
| Title - Sujet DEFIBRILLATORS | |
| Solicitation No. - N° de l'invitation H3618-162811/A | Amendment No. - N° modif. 001 |
| Client Reference No. - N° de référence du client H3618-16-2811 | Date 2016-11-22 |
| GETS Reference No. - N° de référence de SEAG PW-\$KIN-690-7039 | |
| File No. - N° de dossier KIN-6-46089 (690) | CCC No./N° CCC - FMS No./N° VME |
| Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2016-12-12 | |
| Time Zone Fuseau horaire Eastern Standard Time EST | |
| F.O.B. - F.A.B. | |
| Plant-Usine: <input checked="" type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/> | |
| Address Enquiries to: - Adresser toutes questions à: Rider, Kim (Buyer) | Buyer Id - Id de l'acheteur kin690 |
| Telephone No. - N° de téléphone (613) 449-4531 () | FAX No. - N° de FAX (613) 545-8067 |
| Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: | |

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Instructions: See Herein

Instructions: Voir aux présentes

| | |
|---|--|
| Delivery Required - Livraison exigée | Delivery Offered - Livraison proposée |
| Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur | |
| Telephone No. - N° de téléphone Faximile No. - N° de télécopieur | |
| Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie) | |
| Signature | Date |

Issuing Office - Bureau de distribution

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Solicitation No. - N° de l'invitation
H3618-162811/A
Client Ref. No. - N° de réf. du client
H3618-16-2811

Amd. No. - N° de la modif.
001
File No. - N° du dossier
KIN-6-46089

Buyer ID - Id de l'acheteur
KIN690
CCC No./N° CCC - FMS No./N° VME

Question #1

The requirement due in March is for: delivery only, or is invoicing required as well? Please specify.

Answer# 1

Invoicing and delivery is required before the end of the 2016-17 fiscal year end- at least 1 week before March /31st/2017 to allow for processing

Question #2

Is the expectation to have a partial or 100% invoicing, and at which stage (placement of the order, delivery or installation)? Please specify.

Answer#2

We require a complete invoice by the end of the 2016-17 fiscal year- at least 1 week before March 31/2017.

Question #3 Page 10 of 16

1)Requirement“Defibrillation equipment that will allow the delivery of emergency/urgent and routine health care in remote and isolated settings (Nursing Stations) in Ontario, Canada. The defibrillation equipment and accessories must be convertible(adaptable) with current emergency transportation system(i.e. air ambulance- Ornge)”.

Answer #3: We will remove this statement.

Question #4- “Output Current (mA) suggested minimum : 0 - 140 mA”

Q - We would ask that this specification be changed to minimum (10-140mA).There is no clinical evidence that supports that pacing a patient between 0-10mA would ever provide capture. Most patients can be captured at 70mA or more.

Answer #4: We will maintain the current output current at the indicated minimum of 0-140 mA.

Question #5“CO- ability to add on CO monitoring”

Q - Would the customer accept a stand -alone alternative since CO monitoring is not used on a regular basis?

Answer #5 *We cannot accept a stand-alone CO monitoring system.*

Question #6 - Minimum Energy requirement

The current energy specification notes a starting energy level of 1J. The devices currently installed at SLZH start at 2J and represent local protocol. In addition, 2J is consistent with current guidelines recommendations for weight-based pediatric protocols. Please consider revising energy specification for internal/external defibrillation to start at 2J.

Solicitation No. - N° de l'invitation
H3618-162811/A
Client Ref. No. - N° de réf. du client
H3618-16-2811

Amd. No. - N° de la modif.
001
File No. - N° du dossier
KIN-6-46089

Buyer ID - Id de l'acheteur
KIN690
CCC No./N° CCC - FMS No./N° VME

Answer #6

This will be changed to start at 1 or 2 J

Question #7

The current biphasic energy specification references levels to at least 200J. The current SLZH defibrillator installation offers a wide range of escalating biphasic energy through 360J with supporting local medical protocols. Please consider referencing up to 360J for the energy specification consistent with local devices and protocols.

Answer #7

We will accept energy specification up to 360 J