



RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
Room 100,
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6
Bid Fax: (204) 983-0338

SOLICITATION AMENDMENT

MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada -
Western Region
Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6

| | |
|---|---|
| Title - Sujet Tissue Processor | |
| Solicitation No. - N° de l'invitation 01580-170666/A | Amendment No. - N° modif. 001 |
| Client Reference No. - N° de référence du client 01580-170666 | Date 2016-12-15 |
| GETS Reference No. - N° de référence de SEAG PW-\$WPG-119-10065 | |
| File No. - N° de dossier WPG-6-39165 (119) | CCC No./N° CCC - FMS No./N° VME |
| Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-01-11 | Time Zone Fuseau horaire Central Standard Time CST |
| F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/> | |
| Address Enquiries to: - Adresser toutes questions à: Zdan, Tyler | Buyer Id - Id de l'acheteur wpg119 |
| Telephone No. - N° de téléphone (204) 509-5743 () | FAX No. - N° de FAX (204) 983-7796 |
| Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: Agriculture and Agri-Food Canada Room 2252 52 Campus Drive Saskatoon, SK, S7N 5B4 Saskatoon, SK, S7N | |

Instructions: See Herein

Instructions: Voir aux présentes

| | |
|--|--|
| Delivery Required - Livraison exigée | Delivery Offered - Livraison proposée |
| Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur | |
| Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur | |
| Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie) | |
| Signature | Date |

This amendment, No. 001, is raised to address a question regarding Solicitation 01587-177022/A, as follows:

Question 1 (Annex A: Compliance Matrix, Specification 1.9): We presently are unable to connect a computer monitor and keyboard to the tissue processor. The unit has a built in display that has a touchscreen keyboard and all the necessary software pre-loaded so there really is no need. It can be connected to an LIS if required. Is this acceptable?

Response: This spec is for when the touchscreen display not functioning. We need to have the option to get around this problem. The supplier needs to demonstrate that the system can still be used without a functioning touchscreen display.

Question 2 (Annex A: Compliance Matrix, Specification 1.10): Our unit is capable of storing 16 program instructions, not 20. How imperative is this requirement and will you accept 16 program instructions?

Yes, it will be acceptable. Please delete "20" from Specification 1.10 and insert "16".

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED.