



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Travaux publics et Services gouvernementaux
Canada

Place Bonaventure, portail Sud-Est

800, rue de La Gauchetière Ouest

7^{ième} étage

Montréal

Québec

H5A 1L6

FAX pour soumissions: (514) 496-3822

SOLICITATION AMENDMENT

MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Travaux publics et Services gouvernementaux Canada

Place Bonaventure, portail Sud-Est

800, rue de La Gauchetière Ouest

7^{ième} étage

Montréal

Québec

H5A 1L6

| | |
|---|--|
| Title - Sujet System for filtration | |
| Solicitation No. - N° de l'invitation 31241-165994/A | Amendment No. - N° modif. 001 |
| Client Reference No. - N° de référence du client 31241-16-5994 | Date 2016-12-21 |
| GETS Reference No. - N° de référence de SEAG PW-\$MTA-160-14145 | |
| File No. - N° de dossier MTA-6-39206 (160) | CCC No./N° CCC - FMS No./N° VME |
| Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-01-18 | |
| Time Zone Fuseau horaire Heure Normale du l'Est HNE | |
| F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/> | |
| Address Enquiries to: - Adresser toutes questions à: Asquino, Carolina | Buyer Id - Id de l'acheteur mta160 |
| Telephone No. - N° de téléphone (514) 496-3549 () | FAX No. - N° de FAX (514) 496-3822 |
| Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: | |

Instructions: See Herein

Instructions: Voir aux présentes

| | |
|--|--|
| Delivery Required - Livraison exigée | Delivery Offered - Livraison proposée |
| Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur | |
| Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur | |
| Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie) | |
| Signature | Date |

AMENDMENT 001

This amendment is raised to modify a typo in the Mandatory Technical Criteria Table.

The RFP is amended as follows:

1) Under ANNEX "C" MANDATORY TECHNICAL CRITERIA TABLE

DELETE: In its entirety

INSERT:

The bidder must include with its proposal technical datasheets of the products offered. A "Technical datasheet" is defined as a document, such as a brochure, technical document, a drawing, a test report. This document must provide detailed information for each specification and performance requirement.

The bidder must clearly confirm how the products offered meet each of the mandatory technical criteria. Merely mentioning that it meets the criteria is not enough. The bidder must provide sufficient detail to enable a full assessment.

Any proposal that does not clearly demonstrate compliance with each of the technical requirements listed in the table "Mandatory Technical Criteria" detailed below, will be considered non-responsive and its proposal will be rejected.

| No. | Mandatory Technical Specifications | To assist in the evaluation please indicate where it is in the technical proposal |
|-----|---|---|
| 3.1 | The ATF system for filtration must have a pump module. | |
| 3.2 | The pump module must have a silicone diaphragm, an upper hemisphere (liquid side) with a vent port and a lower hemisphere (air side) with an air inlet assembly. | |
| 3.3 | The ATF system for filtration must have a microfiltration polyethersulfone hollow fiber module (0.2 µm) with upper and lower filtrate ports. It sanitarily attaches to the pump module. | |
| 3.4 | The ATF system for filtration must have an ATF hollow fiber module connection to the bioreactor (a silicone tubing with sanitary tri-clamp ends). | |
| 3.5 | The ATF system for filtration must be supplied with a 10mm headplate compression fitting for Applikon vessel. | |

Solicitation No. - N° de l'invitation
31241-165994/A
Client Ref. No. - N° de réf. du client
31241-16-5994

Amd. No. - N° de la modif.
001
File No. - N° du dossier
MTA-6-39206

Buyer ID - Id de l'acheteur
MTA160
CCC No./N° CCC - FMS No./N° VME

| | | |
|------|---|--|
| 3.6 | The ATF system for filtration must be supplied with a dip tube, 450mm length, triclamp top for 5-15 L bioreactor. | |
| 3.7 | The ATF system for filtration must have a controller equipped with a pump line for connection to the ATF system diaphragm pump module. | |
| 3.8 | The controller must have a second stage air pressure gage and a secondary air pressure regulator. | |
| 3.9 | The controller must have an air inlet port and an exhaust vacuum line connection. | |
| 3.10 | The controller must have an HMI display. | |
| 3.11 | The controller must have software programmed to control the ATF system for filtration (similar to the Repligen program for ATF perfusion system or equivalent). | |

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.