



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Address all Bid Inquiries to:
Tammy.Weaver@pwgsc-tpsgc.gc.ca

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Title - Sujet Floating X-Ray Table	
Solicitation No. - N° de l'invitation 21401-172097/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client 21401-17-2097	Date 2017-02-03
GETS Reference No. - N° de référence de SEAG PW-\$KIN-508-7109	
File No. - N° de dossier KIN-6-46183 (508)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-02-21	
Time Zone Fuseau horaire Eastern Standard Time EST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Weaver, Tammy	Buyer Id - Id de l'acheteur kin508
Telephone No. - N° de téléphone (613) 484-1809 ()	FAX No. - N° de FAX (613) 545-8067
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment #001 – Floating X-Ray Table

The following questions have been submitted in reference to the original solicitation:

Question 1:

It is stated in the technical requirements:

Bucky

_ 103 line, 10:1 ratio 34" – 44" focal range grid (page 11 of RFP)

And in another spot it is stated:

Moving Grid (page 12 of RFP)

_ Dimensions 44 x 48 cm.

_ Number of pairs of lines: 103 lines/inch, 40 lines /cm.

_ Grid ratio 10:1,

_ Focalized at 150 cm.

Please specify which focal range is required (standard for table is 34"- 44")?

Answer:

The Grid ratio should read 12:1. NOT 10:1. The focal range is as reads under subtitle Bucky: 34"-44" focal range grid. The line and ratio (103 line, 10:1 ratio) should be moved to the Moving Grid requirements. Please change the ratio to 12:1

Question 2:

Could you clarify the expectation for the remote control for the table top? Is the expectation for the remote to control all movements (i.e elevation, transversal and longitudinal) of the table? If yes can you please verify whether you are seeking a wired or wireless remote control and the range that the remote should be able to accommodate for?

Answer:

Yes, a remote is a foot pedal control. This controls longitudinal and transverse float and elevation.

Question 3:

Is the remote control for the table top a mandatory requirement? We are not aware of any manufacturer having this option on a radiographic elevating table.

Answer:

Foot control so the table operational movements are hands free is a requirement.

Question 4:

Regarding the Table top Height elevation: will an 87cm Maximum be acceptable (90cm Maximum has been indicated)?

Answer:

Maximum floor-Table distance 90cm

Minimum floor-Table distance 55cm

Solicitation No. - N° de l'invitation
21401-170097/A
Client Ref. No. - N° de réf. du client
21401-17-20097

Amd. No. - N° de la modif.
A001
File No. - N° du dossier
KIN-6-46183

Buyer ID - Id de l'acheteur
KIN508
CCC No./N° CCC - FMS No./N° VME

Question 5:

Regarding the Longitudinal Table top movement: will ± 54 cm be acceptable (± 60 cm has been indicated)?

Answer: Longitudinal 60 cm.

Question 6:

Regarding the Transverse Table top movement: will ± 11.4 cm be acceptable (± 12.5 cm has been indicated)?

Answer: Transversal 12.5 cm

Question 7:

Regarding the Vertical Table top movement: will 30.5cm be acceptable (35cm minimum has been indicated)?

Answer: Vertical 35cm and Longitudinal Bucky travel is 61cm

Table top movements are controlled by a foot pedal controller, so the table top movement is hands free for positioning patients when required.