



RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:
Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
Room 100,
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6
Bid Fax: (204) 983-0338

SOLICITATION AMENDMENT

MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada -
Western Region
Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6

Title - Sujet Miscellaneous Lab Equipment	
Solicitation No. - N° de l'invitation A7100-164083/A	Amendment No. - N° modif. 004
Client Reference No. - N° de référence du client A7100-164083	Date 2017-03-01
GETS Reference No. - N° de référence de SEAG PW-\$WPG-016-10118	
File No. - N° de dossier WPG-6-39269 (016)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-03-03	Time Zone Fuseau horaire Central Standard Time CST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Hall, Marlene	Buyer Id - Id de l'acheteur wpg016
Telephone No. - N° de téléphone (204) 230-0147 ()	FAX No. - N° de FAX (204) 983-7796
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: Wills Transfer Ltd. 3100 Swansea Crescent Ottawa, ON. K1G 3W4 c/o Claudette Weedmark	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

N° de l'invitation - Solicitation No.
A7100-164083/A
N° de réf. du client - Client Ref. No.
A7100-164083

N° de la modif - Amd. No.
004
File No. - N° du dossier
WPG-6-39269

Id de l'acheteur - Buyer ID
wpg016
N° CCC / CCC No./ N° VME - FMS

This **amendment # 004** is raised to provide Canada's response to Supplier's question to **Solicitation # A7100-164083/A** as follows:

- Q1.** Item 2 – Ultra-purified ASTM Type I water dispenser, 120L/h (2LPM) capacity. What capacity of production of Type I water is needed?
- A1.** Refer to item 2.2 Production: 2LPM \pm 0.2L (0,5GPM \pm 0,05GPM)
- Q2.** Item 18 – Clean bench laminar flow hood, Class ISO100, 1200MM (4ft) wide work surface. In our product, all of the controls are housed. Is this acceptable?
- A2.** Control board was requested non-proprietary in order to ensure that a replacement could be easily available considering the facility location and the complexity of the supply route. In the context where an in-house control board is proposed, it will be deemed acceptable when a spare replacement control board is included in the bid offer.
- Q3.** Item 4 – Wheeled necropsy table for large animals w/scissor-action lift, 1135kg (2500lbs) load capacity. The width doesn't look standard. Would a larger width be acceptable?
- A3.** A larger width size is not acceptable as it will affect the spatial organization and use of the necropsy room and there is no possibility for enlarging the room (space).
- Q4.** Item 19.3 and 20.3 – 3 pen drawers under the work surface. Is one stainless steel drawer acceptable?
- A4.** One stainless steel drawer is acceptable if it is 150mm (6 in) high or less.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME