



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Travaux publics et Services gouvernementaux
Canada

Place Bonaventure,
800 rue de la Gauchetière Ouest
Voir aux présentes - See herein
Montréal

Québec
H5A 1L6

FAX pour soumissions: (514) 496-3822

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Travaux publics et Services gouvernementaux Canada
Place Bonaventure,
800 rue de la Gauchetière Ouest
Voir aux présentes - See herein
Montréal
Québec
H5A 1L6

Title - Sujet Chambre à temp altitude humidité	
Solicitation No. - N° de l'invitation 01B30-180100/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client 01B30-18-0100	Date 2017-08-10
GETS Reference No. - N° de référence de SEAG PW-SMTA-405-14419	
File No. - N° de dossier MTA-7-40051 (405)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-08-21	
Time Zone Fuseau horaire Heure Avancée de l'Est HAE	
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Séguin, Caroline	Buyer Id - Id de l'acheteur mta405
Telephone No. - N° de téléphone (514) 496-3734 ()	FAX No. - N° de FAX (514) 496-3822
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

AMENDMENT 002

The purpose of this amendment is to publish questions and answers regarding this request.

The purpose of this amendment is to delete three criteria from the Attachment 1.

A) Questions and Answers

Q1: Can you describe with precision the items that will be tested?

A1: Processed foods or not, with or without packaging.

Q2: When testing foods, will they be packaged or a sealed container?

A2: They could be packaged or not.

Q3: Will you test the quality of the foods?

A3: Possible, depending on the protocols.

Q4: Due to technical constraints (due to the fact that humidity and altitude functions will be used at the same time), it will be impossible to provide a heated window or a wiper. Would you accept an equipment without this function?

A4: The humidity and altitude performances were not asked for simultaneously. We must see the product through the window. This criteria is mandatory.

Q5: In simultaneous humidity and altitude mode, the equipment will not be able to meet the criteria 2.2.10 and the criteria 2.2.9. Will you accept a system that

- i. Altitude range: from 0 to 20 000 feet.
- ii. Humidity range: 30% to 95%.
- iii. Dew point range : between +30°C and 47°C
- iv. The three specifications here above with a maximum of 70°C of dry temperature.

A5: The humidity and altitude performances were not asked for simultaneously.

B) Deleted criteria – Attachment 1 “Mandatory Technical Criteria”

At page 16 of 17,

DELETE ENTIRELY:

- 2.1.3 The system must be supplied with a calibration certificate from a recognized organization like A2LA, American Association for Laboratory Accreditation or any equivalent organization;
- 2.1.4 The unit must have CSA certification;
- 2.1.5 Supplier need to deliver hard copy of the certification confirming unit has been pre-tested at the manufacturer's facility.

N.B These criteria will not be evaluated at the time of bid submission, but will remain contract requirements to be met, as mentioned in Annex A “Requirement”.

All other terms and conditions remain unchanged.