



**RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:**

Bid Receiving
PWGSC
33 City Centre Drive
Suite 480C
Mississauga
Ontario
L5B 2N5
Bid Fax: (905) 615-2095

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada
Ontario Region
33 City Centre Drive
Suite 480
Mississauga
Ontario
L5B 2N5

Title - Sujet Growth Chamber Replacement	
Solicitation No. - N° de l'invitation 01689-180221/A	Amendment No. - N° modif. 004
Client Reference No. - N° de référence du client 01689-180221	Date 2017-09-08
GETS Reference No. - N° de référence de SEAG PW-\$TOR-008-7344	
File No. - N° de dossier TOR-7-40027 (008)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-10-02	
Time Zone Fuseau horaire Eastern Daylight Saving Time EDT	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Leslie, Sandra	Buyer Id - Id de l'acheteur tor008
Telephone No. - N° de téléphone (905) 615-2069 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation
01689-180221 /A
Client Ref. No. - N° de réf. du client
01689-180221

Amd. No. - N° de la modif.
004
File No. - N° du dossier
TOR-7-40027

Buyer ID - Id de l'acheteur
TOR008
CCC No./N° CCC - FMS No./N° VME

Solicitation Amendment 004 is being issued to respond to questions received and to make changes on solicitation No. 01689-180221/A.

QUESTION AND ANSWER:

Question: Could you please advise if the following is acceptable? For lines 20, 35 and 49 in Table B could you confirm if temperature ranges of 4-40°C lights off and 10-40°C lights on are acceptable?

Answer: No, the temperature must reach 44°C as per the specifications.

UNDER ANNEX "A" TECHNICAL SPECIFICATIONS

Section 30 00 10, 1.6 Temperature, .2 Chamber Type A – **.6 Cooling/Refrigeration System**

Delete: .5 Refrigeration system must be charged with CFC-free refrigerant.

Insert: .5 Refrigeration system must be charged with HCFC-free and CFC-free refrigerant.

Section 30 00 10, 1.6 Temperature, .3 Chamber Type B - **.6 Cooling/Refrigeration System**

Delete: .5 Refrigeration system must be charged with CFC-free refrigerant.

Insert: .5 Refrigeration system must be charged with HCFC-free and CFC-free refrigerant.

Section 30 00 10, 1.6 Temperature, .4 Chamber Type C – **.4 Cooling/Refrigeration System**

Delete: .5 Refrigeration system must be charged with CFC-free refrigerant.

Insert: .5 Refrigeration system must be charged with HCFC-free and CFC-free refrigerant.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME AND IN FULL EFFECT