



**RETURN BIDS TO:  
RETOURNER LES SOUMISSIONS À:**

**Bid Receiving  
PWGSC  
33 City Centre Drive  
Suite 480C  
Mississauga  
Ontario  
L5B 2N5  
Bid Fax: (905) 615-2095**

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address  
Raison sociale et adresse du  
fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Public Works and Government Services Canada  
Ontario Region  
33 City Centre Drive  
Suite 480  
Mississauga  
Ontario  
L5B 2N5

<b>Title - Sujet</b> Radiation Flux Sensor Suite	
<b>Solicitation No. - N° de l'invitation</b> K8D22-180455/A	<b>Amendment No. - N° modif.</b> 001
<b>Client Reference No. - N° de référence du client</b> K8D22-180455	<b>Date</b> 2017-09-29
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$TOR-031-7361	
<b>File No. - N° de dossier</b> TOR-7-40035 (031)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2017-10-25</b>	
<b>Time Zone</b> <b>Fuseau horaire</b> Eastern Daylight Saving Time EDT	
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Schmidt, Jeff	<b>Buyer Id - Id de l'acheteur</b> tor031
<b>Telephone No. - N° de téléphone</b> (905) 615-2058 ( )	<b>FAX No. - N° de FAX</b> ( ) -
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

**Amendment 001 has been issued to respond to Bidder questions**

**Q1:** Please confirm whether Environment and Climate Change Canada will complete the installation on their own.

**R1:** Yes, Environment and Climate Change Canada will complete the installation

**Q2:** Is there available power supply already at the site? If so, please describe.

**Q2:** Yes, power can be trenched to the site location. Specifications must match Annex A, Requirement 2.1.1, Must have power requirements between 12 to 24 VDC and an AC power supply option.

**Q3:** Will the two snow depth sensors be installed on the same user-supplied tower mast or pole? What is their required range of measurement?

**R3:** Yes, the snow depth sensors will be installed on the user-supplied tower. Its required range must satisfy standard WMO operational guidelines (the sensor be placed 3 m above ground level during operation).