



**RETURN BIDS TO:
RETOURNER LES SOUMISSIONS A :**

Royal Canadian Mounted Police | Gendarmerie royale du Canada
Bid Receiving | Réception des soumissions
Attention: Anna Rozanski
Mailstop | Arrêt postal 15
73 promenade Leikin Drive
Ottawa, Ontario K1A 0R2 Canada

All persons delivering mail, parcels and bids to the Mail Parcel and Screening Facility will be asked to provide government photo identification and a contact number as part of an enhanced security protocol.

Dans le cadre d'un protocole de sécurité amélioré, toute personne qui livre le courrier, les paquets et les soumissions à l'installation d'inspection du courrier et des colis devra désormais présenter une carte d'identité avec photo émise par le gouvernement et un numéro de téléphone.

SOLICITATION AMENDMENT

MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments: - Commentaries :

Title – Sujet Gas Chromatograph / Mass Spectrometer	Date November 10, 2017	
Solicitation No. – Nº de l'invitation 201802694/A	Amendment No. – Nº de modification: 003	
Client Reference No. - Nº de référence du client 201802694		
Solicitation Closes – L'invitation prend fin		
At – à :	2:00 PM	EST (Eastern Standard Time) HNE (heure normale de l'Est)
On – le :	November 29, 2017	
Delivery - Livraison See herein — Voir aux présentes	Taxes See herein — Voir aux présentes	Duty – Droits See herein — Voir aux présentes
Destination of Goods and Services – Destinations des biens et services See herein — Voir aux présentes		
Instructions See herein — Voir aux présentes		
Address Inquiries to – Adresser toute demande de renseignements à anna.rozanski@rcmp-grc.gc.ca		
Telephone No. – No. de téléphone 613-843-6972	Facsimile No. – No. de télécopieur 613-825-0082	

COMPLETE BELOW IN FULL – REMPLISSEZ CI-DESSOUS EN ENTIER	
Vendor/Firm Name, Address and Representative – Raison sociale, adresse et représentant du fournisseur/de l'entrepreneur:	
Telephone No. – No. de téléphone	Facsimile No. – No. de télécopieur
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) – Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date



This amendment is raised to address the following:

- To respond to questions received during the solicitation period; and
- To revise the solicitation accordingly, as applicable.

QUESTIONS AND ANSWERS

Question 1: Could you please clarify the following points regarding this solicitation:

- a) In ATTACHMENT 1 TO PART 4 – TECHNICAL EVALUATION CRITERIA point B) Point Rated Technical Criteria, “R1 Automatic sample changer combines liquid and headspace injections on one sampler so the user can switch from one application to the other on the same workstation”. And then in Annex A, point 2.1 2.1 “The GC/MS system must include an automatic sample changer that can accommodate at least twenty-five (25) different liquid samples in a standard 2ml screw top vial (12x32mm) not including any liquid injector wash or waste vials”. So, could you please clarify whether the customer wants liquid AND Headspace capabilities, or only liquid sampling?
- b) In ATTACHMENT 1 TO PART 4 – TECHNICAL EVALUATION CRITERIA point B) Point Rated Technical Criteria, “R3 Mass spectrometer includes two chemical ionization sources”. And in ANNEX A – PURCHASE SPECIFICATIONS, point “5.1 The MS must operate in electron impact (EI), positive and negative chemical ionization (CI) modes” and point “5.2 The MS source for both EI and CI must be included”. Could you please clarify whether the customer wants 2x CI sources or 1x CI and 1x EI?

Answer 1: Annex A provides the GC/MS mandatory specifications as referenced at M1 on page 9 of the bid solicitation. The point rated criteria are preferred items which are intended to differentiate bids that meet the mandatory technical criteria.

SOLICITATION REVISIONS

There are no solicitation revisions.