



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Title - Sujet Soft magnetic materials testing sys	
Solicitation No. - N° de l'invitation 23584-180654/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client 23584-18-0654	Date 2017-11-20
GETS Reference No. - N° de référence de SEAG PW-\$KIN-930-7337	
File No. - N° de dossier KIN-7-48127 (930)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-12-18	
Time Zone Fuseau horaire Eastern Standard Time EST	
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Littlefield, Mike	Buyer Id - Id de l'acheteur kin930
Telephone No. - N° de téléphone (613) 545-8058 ()	FAX No. - N° de FAX (613) 545-8067
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment 001 is being issued in response to Bidder's question and Canada's response.

Q1: *Could you please let us know if 220V is single phase or 3-phase?*

A1: 220V/60Hz single phase. This is the standard power supply we have.

Refer: Annex A, Section 3.15

Delete: in its entirety

Insert: The equipment must be able to be connected to the following building services: 220V / 60Hz, single phase. The vendor must supply any transformers, power conditioning or other components required to adapt to the available utilities.

Q2: *What is your range of sample thickness?*

A2: The range of sample thickness is: 0.18 mm to 0.79 mm. This complies with the ASTM Standard A343/A343M.

Q3: *What is the range of permeability you are wanting to measure?*

A3: No specific range of permeability is specified as this will highly depend on the materials to be tested.

Q4: *Are you interested in a Laboratory Quality Assurance (QA) or In-process Quality Control (QC) test system?*

A4: We are interested in a Laboratory Quality Assurance test system, i.e. the system will be used in laboratory for testing samples only.

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED