



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
B3J 1T3
Nova Scotia
Bid Fax: (902) 496-5016

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Atlantic Region Acquisitions/Région de l'Atlantique
Acquisitions
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
B3J 3C9
Nova Scot

Title - Sujet Liquid Chromatography (UHPLC) Syst.	
Solicitation No. - N° de l'invitation 39903-180323/A	Amendment No. - N° modif. 005
Client Reference No. - N° de référence du client 39903-18-0323	Date 2017-11-27
GETS Reference No. - N° de référence de SEAG PW-\$HAL-321-10220	
File No. - N° de dossier HAL-7-79084 (321)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2017-12-01	Time Zone Fuseau horaire Atlantic Standard Time AST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Baurin, Bruno	Buyer Id - Id de l'acheteur hal321
Telephone No. - N° de téléphone (902) 402-6891 ()	FAX No. - N° de FAX (902) 496-5016
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Contract No. - N° de l'invitation

39903-180323/A

Amd. No. - N° de la modif.

005

Buyer ID - Id de l'acheteur

HAL321

Client Ref. No. - N° de réf. du client

39903-180323

File No. - N° du dossier

HAL-7-79084

CCC No./N° CCC - FMS No/ N° VME

La modification #005 vise à répondre aux questions.

Question #1

Les articles 5c et 5d sont identiques.

Réponse #1

5c et 5d dans le tableau (p. 23 et 24) devrait être rédigée comme ils sont sur p.14 sur le document (ci-dessous).

Solicitation No. - N° de l'invitation
39903-180323/A
Client Ref. No. - N° de réf. du client
39903-180323

Amd. No. - N° de la modif.
File No. - N° du dossier
HAL-7-79084

Buyer ID - Id de l'acheteur
HAL321
CCC No./N° CCC - FMS No./N° VME

- m. The UHPLC System must be fully controlled by the same application software as the MS allowing for programmable automated sample sequences.
- n. Operator manuals for all equipment and software provided as hard copy or electronic version.

4. Additional Services

- a. Installation and qualification of the instrument will be required. The installing service engineer must connect and verify communication and control between the instrument and data system.
- b. The installing service engineer must provide an on-site user familiarization course covering general use and maintenance of the instrument at the time of installation.
- c. 1 Year Full service warranty from date of instrument commissioning at laboratory site must be provided.
- d. 2 Additional years of coverage under a full service contract must be included in the bid price.
- e. Warranty and service contract (3c & 3d) must cover parts, labour, and travel for the proposed system and include a minimum of 1 preventive maintenance visit per year.
- f. Guaranteed response time to a service call must be two business days or less. The response time is defined as the time from the logging of a problem by lab personnel with the offeror's service centre to the time of arrival of the engineer on-site. The offeror guarantees that actual response times will not exceed two business days.

5. Method Specific Performance Requirements

Please see Annex "B" for full details of the methodology required to demonstrate this performance.

- a. Under the test conditions detailed in Annex "B", the Instrument Detection Limit (IDL) for Chloramphenicol (CAS#: 56-75-7) must be lower than 40 femtograms on column at a 50 ms dwell time while operating in polarity switching mode.
- b. Under the test conditions detailed in Annex "B", the Instrument Detection Limit (IDL) for Sulfisoxazole (CAS#: 127-69-5) must be lower than 20 femtograms on column at a 50 ms dwell time while operating in polarity switching mode.
- c. Under the test conditions detailed in Annex "B", the Instrument Detection Limit (IDL) for Metronidazole (CAS#: 443-48-1) must be lower than 10 femtograms on column at a 50 ms dwell time.
- d. Under the test conditions detailed in Annex "B", the Instrument Detection Limit (IDL) for Diethylstilbestrol (CAS#: 56-53-1) must be lower than 100 femtograms on column at a 50 ms dwell time.
- e. Under the test conditions detailed in Annex "B", the Instrument Detection Limit (IDL) for Okadaic Acid CAS#: 78111-17-8) must be lower than 400 femtograms on column at a 50 ms dwell time.

TOUTES LES AUTRES MODALITÉS ET CONDITIONS RESTENT LES MÊMES.