



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
Pacific Region
401 - 1230 Government Street
Victoria, B.C.
V8W 3X4
Bid Fax: (250) 363-3344

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada - Pacific
Region
401 - 1230 Government Street
Victoria, B. C.
V8W 3X4

Title - Sujet X-Ray Fluorescence Analyzer	
Solicitation No. - N° de l'invitation W0103-18LL09/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client W0103-18LL09	Date 2018-01-04
GETS Reference No. - N° de référence de SEAG PW-\$VIC-240-7398	
File No. - N° de dossier VIC-7-40196 (240)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-01-15	
Time Zone Fuseau horaire Pacific Standard Time PST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Hogg(VIC), Mike	Buyer Id - Id de l'acheteur vic240
Telephone No. - N° de téléphone (250) 217-5640 ()	FAX No. - N° de FAX (250) 363-3344
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation
W0103-18LL09/A
Client Ref. No. - N° de réf. du client
W0103-18LL09

Amd. No. - N° de la modif.
File No. - N° du dossier

Buyer ID - Id de l'acheteur
vic240
CCC No./N° CCC - FMS No./N° VME

This amendment has been raised to incorporate the following changes to Annex A – Statement of Requirement

Please Delete:

Equipment must be no larger than length of 260 mm, width of 240 mm, and a depth of 100 mm; the instrument must weigh less than 1.3 kg (with battery installed);

Please Insert:

Equipment must be no larger than length of 270 mm, width of 240 mm, and a depth of 120 mm; the Instrument must weigh 1.5 kg or less (with battery installed)

All other terms and conditions remain unchanged