



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving
PWGSC
33 City Centre Drive
Suite 480C
Mississauga
Ontario
L5B 2N5
Bid Fax: (905) 615-2095

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Title - Sujet Dissolutions System		
Solicitation No. - N° de l'invitation H4004-173786/A		Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client H4004-173786		Date 2018-02-01
GETS Reference No. - N° de référence de SEAG PW-STOR-009-7462		
File No. - N° de dossier TOR-7-40134 (009)	CCC No./N° CCC - FMS No./N° VME	
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-02-20		Time Zone Fuseau horaire Eastern Standard Time EST
F.O.B. - F.A.B.		
Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>		
Address Enquiries to: - Adresser toutes questions à: Holvec, Monique		Buyer Id - Id de l'acheteur tor009
Telephone No. - N° de téléphone (905) 615-2062 ()		FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:		

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Issuing Office - Bureau de distribution

Public Works and Government Services Canada
Ontario Region
33 City Centre Drive
Suite 480
Mississauga
Ontario
L5B 2N5

Solicitation No. - N° de l'invitation
H4004-173786/A
Client Ref. No. - N° de réf. du client
H4004-173786

Amd. No. - N° de la modif.
002
File No. - N° du dossier
TOR-7-40134

Buyer ID - Id de l'acheteur
tor009
CCC No./N° CCC - FMS No./N° VME

Solicitation Amendment No. 002 is being issued to address the following:

- A) Questions and Answers
- B) Amend Annex "A", Requirement, Section 2.5.1, Dissolution System, item 11
- C) Amend Annex "D" Evaluation Criteria, Item 1.15

Amendment:

A) Questions and Answers

- Q1: With respect Annex "A", Section 2.5.1, Dissolution System, item 3. We have wired temperature sensors and there will be no impact on testing. Can you please change this specification to "Wireless to Wired or Both"?
- A1: Wireless temperature sensors are a key feature affecting the performance and reliability of the instrument. We will not remove them from the specifications.
- Q2: With respect Annex "A", Section 2.5.1, Dissolution System, item 6, "Must have a self-priming circulator/heater". Our heater is not self-priming. Minimal impact on daily work. Our heater has much lower vibration than any other system. Can you please remove the "self-priming" specification?
- A2: A self-priming pump is important for simple operation and robust performance. We will not remove it from the specifications.
- Q3: With respect Annex "A", Section 2.5.1, Dissolution System, item 11, "Must store at least 100 user-programmed methods". Our system offers 30 method slots instead of 100. If you want to interface the system with Dissolution Workstation Software, there are no limits to this. Can you please change the specification to "Must store at least 30 user-programmed methods"?
- A3: After a review of our needs, we would be willing to accept a system that can store 30 user-programmable methods.
- Q3: With respect Annex "A", Section 2.5.1, Dissolution System, item 11, "Must store at least 100 user-programmed methods". Our system offers 30 method slots instead of 100. If you want to interface the system with Dissolution Workstation Software, there are no limits to this. Can you please change the specification to "Must store at least 30 user-programmed methods"?
- A3: After a review of our needs, we would be willing to accept a system that can store 30 user-programmable methods.
- Please see below for further changes/information.
- Q4: With respect Annex "A", Section 2.5.1, Dissolution System, item 12, (a), RPM Control Range: 25-300 RPM or wider. Flow rates for our autosampler are 6-12ml/min. Can you please remove the specification "RPM control range 25-300 RPM or Wider"?
- A4: We cannot agree to the prospective bidder's request to remove the RPM control range from the specifications.

Solicitation No. - N° de l'invitation
H4004-173786/A
Client Ref. No. - N° de réf. du client
H4004-173786

Amd. No. - N° de la modif.
002
File No. - N° du dossier
TOR-7-40134

Buyer ID - Id de l'acheteur
tor009
CCC No./N° CCC - FMS No./N° VME

B) At Annex "A", Requirement, Section 2.5.1, Dissolution System, item 11

Delete in its entirety;

Insert: Must store at least 30 user-programmed methods.

C) At Annex "D" Evaluation Criteria, Item 1.15

Delete in its entirety;

Insert: Dissolution Autosampler must store at least 30 methods.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME