



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Travaux publics et Services gouvernementaux
Canada
Place Bonaventure,
800 rue de la Gauchetière Ouest
Voir aux présentes - See herein
Montréal
Québec
H5A 1L6

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Travaux publics et Services gouvernementaux Canada
Place Bonaventure,
800 rue de la Gauchetière Ouest
Voir aux présentes - See herein
Montréal
Québec
H5A 1L6

Title - Sujet Séchoir atomiseur petit	
Solicitation No. - N° de l'invitation 01B30-181103/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client 01B30-18-1103	Date 2018-02-02
GETS Reference No. - N° de référence de SEAG PW-SMTA-405-14708	
File No. - N° de dossier MTA-7-40337 (405)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-02-21	
Time Zone Fuseau horaire Heure Normale du l'Est HNE	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Séguin, Caroline	Buyer Id - Id de l'acheteur mta405
Telephone No. - N° de téléphone (514) 703-0455 ()	FAX No. - N° de FAX (514) 496-3822
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

AMENDMENT 002

The purpose of this amendment is to publish questions and answers regarding this request.

Questions and Answers

Q1: The weight of a single crate containing the heaviest item (chamber), is 600 kg (± 40 kg). Could you please confirm that your client will have the necessary equipment to unload this crate from the delivery truck and transport it to the location of installation in the facility?

A1: Yes.

Q2: Our Control System will be supplied as a standard for installation in the United States. For operation in Canada however, we will supply the required Transformer. Installation of the Transformer, primary and secondary fusing, and all interconnection piping and wiring would be provided by the customer. Could you please confirm that the Transformer installation would be by the customer?

A2: Yes we will install the supplied transformer.

Q3: The following utilities will be required in order to operate the Small Spray Drying System. Could you please confirm that the customer will have these utilities available onsite at the location of installation and will make the necessary wiring or pneumatic piping connections to our System?

- a. The system's control panel (to be located in the same room as the Small Spray Dryer) will require the following from the transformer: 480 VAC / 60 Hz/ 3 phase 40A
- b. Pneumatic Requirements:

Description	Pressure	Consumption	Units
Chamber Roof	Pressurized Air 0 – 4 bar	0.05	Nm ³ /h
Rotary Atomizer	Pressurized air 0 – 6 bar	17	Nm ³ /h**
Two-Fluid Nozzle, Co-Current	Pressurized air 0 – 6 bar	8 – 15 at 2 bar (10 – 19)	Nm ³ /h (m ³ /h)
Cartridge Filter	Pressurized air 3 – 6 bar	0.5 – 2	Nm ³ /h

Max Pressure: 6 – 8 bar, Max Consumption: 35 – 40 Nm³/h

**Rotary Atomizer and Two-Fluid Nozzle will not be run simultaneously.

A3: Yes, these utilities are all available on site.

All other terms and conditions remain unchanged