

ELECTRICAL SAFETY AUTHORITY REGISTERED



80 Bullock Drive
Unit 1
Markham, Ontario
Canada, L3P 3P7



Phone: 905-294-5274
Fax: 905-294-6096
E-mail: info@cshunt.com
www.cshunt.com

CERTIFICATE OF CALIBRATION

Country of Origin: CANADA

Date of Calibration: 11-Apr-18
Customer: Canadian Coast Guard C.C.G.S. Griffon
Shunt Part #: 90000-968
Ser.#: 110481
Calibration Due: 11-Apr-19

Date of Issue: 11-Apr-18
P.O. #: F2599-18-5601
Invoice #: 070512

Ambient Temp. 23°C ±2°

RANGE: 2000 D.C. Amperes, 50 D.C. Millivolts,
RESISTANCE: 0.000025000Ω, Uncertainty ±0.1%,
k=2, at a level of confidence of approximately 95%
MEASURED USING A CURRENT OF 100 D.C.AMPERES

This measurement is traceable to recognized national standards and to the SI or derived units of measurement realized at the corresponding national standards laboratory of the National Research Council (NRC) or the National Institute of Standards and Technology (NIST). To the best of our knowledge C.S.I. Products do not contain any of the REACH Substances of Very High Concern. C.S.I. is compliant with RoHS Directive 2011/65/EU.

DOCUMENTS ON FILE AT OUR FACILITY

Marc Matthews
Quality Manager

Primary Standards:-

Yokogawa Kelvin Bridge, Model 2752, S/N 60VL0018, Cal. Due 1st Mar., 2019
CSI Standard Shunt, S/N 2812, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 050226-1, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 220212, Cal. Due 7th Feb., 2019

SERVING INDUSTRY SINCE 1972

ELECTRICAL SAFETY AUTHORITY REGISTERED



80 Bullock Drive
Unit 1
Markham, Ontario
Canada, L3P 3P7



Phone: 905-294-5274
Fax: 905-294-6096
E-mail: info@cshunt.com
www.cshunt.com

CERTIFICATE OF CALIBRATION

Country of Origin: CANADA

Date of Calibration: 11-Apr-18
Customer: Canadian Coast Guard C.C.G.S. Griffon
Shunt Part #: 90000-968
Ser.#: 110482
Calibration Due: 11-Apr-19

Date of Issue: 11-Apr-18
P.O. #: F2599-18-5601
Invoice #: 070512

Ambient Temp. 23°C ±2°

RANGE: 2000 D.C. Amperes, 50 D.C. Millivolts,
RESISTANCE: 0.000025038Ω, Uncertainty ±0.1%,
k=2, at a level of confidence of approximately 95%
MEASURED USING A CURRENT OF 100 D.C.AMPERES

This measurement is traceable to recognized national standards and to the SI or derived units of measurement realized at the corresponding national standards laboratory of the National Research Council (NRC) or the National Institute of Standards and Technology (NIST). To the best of our knowledge C.S.I. Products do not contain any of the REACH Substances of Very High Concern. C.S.I. is compliant with RoHS Directive 2011/65/EU.

DOCUMENTS ON FILE AT OUR FACILITY

Marc Matthews
Quality Manager

Primary Standards:-

Yokogawa Kelvin Bridge, Model 2752, S/N 60VL0018, Cal. Due 1st Mar., 2019
CSI Standard Shunt, S/N 2812, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 050226-1, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 220212, Cal. Due 7th Feb., 2019

SERVING INDUSTRY SINCE 1972

ELECTRICAL SAFETY AUTHORITY REGISTERED



80 Bullock Drive
Unit 1
Markham, Ontario
Canada, L3P 3P7



Phone: 905-294-5274
Fax: 905-294-6096
E-mail: info@cshunt.com
www.cshunt.com

CERTIFICATE OF CALIBRATION

Country of Origin: CANADA

Date of Calibration: 11-Apr-18
Customer: Canadian Coast Guard C.C.G.S. Griffon
Shunt Part #: 90000-968
Ser.#: 110483
Calibration Due: 11-Apr-19

Date of Issue: 11-Apr-18
P.O. #: F2599-18-5601
Invoice #: 070512

Ambient Temp. 23°C ±2°

RANGE: 2000 D.C. Amperes, 50 D.C. Millivolts,
RESISTANCE: 0.000025000Ω, Uncertainty ±0.1%,
k=2, at a level of confidence of approximately 95%
MEASURED USING A CURRENT OF 100 D.C.AMPERES

This measurement is traceable to recognized national standards and to the SI or derived units of measurement realized at the corresponding national standards laboratory of the National Research Council (NRC) or the National Institute of Standards and Technology (NIST). To the best of our knowledge C.S.I. Products do not contain any of the REACH Substances of Very High Concern. C.S.I. is compliant with RoHS Directive 2011/65/EU.

DOCUMENTS ON FILE AT OUR FACILITY

Marc Matthews
Quality Manager

Primary Standards:-

Yokogawa Kelvin Bridge, Model 2752, S/N 60VL0018, Cal. Due 1st Mar., 2019
CSI Standard Shunt, S/N 2812, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 050226-1, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 220212, Cal. Due 7th Feb., 2019

SERVING INDUSTRY SINCE 1972

ELECTRICAL SAFETY AUTHORITY REGISTERED



80 Bullock Drive
Unit 1
Markham, Ontario
Canada, L3P 3P7



Phone: 905-294-5274
Fax: 905-294-6096
E-mail: info@cshunt.com
www.cshunt.com

CERTIFICATE OF CALIBRATION

Country of Origin: CANADA

Date of Calibration: 11-Apr-18
Customer: Canadian Coast Guard C.C.G.S. Griffon
Shunt Part #: 90000-968
Ser.#: 110484
Calibration Due: 11-Apr-19

Date of Issue: 11-Apr-18
P.O. #: F2599-18-5601
Invoice #: 070512

Ambient Temp. 23°C ±2°

RANGE: 2000 D.C. Amperes, 50 D.C. Millivolts,
RESISTANCE: 0.000025000Ω, Uncertainty ±0.1%,
k=2, at a level of confidence of approximately 95%
MEASURED USING A CURRENT OF 100 D.C.AMPERES

This measurement is traceable to recognized national standards and to the SI or derived units of measurement realized at the corresponding national standards laboratory of the National Research Council (NRC) or the National Institute of Standards and Technology (NIST). To the best of our knowledge C.S.I. Products do not contain any of the REACH Substances of Very High Concern. C.S.I. is compliant with RoHS Directive 2011/65/EU.

DOCUMENTS ON FILE AT OUR FACILITY

Marc Matthews
Quality Manager

Primary Standards:-

Yokogawa Kelvin Bridge, Model 2752, S/N 60VL0018, Cal. Due 1st Mar., 2019
CSI Standard Shunt, S/N 2812, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 050226-1, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 220212, Cal. Due 7th Feb., 2019

SERVING INDUSTRY SINCE 1972

ELECTRICAL SAFETY AUTHORITY REGISTERED



80 Bullock Drive
Unit 1
Markham, Ontario
Canada, L3P 3P7



Phone: 905-294-5274
Fax: 905-294-6096
E-mail: info@cshunt.com
www.cshunt.com

CERTIFICATE OF CALIBRATION

Country of Origin: CANADA

Date of Calibration: 11-Apr-18
Customer: Canadian Coast Guard C.C.G.S. Griffon
Shunt Part #: 90000-1277
Ser.#: 110485
Calibration Due: 11-Apr-19

Date of Issue: 11-Apr-18
P.O. #: F2599-18-5601
Invoice #: 070512

Ambient Temp. 23°C ±2°

RANGE: 3000 D.C. Amperes, 50 D.C. Millivolts,
RESISTANCE: 0.000016666Ω, Uncertainty ±0.1%,
k=2, at a level of confidence of approximately 95%
MEASURED USING A CURRENT OF 100 D.C.AMPERES

This measurement is traceable to recognized national standards and to the SI or derived units of measurement realized at the corresponding national standards laboratory of the National Research Council (NRC) or the National Institute of Standards and Technology (NIST). To the best of our knowledge C.S.I. Products do not contain any of the REACH Substances of Very High Concern. C.S.I. is compliant with RoHS Directive 2011/65/EU.

DOCUMENTS ON FILE AT OUR FACILITY

Marc Matthews
Quality Manager

Primary Standards:-

Yokogawa Kelvin Bridge, Model 2752, S/N 60VL0018, Cal. Due 1st Mar., 2019
CSI Standard Shunt, S/N 2812, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 050226-1, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 220212, Cal. Due 7th Feb., 2019

SERVING INDUSTRY SINCE 1972

ELECTRICAL SAFETY AUTHORITY REGISTERED



80 Bullock Drive
Unit 1
Markham, Ontario
Canada, L3P 3P7



Phone: 905-294-5274
Fax: 905-294-6096
E-mail: info@cshunt.com
www.cshunt.com

CERTIFICATE OF CALIBRATION

Country of Origin: CANADA

Date of Calibration: 11-Apr-18
Customer: Canadian Coast Guard C.C.G.S. Griffon
Shunt Part #: 90000-1277
Ser.#: 110486
Calibration Due: 11-Apr-19

Date of Issue: 11-Apr-18
P.O. #: F2599-18-5601
Invoice #: 070512

Ambient Temp. 23°C ±2°

RANGE: 3000 D.C. Amperes, 50 D.C. Millivolts,
RESISTANCE: 0.000016666Ω, Uncertainty ±0.1%,
k=2, at a level of confidence of approximately 95%
MEASURED USING A CURRENT OF 100 D.C.AMPERES

This measurement is traceable to recognized national standards and to the SI or derived units of measurement realized at the corresponding national standards laboratory of the National Research Council (NRC) or the National Institute of Standards and Technology (NIST). To the best of our knowledge C.S.I. Products do not contain any of the REACH Substances of Very High Concern. C.S.I. is compliant with RoHS Directive 2011/65/EU.

DOCUMENTS ON FILE AT OUR FACILITY

A handwritten signature in blue ink.

Marc Matthews
Quality Manager

Primary Standards:-

Yokogawa Kelvin Bridge, Model 2752, S/N 60VL0018, Cal. Due 1st Mar., 2019
CSI Standard Shunt, S/N 2812, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 050226-1, Cal. Due 7th Feb., 2019
CSI Standard Shunt, S/N 220212, Cal. Due 7th Feb., 2019

SERVING INDUSTRY SINCE 1972