



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
Cabot Place, Phase II, 2nd Floor
Box 4600
St. John's, NF
A1C 5T2
Bid Fax: (709) 772-4603

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
PWGSC / TPSGC - Nfld. Region
Cabot Place, Phase II, 2nd Floor
Box 4600
St. John's, NF
A1C 5T2

Title - Sujet Atmospheric Visibility Sensors	
Solicitation No. - N° de l'invitation F6839-184064/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client F6839-184064	Date 2018-06-18
GETS Reference No. - N° de référence de SEAG PW-SOLZ-009-7093	
File No. - N° de dossier OLZ-8-41023 (009)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-07-03	
Time Zone Fuseau horaire Newfoundland Daylight Saving Time NDT	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Lacey, Rhonda	Buyer Id - Id de l'acheteur olz009
Telephone No. - N° de téléphone (709) 772-8057 ()	FAX No. - N° de FAX (709) 772-4603
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

AMENDMENT 02

Amendment 02 is being issued to clarify the following:

- Is a calibration fixture required for each unit?

Yes

- Ground transport time estimate alone is (3) to (4) weeks from manufacturing facility in California to Newfoundland. Would we include the transport time in with the manufacturing lead time.

Estimated Time of Delivery required

- Is the requested delivery time flexible?

Six weeks within Contract award

- As we are committed to offer our clients the best possible equipment solution to suit all of their needs, and for ease of integration with the existing equipment can you please give us a basic description of the intended location of your new visibility sensors. Will this equipment be located at sea, on a buoy for instance or perhaps over land, nearby coastline on a meteorological mast? Will it be part of existing marine weather stations?

Nearby Coastline. Fog Detection.

- Can you please be more specific with your instrumentation requirements or would you like to simply negotiate those sort of things separately once you have established your chosen instrument and vendor? Preferences such as:

- Sensor hood heating: Do you require the optional heated version?

Optional Hood Heater Not Required

- Data output preference: RS232 or RS422/RS485 serial output, 0-20mA or 4-20mA?

RS232

- Cable length and cable gland types : Any specific requirements here? Length of cable, plastic or metal cable glands?

Standard Cable Length. Corrosion Resistant Materials

- Do you require self-test monitoring for the instrument? Standard or advanced?

Contract No. - N° de l'invitation

F6839-184064/001/OLZ

Client Ref. No. - N° de réf. du client

F6839-184064

Amd. No. - N° de la modif.

002

File No. - N° du dossier

OLZ-8-41023

Buyer ID - Id de l'acheteur

olz009

CCC No./N° CCC - FMS No./N° VME

Standard

- Do you require any special configuration? With or without weather station module?
Without

- Would you be interested in getting an offer for optional extended warranty, beyond the first two years?

Include in Bid

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED