



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

**Bid Receiving - PWGSC / Réception des soumissions -
TPSGC**

11 Laurier St./ 11 rue, Laurier

Place du Portage, Phase III

Core 0B2 / Noyau 0B2

Gatineau, Québec K1A 0S5

Bid Fax: (819) 997-9776

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

**Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution

Scientific, Medical and Photographic Division /
Division de l'équipement scientifique, des produits
photographiques et pharmaceutiques

11 Laurier St./ 11 rue, Laurier

6A2, Place du Portage

Gatineau, Québec K1A 0S5

Title - Sujet ULTRA CENTRIFUGE AND ROTOR	
Solicitation No. - N° de l'invitation K2A51-181324/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client K2A51-181324	Date 2018-07-18
GETS Reference No. - N° de référence de SEAG PW-\$\$PV-941-75037	
File No. - N° de dossier pv941.K2A51-181324	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-08-07	Time Zone Fuseau horaire Eastern Daylight Saving Time EDT
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Martins, Christina	Buyer Id - Id de l'acheteur pv941
Telephone No. - N° de téléphone (613) 355-1973 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment 001 has been raised to respond to the following questions and to revise the RFP accordingly.

Q1. "Must be able to reach 150,000 rpm in less than 2 minutes".

This is not a typical speed specification, could this required specification criteria reflect that "the instrument proposed MUST be able to reach 150,000 rpm." and remove the section about "less than 2 minutes"?

A1. It is possible to split this criteria in two: This required specification criteria must reflect that;
"The instrument proposed must be able to reach 150,000 rpm."
"The instrument must reach the maximum speed in less than 2 minutes".

Q2. The following specification spec: "Self-locking rotor system that secures the rotor to the drive shaft and eliminates the need to manually secure it".

Could you change the mandatory criteria specification to indicate that "The instrument proposed MUST offer a method of operation that secures the rotor to the drive shaft"?

A2. The mandatory criteria specification will be revised to indicate the following:
"The instrument proposed must offer a method of operation that automatically secures the rotor to the drive shaft."

**

1. Delete:

Page 14 of 20, Annex "A": Part 1 - REQUIREMENT

ANNEX "A" **Part 1 - REQUIREMENT**

Environment and Climate Change Canada has a requirement for the supply of a Centrifuge and rotor which must include manuals and meet all of the mandatory technical requirements as specified below and the mandatory evaluation criteria as specified at Part 2.1 - Mandatory Technical Evaluation Criteria.

The requirement must include all of the following:

- Compact benchtop model;
- Must be able to reach a minimum of 150,000 rpm (in less than 2 min);
- Self-locking rotor system that secures the rotor to the drive shaft and eliminates the need to manually secure it;
- Temperature controlled (4° to 20°C); and
- Must include a fixed angle rotor with a capacity of at least 8 tubes of 5 mL minimum."

Replace with:

Page 14 of 20, Annex "A": Part 1 - REQUIREMENT

ANNEX "A" **Part 1 - REQUIREMENT**

Environment and Climate Change Canada has a requirement for the supply of a Centrifuge and rotor which must include manuals and meet all of the mandatory technical requirements as specified below

and the mandatory evaluation criteria as specified at Part 2.1 - Mandatory Technical Evaluation Criteria.

The requirement must include all of the following:

- Compact benchtop model;
- The instrument proposed must be able to reach 150,000 rpm;
- The instrument must reach the maximum speed in less than 2 minutes;
- The instrument proposed must offer a method of operation that automatically secures the rotor to the drive shaft;
- Temperature controlled (4° to 20°C); and
- Must include a fixed angle rotor with a capacity of at least 8 tubes of 5 mL minimum.”

ALL OTHER TERMS AND CONDITIONS IN THE RFP REMAIN UNCHANGED.