



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
Cabot Place, Phase II, 2nd Floor
Box 4600
St. John's, NL
A1C 5T2
Bid Fax: (709) 772-4603

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
PWGSC / TPGSC - Nfld. Region
Cabot Place, Phase II, 2nd Floor
Box 4600
St. John's, NL
A1C 5T2

| | |
|---|--|
| Title - Sujet Agri Cnd - Reach-In Growth Chamber | |
| Solicitation No. - N° de l'invitation 01800-190450/A | Amendment No. - N° modif. 001 |
| Client Reference No. - N° de référence du client 01800-190450 | Date 2018-10-16 |
| GETS Reference No. - N° de référence de SEAG PW-\$XAQ-040-7185 | |
| File No. - N° de dossier XAQ-8-41120 (040) | CCC No./N° CCC - FMS No./N° VME |
| Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-11-06 | Time Zone Fuseau horaire Newfoundland Standard Time NST |
| F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/> | |
| Address Enquiries to: - Adresser toutes questions à: Connolly, Carolyn | Buyer Id - Id de l'acheteur xaq040 |
| Telephone No. - N° de téléphone (709) 772-5396 () | FAX No. - N° de FAX (709) 772-4603 |
| Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: | |

Instructions: See Herein

Instructions: Voir aux présentes

| | |
|--|--|
| Delivery Required - Livraison exigée | Delivery Offered - Livraison proposée |
| Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur | |
| Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur | |
| Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie) | |
| Signature | Date |

Solicitation No. - N° de l'invitation
01800-190450/A

Amd. No. - N° de la modif.
001

Buyer ID - Id de l'acheteur
xaq040

Client Ref. No. - N° de réf. du client
01800-190450

File No. - N° du dossier
XAQ-8-41120

CCC No./N° CCC - FMS No./N° VME

TABLE OF CONTENTS

AMENDMENT 001..... 2

1.1 AMENDMENT 001.....2

AMENDMENT 001

1.1 Amendment 001

Amendment 001 is raised to address the following:

Regarding specification for temperature range:

Reads -10°C (with lights off), or +2C (lights on) to +45C

Q1. Temperature as low as -10C lights off and 2C lights on will require that the chamber go through periodic defrost cycles. Defrost cycles will occur every 2 -3 hours, typically last 15 minutes during which the temperature could rise by about 7C in the growth area. Is it acceptable to have a temperature spike in the growth area during the defrost cycle.

➤ Answer: Acceptable

Q2. Most manufacturers of horticultural LED lighting specify a maximum ambient condition of +40C. Is this acceptable?

➤ Answer: Acceptable

Regarding specification for additional electrical outlets.

Specification reads 3 electrical outlets for plugging in equipment.

Q3. Would a single 120V 6A circuit feeding all three outlets be acceptable?

➤ Answer: Acceptable

Regarding chamber set-up.

Specification reads: Staff on-site will unload, perform power and water hook-ups.

Q4. Does "Staff" refer to AAFC staff or does it refer to the growth chamber manufactures staff?

➤ Answer: AAFC staff

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED