



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

**Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada**

1713 Bedford Row

Halifax, N.S./Halifax, (N.É.)

Halifax

Nova Scotia

B3J 1T3

Bid Fax: (902) 496-5016

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

**Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution

**Atlantic Region Acquisitions/Région de l'Atlantique
Acquisitions**

1713 Bedford Row

Halifax, N.S./Halifax, (N.É.)

Halifax

Nova Scot

B3J 1T3

Title - Sujet Calorimeter	
Solicitation No. - N° de l'invitation 31028-185597/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client 31028-18-5597	Date 2018-11-05
GETS Reference No. - N° de référence de SEAG PW-SHAL-409-10528	
File No. - N° de dossier HAL-8-81149 (409)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-12-03	Time Zone Fuseau horaire Atlantic Standard Time AST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Taylor, Kathie	Buyer Id - Id de l'acheteur hal409
Telephone No. - N° de téléphone (902) 403-4837 ()	FAX No. - N° de FAX (902) 496-5016
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation
31028-185597
Client Ref. No. - N° de réf. du client
31028-18-5597

Amd. No. - N° de la modif.
001
File No. - N° du dossier
HAL-8-81149

Buyer ID - Id de l'acheteur
HAL409
CCC No./N° CCC - FMS No./N° VME

Amendment 001 is generated to respond to a question raised after the release of the Solicitation.

Q1: The description given is controversial of two of calorimeter models. The first Model that would be used with the 6510 water handling style and the Second model with fixed bomb style and quick lock head. So technically, they cannot be combined in a single machine. I appreciate it if you could please provide us with more information.

A1: The 6510 water handling system is an 'accessory' to the calorimeter but not required for the calorimeter to function if you have access to unlimited and constant temperature freshwater. This is not the case in our remote lab where freshwater must be shipped into our site which is expensive and we do not have a large storage capacity. Thus, we require a calorimeter that can work compatibly with our existing infrastructure to reduce water demands.

All other terms and conditions remain the same.