



**RETURN BIDS TO:**

**RETOURNER LES SOUMISSIONS À:**

Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3  
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**  
**Raison sociale et adresse du**  
**fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3

<b>Title - Sujet</b> Autoclave	
<b>Solicitation No. - N° de l'invitation</b> 01689-190309/A	<b>Amendment No. - N° modif.</b> 002
<b>Client Reference No. - N° de référence du client</b> 01689-19-0309	<b>Date</b> 2018-11-16
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$KIN-620-7591	
<b>File No. - N° de dossier</b> KIN-8-50052 (620)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2018-11-27</b>	
<b>Time Zone</b> <b>Fuseau horaire</b> Eastern Daylight Saving Time EDT	
<b>F.O.B. - F.A.B.</b> Specified Herein - Précisé dans les présentes <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input type="checkbox"/> <b>Other-Autre:</b> <input checked="" type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Porter, Marta M.	<b>Buyer Id - Id de l'acheteur</b> kin620
<b>Telephone No. - N° de téléphone</b> (613) 547-7587 ( )	<b>FAX No. - N° de FAX</b> (613) 545-8067
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

**Amendment 002 – Autoclave, is being raised for the following reason:**

- 1. Provide a response back from Public Works and Government Services Canada to questions received from bidders.**
- 

- Q1.** In Annexe A, section 10 Training, cell 10.1, please confirm that it refers to user/ in-service training.
- R1.** Yes “training” refers to the vendor teaching users how to use the autoclave once it is installed and running.
- Q2.** In Annexe A, section 13 Installation, cell 13.1, please confirm who is responsible for the utility connections from the unit to the building.  
The University's contractor?
- R2.** The building's utilities will be brought within 1 metre of the installation site by our staff, but the vendor will be responsible for connecting the autoclave to the building's utilities. The vendor must coordinate with our Facilities staff and provide a minimum of 3 days notice prior to making this connection.
- Q3.** Referring to the tender criteria – Item # 5 Door and Seal - Could you advise if Public Works will allow for non-hinged door to be supplied for this sterilizer. My manufacturer builds a vertical sliding door for this size 26 x 26 x 39 (B Size) Small Sterilizer. Please advise if this alternate style door is acceptable so that we can provide a quote for this tender.
- R3.** We will accept a vertical door in 5.1 if it is a manual door and the unit meets all of the other specifications.