



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
Halifax
Nova Scotia
B3J 1T3
Bid Fax: (902) 496-5016

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Atlantic Region Acquisitions/Région de l'Atlantique
Acquisitions
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
Halifax
Nova Scot
B3J 1T3

Title - Sujet Microwave Analyzer	
Solicitation No. - N° de l'invitation W355B-192239/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client W355B-19-2239	Date 2018-11-22
GETS Reference No. - N° de référence de SEAG PW-\$HAL-222-10525	
File No. - N° de dossier HAL-8-81076 (222)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2018-12-06	
Time Zone Fuseau horaire Atlantic Standard Time AST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: St-Louis, Catherine	Buyer Id - Id de l'acheteur hal222
Telephone No. - N° de téléphone (902) 240-5159 ()	FAX No. - N° de FAX (902) 496-5016
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

**Microwave Analyzer
W355B-192239/A
Amendment No 002**

The purpose of this amendment is to answer the following questions and extend the solicitation closing date.

At page 1, Solicitation Closes:

Delete: 2018-11-28

Insert: 2018-12-06

Question 2a

The instrument is required to have the Vector Network Analyzer option: can you please specified if DND need the full 2-ports S parameter or only Transmission/Reflexion?

Answer 2a

Both. We need Option 211 (VNA Full 2-Port S-Para) and option 210 (Vector network Analyzer Trans/Reflection) is a requirement for that feature (211).

Question 2b

The instrument is required to have the Spectrum analyzer option: Do you also need the Pre-amp, Interference Analysis or RTSA

Answer 2b

Yes we need N9917A-233 Pre-Amp, Interference Analysis N9917A-236 allows signal playback and no RTSA if the VSA software is purchased.

Question 2c

The instrument is required to have USB Wide Band Power Meter: Can you please indicate the Frequency, the dynamics range and the type of sensor(or P/N) :

Answer 2c

Average, Peak or both, 50 MHz to 18 GHz, -30 to +20 dBm (peak/gated) -35 to +20 dBm (average): we need a USB power sensor support 302 and a U2021XA USB Wideband Power Sensor (50MHz-18GHz)

Question 2d

Are you planning to connect the USB power sensor to the instrument or on a PC? If you need to connect the sensor to the instrument?

Answer 2d

Yes

Question 2e

The following option is a prerequisite and not listed on your requirement: Option 302: External USB Power Sensor Support: Adds Keysight USB power sensor support to FieldFox analyzers. Frequency and power range depend on sensor

Answer 2e

Yes, we need a USB power sensor support 302 and a U2021XA USB Wideband Power Sensor (50MHz-18GHz)

Question 2f

With the option GPS Receiver; do you need the GPS antenna?

Answer 2f

Yes

Question 2g

Do you need a hard transit case?

Answer 2g

Yes

Question 2h

For the VNA option; do you need a calibration kit?

Answer 2h

Not an external cal kit

Question 2i

Do you need RF Test cable with the instrument?

Answer 2i

No

Question 2j

Do you need extra batteries, or DC adapter?

Answer 2j

Extra battery (N9910X-870) and Charger (N9910X-872). Also required is the N9917A-330 pulse measurement.

ALL OTHER TERMS AND CONDITIONS REMANIN UNCHANGED.