



## RETURN BIDS TO:

## RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government  
Services Canada/Réception des soumissions Travaux  
publics et Services gouvernementaux Canada  
800 Burrard Street, Room 219  
800, rue Burrard, pièce 219  
Vancouver, BC V6Z 0B9  
Bid Fax: (604) 775-7526

## SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

### Comments - Commentaires

Vendor/Firm Name and Address  
Raison sociale et adresse du  
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution  
Public Works and Government Services Canada - Pacific  
Region  
219 - 800 Burrard Street  
800, rue Burrard, pièce 219  
Vancouver, BC V6Z 0B9

|   |  |
|---|--|
| <b>Title - Sujet</b><br>Automated Density Meter System  |  |
| <b>Solicitation No. - N° de l'invitation</b><br>K2A15-190285/B  | <b>Amendment No. - N° modif.</b><br>002      |
| <b>Client Reference No. - N° de référence du client</b><br>K2A15-190285   | <b>Date</b><br>2018-11-23                    |
| <b>GETS Reference No. - N° de référence de SEAG</b><br>PW-\$VAN-532-8474  |  |
| <b>File No. - N° de dossier</b><br>VAN-8-41136 (532)  | <b>CCC No./N° CCC - FMS No./N° VME</b>       |
| <b>Solicitation Closes - L'invitation prend fin</b><br><b>at - à 02:00 PM</b><br><b>on - le 2018-12-17</b>  |  |
| <b>Time Zone</b><br>Fuseau horaire<br>Pacific Standard Time<br>PST  |  |
| <b>F.O.B. - F.A.B.</b><br><b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/> |  |
| <b>Address Enquiries to: - Adresser toutes questions à:</b><br>Takasaki, Alan H.  | <b>Buyer Id - Id de l'acheteur</b><br>van532 |
| <b>Telephone No. - N° de téléphone</b><br>(604) 671-1831 ( )  | <b>FAX No. - N° de FAX</b><br>(604) 775-7526 |
| <b>Destination - of Goods, Services, and Construction:</b><br><b>Destination - des biens, services et construction:</b>   |  |

Instructions: See Herein

Instructions: Voir aux présentes

|   |  |
|---|--|
| <b>Delivery Required - Livraison exigée</b>   | <b>Delivery Offered - Livraison proposée</b> |
| <b>Vendor/Firm Name and Address</b><br>Raison sociale et adresse du fournisseur/de l'entrepreneur   |  |
| <b>Telephone No. - N° de téléphone</b><br><b>Facsimile No. - N° de télécopieur</b>  |  |
| <b>Name and title of person authorized to sign on behalf of Vendor/Firm</b><br>(type or print)<br><b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b><br><b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b> |  |
| <b>Signature</b>  | <b>Date</b>                                  |

Solicitation No. - N° de l'invitation  
K2A25-190285/B  
Client Ref. No. - N° de réf. du client  
K2A25-190285

Amd. No. - N° de la modif.  
A002  
File No. - N° du dossier

Buyer ID - Id de l'acheteur  
VAN532  
CCC No./N° CCC - FMS No./N° VME

---

**Solicitation #: K2A25-190285/B**

The following questions and answers are for clarification purposes.

If due to the clarification an amendment to the solicitation is required, the amendment will be stated herein as "The above referenced solicitation is amended as follows":

**Question 1:** - Annex A: Section 3e to 3h: must include an external computer equipped with software and licence for instrument's operation Is providing an external computer equipped with software and licence mandatory if all software and calculations capability is integrated directly into the instrument (essentially making the external computer, software and licence unnecessary for the instrument's operation) ?

**Response 1:** An external computer system (or laptop) is required, as the system must be able to connect to our Laboratory Information Management System (LIMS) for automatic data upload. Computer system (or laptop) is also required for the requirement described in paragraph 3g of Annex A

**Question 2:** Annex A - In order to offer the proper density meter model, we need to know the expected density accuracy. What is the required density accuracy?

**Response 2:** Required accuracy is 4 (four) digits.

**Question 3: - Section 6.6.3 Method of Payment**

While SACC Clause H1000C describes that a single payment will be made upon delivery and work, it does not specify \*when\* this will actually be done. Could you let us know how long after all the work and documents have been completed that the payment will be made?

<https://buyandsell.gc.ca/policy-and-guidelines/standard-acquisition-clauses-and-conditions-manual/5/H/H1000C/2>

**Response 3:**

See Article 15 - 2010A – Payment Period

<https://buyandsell.gc.ca/policy-and-guidelines/standard-acquisition-clauses-and-conditions-manual/3/2010A/18#payment-period>

**Question 4:** - section 6.4.1 Delivery date.

In the bid document, the section is blank. What is the required delivery date?

**Response 4:** March 31, 2019

**The above solicitation is amended as follows:**

**PART 6 – RESULTING CONTRACT CLAUSING**

Section 6.4.1 Delivery date

**Suppression :** All the deliverables must be received on or before \_\_\_\_\_ (*insert the date*).

Solicitation No. - N° de l'invitation  
K2A25-190285/B  
Client Ref. No. - N° de réf. du client  
K2A25-190285

Amd. No. - N° de la modif.  
A002  
File No. - N° du dossier

Buyer ID - Id de l'acheteur  
VAN532  
CCC No./N° CCC - FMS No./N° VME

---

**Insertion:** All the deliverables must be received on or before March 31, 2019.

#### **Annex A**

**1. Replace 3 e. with:**

The system must include an external computer system or laptop for Laboratory Information Management System (LIMS) connectivity for data upload and security, as well as external data manipulation and statistical analysis capability. The computer system must be Windows 10 or newer and with at least 1 free USB port.

**2. Insert in Section 3 Mandatory Specifications - Annex A**

i. The system must have an accuracy of 4 digits

#### **FORM B**

**1. Replace 3 e. with**

The system must include an external computer system or laptop for Laboratory Information Management System (LIMS) connectivity for data upload and security, as well as external data manipulation and statistical analysis capability. The computer system must be Windows 10 or newer and with at least 1 free USB port.

**2. Insert in Section 3 Mandatory Specifications:**

i. The system must have an accuracy of 4 digits

All other terms and condition remain unchanged.